

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029991.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.141	9.044	9.244	736.550	712.500	3.4
2,4,5-TP(Silvex)	8.858	8.762	8.962	730.550	712.500	2.5
2,4-D	8.016	7.919	8.119	715.650	705.000	1.5
2,4-DB	9.699	9.603	9.803	764.140	712.500	7.2
2,4-DCAA	6.939	6.842	7.042	758.550	750.000	1.1
DICAMBA	7.116	7.018	7.218	723.860	705.000	2.7
DICHLORPROP	7.796	7.699	7.899	704.420	705.000	-0.1
Dinoseb	10.861	10.766	10.966	714.800	705.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029991.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.938	9.847	10.047	776.840	712.500	9.0
2,4,5-TP(Silvex)	9.534	9.442	9.642	786.260	712.500	10.4
2,4-D	8.655	8.561	8.761	748.240	705.000	6.1
2,4-DB	10.497	10.407	10.607	729.270	712.500	2.4
2,4-DCAA	7.463	7.367	7.567	801.050	750.000	6.8
DICAMBA	7.649	7.553	7.753	781.610	705.000	10.9
DICHLORPROP	8.342	8.248	8.448	757.570	705.000	7.5
Dinoseb	10.870	10.781	10.981	758.310	705.000	7.6