

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030002.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.140	9.044	9.244	722.200	712.500	1.4
2,4,5-TP(Silvex)	8.858	8.762	8.962	720.000	712.500	1.1
2,4-D	8.015	7.919	8.119	700.150	705.000	-0.7
2,4-DB	9.698	9.603	9.803	748.260	712.500	5.0
2,4-DCAA	6.939	6.842	7.042	735.570	750.000	-1.9
DICAMBA	7.115	7.018	7.218	705.880	705.000	0.1
DICHLORPROP	7.796	7.699	7.899	687.850	705.000	-2.4
Dinoseb	10.860	10.766	10.966	690.110	705.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030002.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.938	9.847	10.047	770.350	712.500	8.1
2,4,5-TP(Silvex)	9.534	9.442	9.642	779.550	712.500	9.4
2,4-D	8.655	8.561	8.761	741.780	705.000	5.2
2,4-DB	10.496	10.407	10.607	739.380	712.500	3.8
2,4-DCAA	7.462	7.367	7.567	794.140	750.000	5.9
DICAMBA	7.648	7.553	7.753	780.920	705.000	10.8
DICHLORPROP	8.342	8.248	8.448	760.750	705.000	7.9
Dinoseb	10.870	10.781	10.981	739.550	705.000	4.9