

### CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030013.D Time Analyzed: 12:12

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|--------|----------------------|--------|--------------------|-------------------|-----|
| 2,4,5-T          | 9.140  | 9.044                | 9.244  | 745.600            | 712.500           | 4.6 |
| 2,4,5-TP(Silvex) | 8.858  | 8.762                | 8.962  | 742.010            | 712.500           | 4.1 |
| 2,4-D            | 8.015  | 7.919                | 8.119  | 720.060            | 705.000           | 2.1 |
| 2,4-DB           | 9.698  | 9.603                | 9.803  | 763.740            | 712.500           | 7.2 |
| 2,4-DCAA         | 6.939  | 6.842                | 7.042  | 760.450            | 750.000           | 1.4 |
| DICAMBA          | 7.116  | 7.018                | 7.218  | 728.740            | 705.000           | 3.4 |
| DICHLORPROP      | 7.796  | 7.699                | 7.899  | 708.330            | 705.000           | 0.5 |
| Dinoseb          | 10.861 | 10.766               | 10.966 | 719.310            | 705.000           | 2.0 |

### CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030013.D Time Analyzed: 12:12

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 9.937  | 9.847                | 10.047 | 784.080            | 712.500           | 10.0 |
| 2,4,5-TP(Silvex) | 9.533  | 9.442                | 9.642  | 796.690            | 712.500           | 11.8 |
| 2,4-D            | 8.654  | 8.561                | 8.761  | 757.120            | 705.000           | 7.4  |
| 2,4-DB           | 10.495 | 10.407               | 10.607 | 757.220            | 712.500           | 6.3  |
| 2,4-DCAA         | 7.462  | 7.367                | 7.567  | 820.590            | 750.000           | 9.4  |
| DICAMBA          | 7.648  | 7.553                | 7.753  | 800.730            | 705.000           | 13.6 |
| DICHLORPROP      | 8.341  | 8.248                | 8.448  | 768.250            | 705.000           | 9.0  |
| Dinoseb          | 10.869 | 10.781               | 10.981 | 755.960            | 705.000           | 7.2  |