

Surrogate Summary

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029918.D	PIBLK-PS029918.D	2,4-DCAA	1	500	433	87		39	175
		2,4-DCAA	2	500	433	87		39	175
I.BLK-PS029990.D	PIBLK-PS029990.D	2,4-DCAA	1	500	525	105		39	175
		2,4-DCAA	2	500	558	112		39	175
Q1906-05MS	WC-5MS	2,4-DCAA	1	500	956	191	*	10	141
		2,4-DCAA	2	500	1520	303	*	10	141
Q1906-05MSD	WC-5MSD	2,4-DCAA	1	500	969	194	*	10	141
		2,4-DCAA	2	500	1540	308	*	10	141
I.BLK-PS030001.D	PIBLK-PS030001.D	2,4-DCAA	1	500	518	104		39	175
		2,4-DCAA	2	500	565	113		39	175
Q1907-01	CO-8R-WC	2,4-DCAA	1	500	264	53		10	141
		2,4-DCAA	2	500	294	59		10	141
PB167796BL	PB167796BL	2,4-DCAA	1	500	433	87		10	141
		2,4-DCAA	2	500	413	83		10	141
I.BLK-PS030012.D	PIBLK-PS030012.D	2,4-DCAA	1	500	551	110		39	175
		2,4-DCAA	2	500	588	118		39	175
I.BLK-PS030028.D	PIBLK-PS030028.D	2,4-DCAA	1	500	530	106		39	175
		2,4-DCAA	2	500	553	111		39	175
PB167796BS	PB167796BS	2,4-DCAA	1	500	530	106		10	141
		2,4-DCAA	2	500	540	108		10	141
I.BLK-PS030040.D	PIBLK-PS030040.D	2,4-DCAA	1	500	529	106		39	175
		2,4-DCAA	2	500	552	110		39	175