

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM **Case No.:** Q1907 **SAS No.:** Q1907 **SDG NO.:** Q1907

Instrument ID: ECD_P **Date(s) Analyzed:** 04/22/2025 04/22/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.72	4.62	4.82	24116800
		2	4.80	4.70	4.90	18016900
		3	4.88	4.78	4.98	56611200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	44912000
		2	5.40	5.30	5.50	23013000
		3	5.69	5.59	5.79	47217600
		4	5.85	5.75	5.95	23179400
		5	5.94	5.84	6.04	16145700
Aroclor-1262	500	1	8.10	8.00	8.20	144491000
		2	8.42	8.32	8.52	274186000
		3	8.74	8.64	8.84	182878000
		4	8.82	8.72	8.92	134211000
		5	9.48	9.38	9.58	88162000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: WALS01

Lab Code: CHEM **Case No.:** Q1907 **SAS No.:** Q1907 **SDG NO.:** Q1907

Instrument ID: ECD_P **Date(s) Analyzed:** 04/22/2025 04/22/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	21995000
		2	4.11	4.01	4.21	16561700
		3	4.19	4.09	4.29	48477000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	38317800
		2	4.92	4.82	5.02	38187400
		3	5.09	4.99	5.19	21175200
		4	5.18	5.08	5.28	18540400
		5	5.35	5.25	5.45	19901900
Aroclor-1262	500	1	6.94	6.84	7.04	113612000
		2	7.20	7.10	7.30	92339400
		3	7.72	7.62	7.82	75694400
		4	7.79	7.69	7.89	124826000
		5	8.29	8.19	8.39	55493200