

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 04/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP071650.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.662	5.571	5.771	520.080	500.000	4.0
Aroclor-1016-2	5.683	5.593	5.793	514.850	500.000	3.0
Aroclor-1016-3	5.746	5.654	5.854	510.100	500.000	2.0
Aroclor-1016-4	5.843	5.752	5.952	517.690	500.000	3.5
Aroclor-1016-5	6.136	6.045	6.245	518.100	500.000	3.6
Aroclor-1260-1	7.255	7.165	7.365	514.390	500.000	2.9
Aroclor-1260-2	7.508	7.418	7.618	505.470	500.000	1.1
Aroclor-1260-3	7.867	7.777	7.977	530.650	500.000	6.1
Aroclor-1260-4	8.092	8.002	8.202	517.190	500.000	3.4
Aroclor-1260-5	8.411	8.322	8.522	531.830	500.000	6.4
Decachlorobiphenyl	10.225	10.137	10.337	54.080	50.000	8.2
Tetrachloro-m-xylene	4.509	4.417	4.617	51.770	50.000	3.5

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 04/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP071650.D Time Analyzed: 16:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	547.990	500.000	9.6
Aroclor-1016-2	4.911	4.816	5.016	546.060	500.000	9.2
Aroclor-1016-3	5.088	4.993	5.193	558.800	500.000	11.8
Aroclor-1016-4	5.130	5.036	5.236	557.840	500.000	11.6
Aroclor-1016-5	5.344	5.250	5.450	553.040	500.000	10.6
Aroclor-1260-1	6.379	6.285	6.485	543.610	500.000	8.7
Aroclor-1260-2	6.567	6.473	6.673	540.470	500.000	8.1
Aroclor-1260-3	6.720	6.627	6.827	543.100	500.000	8.6
Aroclor-1260-4	7.191	7.098	7.298	539.760	500.000	8.0
Aroclor-1260-5	7.433	7.339	7.539	541.090	500.000	8.2
Decachlorobiphenyl	8.840	8.749	8.949	54.550	50.000	9.1
Tetrachloro-m-xylene	3.808	3.712	3.912	55.400	50.000	10.8