

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP071665.D Time Analyzed: 21:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.666	5.571	5.771	503.260	500.000	0.7
Aroclor-1016-2	5.688	5.593	5.793	504.310	500.000	0.9
Aroclor-1016-3	5.750	5.654	5.854	502.650	500.000	0.5
Aroclor-1016-4	5.847	5.752	5.952	505.440	500.000	1.1
Aroclor-1016-5	6.140	6.045	6.245	494.380	500.000	-1.1
Aroclor-1260-1	7.259	7.165	7.365	485.270	500.000	-2.9
Aroclor-1260-2	7.513	7.418	7.618	475.760	500.000	-4.8
Aroclor-1260-3	7.871	7.777	7.977	492.240	500.000	-1.6
Aroclor-1260-4	8.096	8.002	8.202	472.680	500.000	-5.5
Aroclor-1260-5	8.414	8.322	8.522	498.970	500.000	-0.2
Decachlorobiphenyl	10.229	10.137	10.337	51.660	50.000	3.3
Tetrachloro-m-xylene	4.513	4.417	4.617	50.510	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP071665.D Time Analyzed: 21:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.893	4.798	4.998	487.020	500.000	-2.6
Aroclor-1016-2	4.911	4.816	5.016	495.010	500.000	-1.0
Aroclor-1016-3	5.088	4.993	5.193	502.210	500.000	0.4
Aroclor-1016-4	5.130	5.036	5.236	504.360	500.000	0.9
Aroclor-1016-5	5.344	5.250	5.450	527.960	500.000	5.6
Aroclor-1260-1	6.378	6.285	6.485	500.380	500.000	0.1
Aroclor-1260-2	6.567	6.473	6.673	495.690	500.000	-0.9
Aroclor-1260-3	6.720	6.627	6.827	486.150	500.000	-2.8
Aroclor-1260-4	7.191	7.098	7.298	501.720	500.000	0.3
Aroclor-1260-5	7.432	7.339	7.539	490.810	500.000	-1.8
Decachlorobiphenyl	8.840	8.749	8.949	51.180	50.000	2.4
Tetrachloro-m-xylene	3.809	3.712	3.912	49.930	50.000	-0.1