

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD088351.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.606	6.806	59.640	50.000	19.3
4,4'-DDE	6.197	6.097	6.297	54.940	50.000	9.9
4,4'-DDT	7.022	6.922	7.122	54.540	50.000	9.1
Aldrin	5.273	5.173	5.373	57.040	50.000	14.1
alpha-BHC	4.000	3.901	4.101	56.560	50.000	13.1
alpha-Chlordane	6.026	5.929	6.129	57.020	50.000	14.0
beta-BHC	4.516	4.416	4.616	55.050	50.000	10.1
Decachlorobiphenyl	9.074	8.975	9.175	51.210	50.000	2.4
delta-BHC	4.763	4.665	4.865	57.000	50.000	14.0
Dieldrin	6.348	6.249	6.449	57.500	50.000	15.0
Endosulfan I	6.075	5.976	6.176	57.810	50.000	15.6
Endosulfan II	6.787	6.687	6.887	55.820	50.000	11.6
Endosulfan sulfate	7.150	7.050	7.250	55.050	50.000	10.1
Endrin	6.575	6.476	6.676	56.260	50.000	12.5
Endrin aldehyde	6.916	6.816	7.016	53.480	50.000	7.0
Endrin ketone	7.631	7.531	7.731	55.000	50.000	10.0
gamma-BHC (Lindane)	4.331	4.232	4.432	56.950	50.000	13.9
gamma-Chlordane	5.946	5.847	6.047	55.980	50.000	12.0
Heptachlor	4.931	4.831	5.031	55.920	50.000	11.8
Heptachlor epoxide	5.692	5.592	5.792	58.640	50.000	17.3
Methoxychlor	7.494	7.395	7.595	50.200	50.000	0.4
Tetrachloro-m-xylene	3.551	3.452	3.652	53.840	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD088351.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	49.490	50.000	-1.0
4,4'-DDE	5.376	5.280	5.480	48.470	50.000	-3.1
4,4'-DDT	6.187	6.088	6.288	47.090	50.000	-5.8
Aldrin	4.370	4.273	4.473	49.060	50.000	-1.9
alpha-BHC	3.395	3.296	3.496	48.680	50.000	-2.6
alpha-Chlordane	5.193	5.094	5.294	48.330	50.000	-3.3
beta-BHC	4.027	3.928	4.128	48.520	50.000	-3.0
Decachlorobiphenyl	8.077	7.977	8.177	45.030	50.000	-9.9
delta-BHC	4.264	4.165	4.365	49.540	50.000	-0.9
Dieldrin	5.516	5.417	5.617	48.950	50.000	-2.1
Endosulfan I	5.249	5.151	5.351	40.760	50.000	-18.5
Endosulfan II	6.083	5.984	6.184	48.410	50.000	-3.2
Endosulfan sulfate	6.485	6.386	6.586	47.170	50.000	-5.7
Endrin	5.792	5.693	5.893	49.730	50.000	-0.5
Endrin aldehyde	6.261	6.163	6.363	46.110	50.000	-7.8
Endrin ketone	6.995	6.895	7.095	46.820	50.000	-6.4
gamma-BHC (Lindane)	3.731	3.633	3.833	48.270	50.000	-3.5
gamma-Chlordane	5.128	5.030	5.230	48.800	50.000	-2.4
Heptachlor	4.085	3.986	4.186	48.090	50.000	-3.8
Heptachlor epoxide	4.875	4.777	4.977	48.700	50.000	-2.6
Methoxychlor	6.758	6.659	6.859	45.340	50.000	-9.3
Tetrachloro-m-xylene	2.881	2.783	2.983	47.020	50.000	-6.0