

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD088362.D Time Analyzed: 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	57.840	50.000	15.7
4,4'-DDE	6.197	6.097	6.297	53.490	50.000	7.0
4,4'-DDT	7.022	6.922	7.122	47.800	50.000	-4.4
Aldrin	5.273	5.173	5.373	57.410	50.000	14.8
alpha-BHC	4.000	3.901	4.101	58.390	50.000	16.8
alpha-Chlordane	6.028	5.929	6.129	56.390	50.000	12.8
beta-BHC	4.516	4.416	4.616	56.290	50.000	12.6
Decachlorobiphenyl	9.075	8.975	9.175	49.980	50.000	0.0
delta-BHC	4.764	4.665	4.865	57.590	50.000	15.2
Dieldrin	6.349	6.249	6.449	56.000	50.000	12.0
Endosulfan I	6.076	5.976	6.176	56.320	50.000	12.6
Endosulfan II	6.787	6.687	6.887	53.510	50.000	7.0
Endosulfan sulfate	7.150	7.050	7.250	52.660	50.000	5.3
Endrin	6.576	6.476	6.676	54.200	50.000	8.4
Endrin aldehyde	6.916	6.816	7.016	51.240	50.000	2.5
Endrin ketone	7.631	7.531	7.731	51.970	50.000	3.9
gamma-BHC (Lindane)	4.331	4.232	4.432	57.410	50.000	14.8
gamma-Chlordane	5.946	5.847	6.047	54.870	50.000	9.7
Heptachlor	4.931	4.831	5.031	55.920	50.000	11.8
Heptachlor epoxide	5.692	5.592	5.792	55.530	50.000	11.1
Methoxychlor	7.494	7.395	7.595	46.280	50.000	-7.4
Tetrachloro-m-xylene	3.551	3.452	3.652	55.270	50.000	10.5

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD088362.D Time Analyzed: 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	49.270	50.000	-1.5
4,4'-DDE	5.377	5.280	5.480	48.770	50.000	-2.5
4,4'-DDT	6.186	6.088	6.288	44.410	50.000	-11.2
Aldrin	4.370	4.273	4.473	50.520	50.000	1.0
alpha-BHC	3.395	3.296	3.496	51.110	50.000	2.2
alpha-Chlordane	5.193	5.094	5.294	48.900	50.000	-2.2
beta-BHC	4.027	3.928	4.128	50.090	50.000	0.2
Decachlorobiphenyl	8.076	7.977	8.177	40.110	50.000	-19.8
delta-BHC	4.264	4.165	4.365	50.700	50.000	1.4
Dieldrin	5.515	5.417	5.617	49.010	50.000	-2.0
Endosulfan I	5.249	5.151	5.351	41.280	50.000	-17.4
Endosulfan II	6.083	5.984	6.184	47.640	50.000	-4.7
Endosulfan sulfate	6.485	6.386	6.586	45.930	50.000	-8.1
Endrin	5.792	5.693	5.893	49.380	50.000	-1.2
Endrin aldehyde	6.261	6.163	6.363	45.320	50.000	-9.4
Endrin ketone	6.994	6.895	7.095	44.690	50.000	-10.6
gamma-BHC (Lindane)	3.731	3.633	3.833	50.100	50.000	0.2
gamma-Chlordane	5.128	5.030	5.230	49.460	50.000	-1.1
Heptachlor	4.085	3.986	4.186	48.710	50.000	-2.6
Heptachlor epoxide	4.875	4.777	4.977	49.540	50.000	-0.9
Methoxychlor	6.757	6.659	6.859	42.800	50.000	-14.4
Tetrachloro-m-xylene	2.882	2.783	2.983	50.470	50.000	0.9