

Surrogate Summary

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	22.7	113		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
I.BLK-PD088350.D	PIBLK-PD088350.D	Decachlorobiphenyl	1	20	20.0	100		43	140
		Tetrachloro-m-xylene	1	20	19.7	99		77	126
		Decachlorobiphenyl	2	20	17.9	89		43	140
		Tetrachloro-m-xylene	2	20	18.9	95		77	126
PB167795BL	PB167795BL	Decachlorobiphenyl	1	20	18.6	93		20	144
		Tetrachloro-m-xylene	1	20	17.6	88		19	148
		Decachlorobiphenyl	2	20	16.4	82		20	144
		Tetrachloro-m-xylene	2	20	17.6	88		19	148
PB167795BS	PB167795BS	Decachlorobiphenyl	1	20	20.9	104		20	144
		Tetrachloro-m-xylene	1	20	19.9	100		19	148
		Decachlorobiphenyl	2	20	18.9	94		20	144
		Tetrachloro-m-xylene	2	20	20.0	100		19	148
Q1907-01	CO-8R-WC	Decachlorobiphenyl	1	20	12.5	63		20	144
		Tetrachloro-m-xylene	1	20	12.3	61		19	148
		Decachlorobiphenyl	2	20	9.39	47		20	144
		Tetrachloro-m-xylene	2	20	12.2	61		19	148
Q1912-01MS	MH-EMS	Decachlorobiphenyl	1	20	28.2	141		20	144
		Tetrachloro-m-xylene	1	20	18.7	93		19	148
		Decachlorobiphenyl	2	20	15.0	75		20	144
		Tetrachloro-m-xylene	2	20	18.9	95		19	148
Q1912-01MSD	MH-EMSD	Decachlorobiphenyl	1	20	25.0	125		20	144
		Tetrachloro-m-xylene	1	20	18.8	94		19	148
		Decachlorobiphenyl	2	20	15.1	75		20	144
		Tetrachloro-m-xylene	2	20	18.9	95		19	148
I.BLK-PD088361.D	PIBLK-PD088361.D	Decachlorobiphenyl	1	20	19.2	96		43	140
		Tetrachloro-m-xylene	1	20	19.7	98		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126