

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

DataFile : PD088359.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1912-01MSD	MH-EMSD											
	alpha-BHC	17.99	0	18.3	ug/kg	102		0		60	144	20
	beta-BHC	17.99	0	16.4	ug/kg	91		1		54	143	20
	delta-BHC	17.99	0	18.1	ug/kg	101		1		29	151	20
	gamma-BHC (Lindane)	17.99	0	18.9	ug/kg	105		0		61	140	20
	Heptachlor	17.99	0	17.4	ug/kg	97		1		63	135	20
	Aldrin	17.99	0	18.6	ug/kg	103		0		49	139	20
	Heptachlor epoxide	17.99	0	17.4	ug/kg	97		1		41	156	20
	Endosulfan I	17.99	0	20.5	ug/kg	114		0		56	142	20
	Dieldrin	17.99	1.9	17.8	ug/kg	88		1		47	161	20
	4,4'-DDE	17.99	0	14.4	ug/kg	80		0		55	136	20
	Endrin	17.99	1.4	18.1	ug/kg	93		0		57	139	20
	Endosulfan II	17.99	0	15.9	ug/kg	88		1		40	163	20
	4,4'-DDD	17.99	0	19.2	ug/kg	107		0		47	163	20
	Endosulfan sulfate	17.99	0.2342	17.4	ug/kg	95		0		62	139	20
	4,4'-DDT	17.99	0.7385	15.5	ug/kg	82		1		51	146	20
	Methoxychlor	17.99	0	13.8	ug/kg	77		5		54	136	20
	Endrin ketone	17.99	0	17.1	ug/kg	95		0		60	129	20
	Endrin aldehyde	17.99	0.897	18.5	ug/kg	98		1		59	132	20
	alpha-Chlordane	17.99	0	18.2	ug/kg	101		0		39	166	20
	gamma-Chlordane	17.99	0	18.2	ug/kg	101		1		44	175	20