

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: **8081B** Datafile : PD088353.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167795BS	alpha-BHC	16.66	17.2	ug/kg	103				84	123	
	beta-BHC	16.66	17.0	ug/kg	102				82	123	
	delta-BHC	16.66	18.1	ug/kg	109				83	126	
	gamma-BHC (Lindane)	16.66	17.2	ug/kg	103				83	125	
	Heptachlor	16.66	17.4	ug/kg	104				83	122	
	Aldrin	16.66	17.6	ug/kg	106				82	124	
	Heptachlor epoxide	16.66	17.6	ug/kg	106				83	120	
	Endosulfan I	16.66	17.7	ug/kg	106				81	124	
	Dieldrin	16.66	18.0	ug/kg	108				85	121	
	4,4'-DDE	16.66	17.4	ug/kg	104				81	123	
	Endrin	16.66	17.3	ug/kg	104				76	130	
	Endosulfan II	16.66	17.2	ug/kg	103				80	125	
	4,4'-DDD	16.66	18.0	ug/kg	108				80	131	
	Endosulfan sulfate	16.66	17.6	ug/kg	106				81	122	
	4,4'-DDT	16.66	16.3	ug/kg	98				70	129	
	Methoxychlor	16.66	16.2	ug/kg	97				60	119	
	Endrin ketone	16.66	17.7	ug/kg	106				77	132	
	Endrin aldehyde	16.66	17.8	ug/kg	107				79	124	
	alpha-Chlordane	16.66	17.5	ug/kg	105				84	120	
	gamma-Chlordane	16.66	17.5	ug/kg	105				83	122	