

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01
Lab Code: CHEM **Case No.:** Q1907 **SAS No.:** Q1907 **SDG NO.:** Q1907
Instrument ID: ECD_S **Calibration Date(s):** 04/23/2025 04/23/2025
Calibration Times: 11:01 12:37

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS029919.D</u>	CF 500 = <u>PS029920.D</u>				
CF 750 = <u>PS029921.D</u>		CF 1000 = <u>PS029922.D</u>	CF 1500 = <u>PS029923.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	16213200000	13862200000	13431300000	12910500000	12897000000	13862800000	10
2,4-D	3379530000	2797950000	2695740000	2590800000	2600340000	2812870000	12
2,4-DCAA	3038250000	2463560000	2343090000	2233670000	2226830000	2461080000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01
Lab Code: CHEM **Case No.:** Q1907 **SAS No.:** Q1907 **SDG NO.:** Q1907
Instrument ID: ECD_S **Calibration Date(s):** 04/23/2025 04/23/2025
Calibration Times: 11:01 12:37

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029919.D</u>	CF 500 =	<u>PS029920.D</u>		
CF 750 =		<u>PS029921.D</u>	CF 1000 =	<u>PS029922.D</u>	CF 1500 =	<u>PS029923.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	8841790000	8081200000	8077540000	7871670000	8036560000	8181750000	5
2,4-D	1299090000	1113240000	1091000000	1054800000	1077400000	1127110000	9
2,4-DCAA	807639000	692361000	682858000	661991000	677736000	704517000	8