

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

DataFile : PS030063.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1901-08MS	B-167-SB01MS											
	2,4-D	50	0	64.2	ug/L	128				65	135	
	2,4,5-TP(Silvex)	50	0	63.4	ug/L	127				62	139	