

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 12:09 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.01
Methoxychlor	7.49	7.50	7.40	7.60	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 12:09 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.01
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.09	3.99	4.19	0.01
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00