

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088384.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.077	8.975	9.175	47.420	50.000	-5.2
Endrin	6.577	6.476	6.676	52.460	50.000	4.9
gamma-BHC (Lindane)	4.333	4.232	4.432	53.820	50.000	7.6
Heptachlor	4.933	4.831	5.031	53.190	50.000	6.4
Heptachlor epoxide	5.694	5.592	5.792	53.690	50.000	7.4
Methoxychlor	7.496	7.395	7.595	47.380	50.000	-5.2
Tetrachloro-m-xylene	3.553	3.452	3.652	52.240	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088384.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	43.900	50.000	-12.2
Endrin	5.791	5.693	5.893	47.310	50.000	-5.4
gamma-BHC (Lindane)	3.731	3.633	3.833	48.110	50.000	-3.8
Heptachlor	4.085	3.986	4.186	46.890	50.000	-6.2
Heptachlor epoxide	4.875	4.777	4.977	48.060	50.000	-3.9
Methoxychlor	6.757	6.659	6.859	44.580	50.000	-10.8
Tetrachloro-m-xylene	2.882	2.783	2.983	48.420	50.000	-3.2