

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: WALS01
ProjectID: Walsh CO-032 Sampling
Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG No.: Q1907
DataFile: FG015772.D Analyst Name: YP\AJ Analyst Date: 04-29-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	103719858	122023	121213	0.668