

**Matrix Spike/Matrix Spike Duplicate Summary**

**SW-846**

**SDG No.:** Q1907

**Client:** Walsh Construction Company II, LLC

**Analytical Method:** SW8270E

| Parameter             | Spike             | Sample<br>Result         | Result              | Units | Rec | Rec<br>Qual | RPD | RPD<br>Qual      | Low               | Limits<br>High | RPD |
|-----------------------|-------------------|--------------------------|---------------------|-------|-----|-------------|-----|------------------|-------------------|----------------|-----|
| <b>Lab Sample ID:</b> | <b>Q1901-08MS</b> | <b>Client Sample ID:</b> | <b>B-167-SB01MS</b> |       |     |             |     | <b>DataFile:</b> | <b>BF142277.D</b> |                |     |
| Pyridine              | 500               | 24.1                     | 270                 | ug/L  | 49  |             |     |                  | 10                | 109            |     |
| 1,4-Dichlorobenzene   | 500               | 0                        | 340                 | ug/L  | 68  |             |     |                  | 55                | 125            |     |
| 2-Methylphenol        | 500               | 0                        | 360                 | ug/L  | 72  |             |     |                  | 60                | 131            |     |
| 3+4-Methylphenols     | 500               | 0                        | 360                 | ug/L  | 72  |             |     |                  | 54                | 136            |     |
| Hexachloroethane      | 500               | 0                        | 350                 | ug/L  | 70  |             |     |                  | 19                | 146            |     |
| Nitrobenzene          | 500               | 0                        | 440                 | ug/L  | 88  |             |     |                  | 62                | 112            |     |
| Hexachlorobutadiene   | 500               | 0                        | 380                 | ug/L  | 76  |             |     |                  | 52                | 125            |     |
| 2,4,6-Trichlorophenol | 500               | 0                        | 470                 | ug/L  | 94  |             |     |                  | 78                | 112            |     |
| 2,4,5-Trichlorophenol | 500               | 0                        | 460                 | ug/L  | 92  |             |     |                  | 71                | 111            |     |
| 2,4-Dinitrotoluene    | 500               | 0                        | 500                 | ug/L  | 100 |             |     |                  | 74                | 137            |     |
| Hexachlorobenzene     | 500               | 0                        | 440                 | ug/L  | 88  |             |     |                  | 72                | 115            |     |
| Pentachlorophenol     | 1000              | 0                        | 940                 | ug/L  | 94  |             |     |                  | 52                | 162            |     |