

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VX045989.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0430WBSD01	Vinyl chloride	20	16.6	ug/L	83	4		65	117	20
	1,1-Dichloroethene	20	18.5	ug/L	93	0		74	110	20
	2-Butanone	100	120	ug/L	120	9		65	122	20
	Carbon Tetrachloride	20	21.8	ug/L	109	0		77	113	20
	Chloroform	20	21.0	ug/L	105	3		79	113	20
	Benzene	20	20.1	ug/L	101	0		82	109	20
	1,2-Dichloroethane	20	22.6	ug/L	113	1		80	115	20
	Trichloroethene	20	20.2	ug/L	101	1		77	113	20
	Tetrachloroethene	20	20.1	ug/L	101	6		67	123	20
	Chlorobenzene	20	20.8	ug/L	104	2		82	109	20