

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907
Lab File ID: VX045524.D BFB Injection Date: 04/01/2025
Instrument ID: MSVOA_X BFB Injection Time: 16:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.2 (7.8) 1
176	95.0 - 101.0% of mass 174	65.3 (97.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045525.D	04/01/2025	17:06
VSTDIC005	VSTDIC005	VX045526.D	04/01/2025	17:29
VSTDIC020	VSTDIC020	VX045527.D	04/01/2025	17:52
VSTDIC050	VSTDIC050	VX045528.D	04/01/2025	18:15
VSTDIC100	VSTDIC100	VX045529.D	04/01/2025	18:38
VSTDIC150	VSTDIC150	VX045530.D	04/01/2025	19:02