

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907
Lab File ID: VX045984.D BFB Injection Date: 04/30/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:32
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1 (1.4) 1
174	50.0 - 100.0% of mass 95	69.5
175	5.0 - 9.0% of mass 174	5.3 (7.7) 1
176	95.0 - 101.0% of mass 174	67.1 (96.5) 1
177	5.0 - 9.0% of mass 176	4.5 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045985.D	04/30/2025	10:01
VX0430WBL01	VX0430WBL01	VX045987.D	04/30/2025	10:55
VX0430WBS01	VX0430WBS01	VX045988.D	04/30/2025	11:19
VX0430WBSD01	VX0430WBSD01	VX045989.D	04/30/2025	11:46
CO-8R-WC	Q1907-02	VX046003.D	04/30/2025	17:13