

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY022072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBS01	Dichlorodifluoromethane	20	19.3	ug/Kg	97			64	136	
	Chloromethane	20	22.2	ug/Kg	111			70	130	
	Vinyl chloride	20	20.7	ug/Kg	104			72	129	
	Bromomethane	20	22.8	ug/Kg	114			58	141	
	Chloroethane	20	21.2	ug/Kg	106			69	130	
	Trichlorofluoromethane	20	21.8	ug/Kg	109			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102			81	123	
	1,1-Dichloroethene	20	19.7	ug/Kg	99			79	121	
	Acetone	100	94.5	ug/Kg	95			60	131	
	Carbon disulfide	20	19.9	ug/Kg	100			45	154	
	Methyl tert-butyl Ether	20	19.0	ug/Kg	95			77	129	
	Methyl Acetate	20	20.3	ug/Kg	102			69	149	
	Methylene Chloride	20	20.0	ug/Kg	100			56	174	
	trans-1,2-Dichloroethene	20	19.5	ug/Kg	98			80	123	
	1,1-Dichloroethane	20	19.9	ug/Kg	100			82	123	
	Cyclohexane	20	18.9	ug/Kg	95			76	122	
	2-Butanone	100	92.7	ug/Kg	93			69	131	
	Carbon Tetrachloride	20	21.3	ug/Kg	106			76	129	
	cis-1,2-Dichloroethene	20	19.5	ug/Kg	98			82	123	
	Bromochloromethane	20	20.0	ug/Kg	100			80	127	
	Chloroform	20	20.4	ug/Kg	102			82	125	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			80	126	
	Methylcyclohexane	20	19.2	ug/Kg	96			77	123	
	Benzene	20	20.6	ug/Kg	103			84	121	
	1,2-Dichloroethane	20	21.2	ug/Kg	106			81	126	
	Trichloroethene	20	20.9	ug/Kg	104			83	122	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			83	122	
	Bromodichloromethane	20	20.9	ug/Kg	104			82	123	
	4-Methyl-2-Pentanone	100	100	ug/Kg	100			70	135	
	Toluene	20	21.0	ug/Kg	105			83	122	
	t-1,3-Dichloropropene	20	20.6	ug/Kg	103			78	124	
	cis-1,3-Dichloropropene	20	20.6	ug/Kg	103			81	122	
	1,1,2-Trichloroethane	20	20.9	ug/Kg	104			82	125	
	2-Hexanone	100	99.9	ug/Kg	100			66	138	
	Dibromochloromethane	20	21.1	ug/Kg	106			79	125	
	1,2-Dibromoethane	20	20.6	ug/Kg	103			80	125	
	Tetrachloroethene	20	21.1	ug/Kg	106			83	125	
	Chlorobenzene	20	20.1	ug/Kg	101			84	122	
	Ethyl Benzene	20	19.3	ug/Kg	97			82	124	
	m/p-Xylenes	40	40.6	ug/Kg	102			83	124	
	o-Xylene	20	19.7	ug/Kg	99			83	123	
	Styrene	20	20.0	ug/Kg	100			82	124	
	Bromoform	20	20.9	ug/Kg	104			75	127	
	Isopropylbenzene	20	18.8	ug/Kg	94			82	124	
	1,1,2,2-Tetrachloroethane	20	19.7	ug/Kg	99			77	127	
	1,3-Dichlorobenzene	20	19.9	ug/Kg	100			83	122	
	1,4-Dichlorobenzene	20	19.7	ug/Kg	99			84	121	
	1,2-Dichlorobenzene	20	20.1	ug/Kg	101			83	124	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY022072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBS01	1,2-Dibromo-3-Chloropropane	20	18.2	ug/Kg	91			66	134	
	1,2,4-Trichlorobenzene	20	19.2	ug/Kg	96			78	127	
	1,2,3-Trichlorobenzene	20	19.7	ug/Kg	99			70	137	