

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY022073.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBSD01	Dichlorodifluoromethane	20	20.3	ug/Kg	102	5		64	136	20
	Chloromethane	20	19.5	ug/Kg	98	12		70	130	20
	Vinyl chloride	20	19.7	ug/Kg	99	5		72	129	20
	Bromomethane	20	19.5	ug/Kg	98	15		58	141	20
	Chloroethane	20	19.5	ug/Kg	98	8		69	130	20
	Trichlorofluoromethane	20	20.4	ug/Kg	102	7		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102	0		81	123	20
	1,1-Dichloroethene	20	20.2	ug/Kg	101	2		79	121	20
	Acetone	100	89.4	ug/Kg	89	7		60	131	20
	Carbon disulfide	20	19.7	ug/Kg	99	1		45	154	20
	Methyl tert-butyl Ether	20	18.6	ug/Kg	93	2		77	129	20
	Methyl Acetate	20	21.5	ug/Kg	108	6		69	149	20
	Methylene Chloride	20	20.1	ug/Kg	101	1		56	174	20
	trans-1,2-Dichloroethene	20	20.0	ug/Kg	100	2		80	123	20
	1,1-Dichloroethane	20	19.8	ug/Kg	99	1		82	123	20
	Cyclohexane	20	18.9	ug/Kg	95	0		76	122	20
	2-Butanone	100	91.4	ug/Kg	91	2		69	131	20
	Carbon Tetrachloride	20	20.7	ug/Kg	104	2		76	129	20
	cis-1,2-Dichloroethene	20	19.7	ug/Kg	99	1		82	123	20
	Bromochloromethane	20	20.0	ug/Kg	100	0		80	127	20
	Chloroform	20	20.1	ug/Kg	101	1		82	125	20
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101	0		80	126	20
	Methylcyclohexane	20	19.0	ug/Kg	95	1		77	123	20
	Benzene	20	20.5	ug/Kg	103	0		84	121	20
	1,2-Dichloroethane	20	20.5	ug/Kg	103	3		81	126	20
	Trichloroethene	20	20.5	ug/Kg	103	1		83	122	20
	1,2-Dichloropropane	20	20.4	ug/Kg	102	0		83	122	20
	Bromodichloromethane	20	20.6	ug/Kg	103	1		82	123	20
	4-Methyl-2-Pentanone	100	98.5	ug/Kg	99	1		70	135	20
	Toluene	20	20.5	ug/Kg	103	2		83	122	20
	t-1,3-Dichloropropene	20	20.2	ug/Kg	101	2		78	124	20
	cis-1,3-Dichloropropene	20	20.1	ug/Kg	101	2		81	122	20
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103	1		82	125	20
	2-Hexanone	100	94.3	ug/Kg	94	6		66	138	20
	Dibromochloromethane	20	20.2	ug/Kg	101	5		79	125	20
	1,2-Dibromoethane	20	20.1	ug/Kg	101	2		80	125	20
	Tetrachloroethene	20	21.3	ug/Kg	106	0		83	125	20
	Chlorobenzene	20	20.3	ug/Kg	102	1		84	122	20
	Ethyl Benzene	20	19.4	ug/Kg	97	0		82	124	20
	m/p-Xylenes	40	40.3	ug/Kg	101	1		83	124	20
	o-Xylene	20	19.4	ug/Kg	97	2		83	123	20
	Styrene	20	20.3	ug/Kg	102	2		82	124	20
	Bromoform	20	20.3	ug/Kg	102	2		75	127	20
	Isopropylbenzene	20	19.3	ug/Kg	97	3		82	124	20
	1,1,2,2-Tetrachloroethane	20	19.4	ug/Kg	97	2		77	127	20
	1,3-Dichlorobenzene	20	20.0	ug/Kg	100	0		83	122	20
	1,4-Dichlorobenzene	20	20.0	ug/Kg	100	1		84	121	20
	1,2-Dichlorobenzene	20	20.4	ug/Kg	102	1		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1907

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY022073.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0430SBSD01	1,2-Dibromo-3-Chloropropane	20	18.4	ug/Kg	92	1		66	134	20
	1,2,4-Trichlorobenzene	20	18.8	ug/Kg	94	2		78	127	20
	1,2,3-Trichlorobenzene	20	18.9	ug/Kg	95	4		70	137	20