

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: Q1907 SAS No.: Q1907 SDG NO.: Q1907
 Lab File ID: VY022070.D Date Analyzed: 04/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	299393	7.71	438020	8.62	420486	11.41
UPPER LIMIT	598786	8.207	876040	9.116	840972	11.914
LOWER LIMIT	149697	7.207	219010	8.116	210243	10.914
EPA SAMPLE NO.						
CO-8R-WC	342212	7.71	633894	8.61	560130	11.41
VY0430SBL01	311598	7.71	580894	8.62	511027	11.41
VY0430SBS01	286487	7.71	433702	8.62	405299	11.41
VY0430SBSD01	289313	7.71	439988	8.61	406945	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.