

Duplicate Sample Summary

Lab Sample Id :	Q1908-01DUP	Q1908-01
Client Id :	B-1DUP	B-1
DF :	1	1
Datafile :	VL042454.D	VL042443.D
Anal Date & Time :	04/29/2025 19:57	04/29/2025 13:11

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.11	0.13	16.7
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	2.3	2.4	4.3
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	10.9	11	0.91
Allyl Chloride	0	0	0
Benzene	0	0	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.06	0.06	0

Duplicate Sample Summary

Lab Sample Id :	Q1908-01DUP	Q1908-01
Client Id :	B-1DUP	B-1
DF :	1	1
Datafile :	VL042454.D	VL042443.D
Anal Date & Time :	04/29/2025 19:57	04/29/2025 13:11

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.37	0.41	10.3
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.44	0.49	10.8
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.32	0.33	3.1
m/p-Xylene	0	0	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.57	0.54	5.4
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.05	0.05	0
Tetrahydrofuran	0	0	0
Toluene	0.13	0.15	14.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.22	0.21	4.7
Vinyl Chloride	0	0	0