

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135606

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/30/2025	11:25
2	CAL2	1	Water	NA	NA	20.2	6.99	04/30/2025	11:26
3	CAL3	1	Water	NA	NA	20.3	10.02	04/30/2025	11:28
4	ICV	1	Water	NA	NA	20.2	7.01	04/30/2025	11:30
5	CCV1	1	Water	NA	NA	20.3	2.01	04/30/2025	11:32
6	Q1913-02	1	Water	NA	NA	20.4	11.91	04/30/2025	11:48
7	Q1913-04	1	Water	NA	NA	20.1	12.19	04/30/2025	12:00
8	Q1913-04DUP	1	Water	NA	NA	20.2	12.20	04/30/2025	12:01
9	CCV2	1	Water	NA	NA	20.2	12.02	04/30/2025	12:05

WORKLIST(Hardcopy Internal Chain)

135606

WorkList Name : PH Q1913 W WorkList ID : 189232 Department : Wet-Chemistry Date : 04-30-2025 11:04:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-02	WC-12-A-202504	Water	pH	Cool 4 deg C	AECO02	L21	04/29/2025	9040C
Q1913-04	WC-13-A-202504	Water	pH	Cool 4 deg C	AECO02	L21	04/29/2025	9040C

Date/Time 04/30/25 11:15
 Raw Sample Received by: SB (wpc)
 Raw Sample Relinquished by: JDCSM

Date/Time 04/30/25 13:30
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: SB (wpc)