



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135606

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	04/30/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	04/30/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	04/30/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135608

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	04/30/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	04/30/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135628

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.1	81	101	78-84	05/01/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135649

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	05/02/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	05/02/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	05/02/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135650

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	05/02/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/02/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/02/2025