

Initial and Continuing Calibration Blank Summary

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135608

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025

Initial and Continuing Calibration Blank Summary

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB135650

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0016	0.0025	J	0.00096	0.005	05/02/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	05/02/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.00098	0.0025	J	0.00096	0.005	05/02/2025