


LB135662

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 2.1388e-004
 C= -1.4764e-002
 Rho= 0.9991178
 Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.005	-0.005	46	0.000	46	0				0/11
0.2	0.200	0.202	0.002	1015	0.0 %	1015					1
2.5	2.500	2.403	-0.097	11304	0.0 %	11304					-4
5.0	5.000	5.036	0.036	23614	0.0 %	23614					-1
7.5	7.500	7.788	0.288	36484	0.0 %	36484					4
10.0	10.000	9.776	-0.224	45777	0.0 %	45777					-2

LB135662 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 May 2025 10:55:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	05 May 2025 10:59:09	S	Std
0.2 - 1	502	1015	-	0.2000	7470A	PPB	05 May 2025 11:01:27	S	Std
2.5 - 1	52.5	11304	-	2.5000	7470A	PPB	05 May 2025 11:03:47	S	Std
5.0 - 1	55	23614	-	5.0000	7470A	PPB	05 May 2025 11:06:04	S	Std
7.5 - 1	57.5	36484	-	7.5000	7470A	PPB	05 May 2025 11:08:21	S	Std
10.0 - 1	57.5	45777	-	10.0000	7470A	PPB	05 May 2025 11:13:18	S	Std
ICV114 - 1	ICV114	17467	3.7210	-7470A	PPB	PPB	05 May 2025 11:16:08	U	SMPL
ICB114 - 1	ICB114	-243	-0.0667	-7470A	PPB	PPB	05 May 2025 11:18:24	U	SMPL
CCV71 - 1	CCV71	22760	4.8530	-7470A	PPB	PPB	05 May 2025 11:25:55	U	SMPL
CCB71 - 1	CCB71	-108	-0.0379	-7470A	PPB	PPB	05 May 2025 11:30:48	U	SMPL
CRA - 1	CRA	980	0.1948	-7470A	PPB	PPB	05 May 2025 11:33:04	U	SMPL
HighStd - 1	HighStd	43610	9.3123	-7470A	PPB	PPB	05 May 2025 11:35:19	U	SMPL
ChkStd - 1	ChkStd	34184	7.2963	-7470A	PPB	PPB	05 May 2025 11:37:34	U	SMPL
PB167839BL - 1	PBW	-112	-0.0387	-7470A	PPB	PPB	05 May 2025 11:42:28	U	SMPL
PB167839BS - 1	LCSW	19000	4.0489	-7470A	PPB	PPB	05 May 2025 11:44:43	U	SMPL
Q1916-04 - 1	WC-12	-323	-0.0838	-7470A	PPB	PPB	05 May 2025 11:47:00	U	SMPL
Q1917-04 - 1	MH-JJ	-66	-0.0289	-7470A	PPB	PPB	05 May 2025 11:49:24	U	SMPL
Q1922-04 - 1	MH-R	-22	-0.0195	-7470A	PPB	PPB	05 May 2025 11:51:41	U	SMPL
Q1922-08 - 1	MH-S	-10	-0.0169	-7470A	PPB	PPB	05 May 2025 11:53:58	U	SMPL
Q1925-01 - 1	AUD-25-0068	76	0.0015	-7470A	PPB	PPB	05 May 2025 11:56:14	U	SMPL
CCV72 - 1	CCV72	23957	5.1090	-7470A	PPB	PPB	05 May 2025 11:58:29	U	SMPL
CCB72 - 1	CCB72	2	-0.0143	-7470A	PPB	PPB	05 May 2025 12:03:22	U	SMPL
Q1925-02 - 1	AUD-25-0069	71	0.0004	-7470A	PPB	PPB	05 May 2025 12:05:37	U	SMPL
Q1925-03 - 1	AUD-25-0070	48	-0.0045	-7470A	PPB	PPB	05 May 2025 12:07:52	U	SMPL
Q1928-04 - 1	MH-Q	25	-0.0094	-7470A	PPB	PPB	05 May 2025 12:10:08	U	SMPL
Q1929-03 - 1	WC-A4-02-C	12	-0.0122	-7470A	PPB	PPB	05 May 2025 12:12:24	U	SMPL
Q1929-07 - 1	WC-A1-03-C	-11	-0.0171	-7470A	PPB	PPB	05 May 2025 12:14:40	U	SMPL
Q1929-11 - 1	WC-A1-04-C	53	-0.0034	-7470A	PPB	PPB	05 May 2025 12:16:56	U	SMPL
Q1932-02 - 1	COMP-4	53	-0.0034	-7470A	PPB	PPB	05 May 2025 12:19:13	U	SMPL
Q1932-04 - 1	COMP-5	55	-0.0030	-7470A	PPB	PPB	05 May 2025 12:21:30	U	SMPL
Q1932-06 - 1	COMP-6	65	-0.0009	-7470A	PPB	PPB	05 May 2025 12:23:49	U	SMPL
Q1932-08 - 1	COMP-7	62	-0.0015	-7470A	PPB	PPB	05 May 2025 12:26:06	U	SMPL
CCV73 - 1	CCV73	22286	4.7517	-7470A	PPB	PPB	05 May 2025 12:28:22	U	SMPL
CCB73 - 1	CCB73	19	-0.0107	-7470A	PPB	PPB	05 May 2025 12:38:42	U	SMPL
Q1935-04 - 1	MH-NN	52	-0.0036	-7470A	PPB	PPB	05 May 2025 12:40:57	U	SMPL
Q1935-08 - 1	MH-MM	129	0.0128	-7470A	PPB	PPB	05 May 2025 12:43:13	U	SMPL
Q1935-08DUP - 1	MH-MMDUP	87	0.0038	-7470A	PPB	PPB	05 May 2025 12:45:29	U	SMPL
Q1935-08MS - 1	MH-MMMS	21031	4.4832	-7470A	PPB	PPB	05 May 2025 12:47:45	U	SMPL
Q1935-08MSD - 1	MH-MMMSD	15092	3.2130	-7470A	PPB	PPB	05 May 2025 12:50:01	U	SMPL
PB167866BL - 1	PBW	-373	-0.0945	-7470A	PPB	PPB	05 May 2025 12:52:19	U	SMPL
PB167866BS - 1	LCSW	17471	3.7218	-7470A	PPB	PPB	05 May 2025 12:54:37	U	SMPL
Q1944-02 - 1	CITY-WATER	-387	-0.0975	-7470A	PPB	PPB	05 May 2025 12:56:54	U	SMPL
Q1944-03 - 1	CHILLER-WATER	239	0.0364	-7470A	PPB	PPB	05 May 2025 12:59:13	U	SMPL
Q1945-01 - 1	GB1W	21242	4.5284	-7470A	PPB	PPB	05 May 2025 13:01:29	U	SMPL
CCV74 - 1	CCV74	21313	4.5436	-7470A	PPB	PPB	05 May 2025 13:03:44	U	SMPL
CCB74 - 1	CCB74	-319	-0.0830	-7470A	PPB	PPB	05 May 2025 13:06:03	U	SMPL
Q1945-01DUP - 1	GB1WDUP	21192	4.5177	-7470A	PPB	PPB	05 May 2025 13:17:16	U	SMPL
Q1945-01MS - 1	GB1WMS	30196	6.4434	-7470A	PPB	PPB	05 May 2025 13:19:31	U	SMPL
Q1945-01MSD - 1	GB1WMSD	29844	6.3681	-7470A	PPB	PPB	05 May 2025 13:21:50	U	SMPL
Q1956-07 - 1	FB-05022025	-607	-0.1446	-7470A	PPB	PPB	05 May 2025 13:24:09	U	SMPL
PB167815TB - 1	PB167815TB	-41	-0.0235	-7470A	PPB	PPB	05 May 2025 13:26:32	U	SMPL
Q1935-08LX5 - 1		7	-0.0133	-7470A	PPB	PPB	05 May 2025 13:28:48	U	SMPL
Q1935-08A - 1		18282	3.8953	-7470A	PPB	PPB	05 May 2025 13:31:05	U	SMPL
Q1945-01LX5 - 1		4179	0.8790	-7470A	PPB	PPB	05 May 2025 13:36:48	U	SMPL
Q1945-01A - 1		38182	8.1514	-7470A	PPB	PPB	05 May 2025 13:39:05	U	SMPL
CCV75 - 1	CCV75	25440	5.4262	-7470A	PPB	PPB	05 May 2025 13:41:21	U	SMPL
CCB75 - 1	CCB75	79	0.0021	-7470A	PPB	PPB	05 May 2025 13:43:41	U	SMPL
Q1913-01 - 1		-112	-0.0387	-7470A	PPB	PPB	05 May 2025 13:45:59	U	SMPL
Q1913-03 - 1		-51	-0.0257	-7470A	PPB	PPB	05 May 2025 13:48:15	U	SMPL
PB167774TB - 1	PB167774TB	-98	-0.0357	-7470A	PPB	PPB	05 May 2025 13:50:30	U	SMPL
CCV76 - 1	CCV76	25307	5.3978	-7470A	PPB	PPB	05 May 2025 13:52:46	U	SMPL
CCB76 - 1	CCB76	-171	-0.0513	-7470A	PPB	PPB	05 May 2025 13:55:01	U	SMPL
CCB71 - 1	CCB71	0	-0.0148	-7470A	PPB	PPB	05 May 2025 11:22:56	U	SMPL