

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV112	Mercury	3.81	4.0	95	90 - 110	CV	05/01/2025	11:47	LB135623

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV114	Mercury	3.72	4.0	93	90 - 110	CV	05/05/2025	11:16	LB135662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM SDG No.: Q1913
Contract: AECO02 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV71	Mercury	4.85	5.0	97	90 - 110	CV	05/05/2025	11:25	LB135662
CCV72	Mercury	5.11	5.0	102	90 - 110	CV	05/05/2025	11:58	LB135662
CCV73	Mercury	4.75	5.0	95	90 - 110	CV	05/05/2025	12:28	LB135662
CCV74	Mercury	4.54	5.0	91	90 - 110	CV	05/05/2025	13:03	LB135662
CCV75	Mercury	5.43	5.0	108	90 - 110	CV	05/05/2025	13:41	LB135662
CCV76	Mercury	5.40	5.0	108	90 - 110	CV	05/05/2025	13:52	LB135662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1010	1000	100	90 - 110	P	05/01/2025	13:44	LB135636
	Barium	495	520	95	90 - 110	P	05/01/2025	13:44	LB135636
	Cadmium	508	510	100	90 - 110	P	05/01/2025	13:44	LB135636
	Chromium	506	520	97	90 - 110	P	05/01/2025	13:44	LB135636
	Lead	989	1000	99	90 - 110	P	05/01/2025	13:44	LB135636
	Selenium	1040	1000	104	90 - 110	P	05/01/2025	13:44	LB135636
	Silver	240	250	96	90 - 110	P	05/01/2025	13:44	LB135636



- 2a -

Client:	<u>AECOM</u>	SDG No.:	<u>Q1913</u>				
Contract:	<u>AECO02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1913</u>	SAS No.:	<u>Q1913</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
LLICV01	Arsenic	23.7	20.0	119	80 - 120	P	05/01/2025	14:01	LB135636
	Barium	95.8	100	96	80 - 120	P	05/01/2025	14:01	LB135636
	Cadmium	6.39	6.0	107	80 - 120	P	05/01/2025	14:01	LB135636
	Chromium	10.5	10.0	105	80 - 120	P	05/01/2025	14:01	LB135636
	Lead	13.1	12.0	109	80 - 120	P	05/01/2025	14:01	LB135636
	Selenium	22.9	20.0	114	80 - 120	P	05/01/2025	14:01	LB135636
	Silver	10.2	10.0	102	80 - 120	P	05/01/2025	14:01	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM SDG No.: Q1913
Contract: AECO02 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	15:05	LB135636
	Barium	9650	10000	96	90 - 110	P	05/01/2025	15:05	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	15:05	LB135636
	Chromium	1000	1000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Selenium	5140	5000	103	90 - 110	P	05/01/2025	15:05	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	15:05	LB135636
CCV02	Arsenic	4930	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Barium	10200	10000	102	90 - 110	P	05/01/2025	16:00	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	16:00	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Selenium	4960	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Silver	1220	1250	98	90 - 110	P	05/01/2025	16:00	LB135636
CCV03	Arsenic	5040	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Cadmium	2510	2500	100	90 - 110	P	05/01/2025	17:03	LB135636
	Chromium	988	1000	99	90 - 110	P	05/01/2025	17:03	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Selenium	5060	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Silver	1220	1250	97	90 - 110	P	05/01/2025	17:03	LB135636
CCV04	Arsenic	5160	5000	103	90 - 110	P	05/01/2025	17:49	LB135636
	Barium	10000	10000	100	90 - 110	P	05/01/2025	17:49	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	17:49	LB135636
	Chromium	987	1000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Lead	4950	5000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Selenium	5240	5000	105	90 - 110	P	05/01/2025	17:49	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	17:49	LB135636
CCV05	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	18:36	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	18:36	LB135636
	Chromium	984	1000	98	90 - 110	P	05/01/2025	18:36	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Selenium	5150	5000	103	90 - 110	P	05/01/2025	18:36	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM SDG No.: Q1913
Contract: AECO02 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1230	1250	98	90 - 110	P	05/01/2025	18:36	LB135636
CCV06	Arsenic	4990	5000	100	90 - 110	P	05/01/2025	19:23	LB135636
	Barium	9830	10000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Cadmium	2460	2500	98	90 - 110	P	05/01/2025	19:23	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Lead	4920	5000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Selenium	5030	5000	101	90 - 110	P	05/01/2025	19:23	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	19:23	LB135636
CCV07	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Barium	9790	10000	98	90 - 110	P	05/01/2025	20:11	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	20:11	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Selenium	5120	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	20:11	LB135636
CCV08	Arsenic	5180	5000	104	90 - 110	P	05/01/2025	20:24	LB135636
	Barium	9880	10000	99	90 - 110	P	05/01/2025	20:24	LB135636
	Cadmium	2490	2500	100	90 - 110	P	05/01/2025	20:24	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	20:24	LB135636
	Lead	4990	5000	100	90 - 110	P	05/01/2025	20:24	LB135636
	Selenium	5270	5000	105	90 - 110	P	05/01/2025	20:24	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	20:24	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1060	1000	106	90 - 110	P	05/02/2025	13:47	LB135654
	Barium	518	520	100	90 - 110	P	05/02/2025	13:47	LB135654
	Cadmium	493	510	97	90 - 110	P	05/02/2025	13:47	LB135654
	Chromium	503	520	97	90 - 110	P	05/02/2025	13:47	LB135654
	Lead	966	1000	97	90 - 110	P	05/02/2025	13:47	LB135654
	Selenium	1040	1000	104	90 - 110	P	05/02/2025	13:47	LB135654
	Silver	230	250	92	90 - 110	P	05/02/2025	13:47	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.0	20.0	95	80 - 120	P	05/02/2025	15:03	LB135654
	Barium	81.5	100	82	80 - 120	P	05/02/2025	15:03	LB135654
	Cadmium	5.51	6.0	92	80 - 120	P	05/02/2025	15:03	LB135654
	Chromium	10.1	10.0	101	80 - 120	P	05/02/2025	15:03	LB135654
	Lead	11.1	12.0	93	80 - 120	P	05/02/2025	15:03	LB135654
	Selenium	16.8	20.0	84	80 - 120	P	05/02/2025	15:03	LB135654
	Silver	10.8	10.0	108	80 - 120	P	05/02/2025	15:03	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM SDG No.: Q1913
Contract: AECO02 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5450	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Barium	9540	10000	95	90 - 110	P	05/02/2025	16:01	LB135654
	Cadmium	2440	2500	98	90 - 110	P	05/02/2025	16:01	LB135654
	Chromium	1010	1000	101	90 - 110	P	05/02/2025	16:01	LB135654
	Lead	4820	5000	96	90 - 110	P	05/02/2025	16:01	LB135654
	Selenium	5460	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Silver	1270	1250	102	90 - 110	P	05/02/2025	16:01	LB135654
CCV02	Arsenic	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Barium	9720	10000	97	90 - 110	P	05/02/2025	16:34	LB135654
	Cadmium	2390	2500	96	90 - 110	P	05/02/2025	16:34	LB135654
	Chromium	1000	1000	100	90 - 110	P	05/02/2025	16:34	LB135654
	Lead	4740	5000	95	90 - 110	P	05/02/2025	16:34	LB135654
	Selenium	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Silver	1260	1250	101	90 - 110	P	05/02/2025	16:34	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM SDG No.: Q1913
 Contract: AECO02 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	973	1000	97	90 - 110	P	05/05/2025	14:18	LB135674
	Barium	522	520	100	90 - 110	P	05/05/2025	14:18	LB135674
	Cadmium	501	510	98	90 - 110	P	05/05/2025	14:18	LB135674
	Chromium	508	520	98	90 - 110	P	05/05/2025	14:18	LB135674
	Lead	988	1000	99	90 - 110	P	05/05/2025	14:18	LB135674
	Selenium	1020	1000	102	90 - 110	P	05/05/2025	14:18	LB135674
	Silver	228	250	91	90 - 110	P	05/05/2025	14:18	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.4	20.0	117	80 - 120	P	05/05/2025	14:36	LB135674
	Barium	95.2	100	95	80 - 120	P	05/05/2025	14:36	LB135674
	Cadmium	5.83	6.0	97	80 - 120	P	05/05/2025	14:36	LB135674
	Chromium	10.2	10.0	102	80 - 120	P	05/05/2025	14:36	LB135674
	Lead	10.9	12.0	91	80 - 120	P	05/05/2025	14:36	LB135674
	Selenium	22.6	20.0	113	80 - 120	P	05/05/2025	14:36	LB135674
	Silver	10.2	10.0	102	80 - 120	P	05/05/2025	14:36	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913

Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4930	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Barium	9850	10000	98	90 - 110	P	05/05/2025	15:20	LB135674
	Cadmium	2450	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Chromium	996	1000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Lead	4940	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Selenium	4960	5000	99	90 - 110	P	05/05/2025	15:20	LB135674
	Silver	1240	1250	99	90 - 110	P	05/05/2025	15:20	LB135674
CCV02	Arsenic	4880	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Barium	9880	10000	99	90 - 110	P	05/05/2025	16:07	LB135674
	Cadmium	2410	2500	96	90 - 110	P	05/05/2025	16:07	LB135674
	Chromium	974	1000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Lead	4870	5000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Selenium	4920	5000	98	90 - 110	P	05/05/2025	16:07	LB135674
	Silver	1210	1250	97	90 - 110	P	05/05/2025	16:07	LB135674
CCV03	Arsenic	5050	5000	101	90 - 110	P	05/05/2025	16:56	LB135674
	Barium	9910	10000	99	90 - 110	P	05/05/2025	16:56	LB135674
	Cadmium	2410	2500	96	90 - 110	P	05/05/2025	16:56	LB135674
	Chromium	975	1000	98	90 - 110	P	05/05/2025	16:56	LB135674
	Lead	4880	5000	98	90 - 110	P	05/05/2025	16:56	LB135674
	Selenium	5160	5000	103	90 - 110	P	05/05/2025	16:56	LB135674
	Silver	1230	1250	98	90 - 110	P	05/05/2025	16:56	LB135674
CCV04	Arsenic	4930	5000	99	90 - 110	P	05/05/2025	17:55	LB135674
	Barium	9690	10000	97	90 - 110	P	05/05/2025	17:55	LB135674
	Cadmium	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
	Chromium	977	1000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Lead	4870	5000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Selenium	4980	5000	100	90 - 110	P	05/05/2025	17:55	LB135674
	Silver	1200	1250	96	90 - 110	P	05/05/2025	17:55	LB135674
CCV05	Arsenic	4990	5000	100	90 - 110	P	05/05/2025	18:43	LB135674
	Barium	9570	10000	96	90 - 110	P	05/05/2025	18:43	LB135674
	Cadmium	2420	2500	97	90 - 110	P	05/05/2025	18:43	LB135674
	Chromium	970	1000	97	90 - 110	P	05/05/2025	18:43	LB135674
	Lead	4860	5000	97	90 - 110	P	05/05/2025	18:43	LB135674
	Selenium	5070	5000	101	90 - 110	P	05/05/2025	18:43	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913

Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1220	1250	98	90 - 110	P	05/05/2025	18:43	LB135674
CCV06	Arsenic	4750	5000	95	90 - 110	P	05/05/2025	19:32	LB135674
	Barium	9680	10000	97	90 - 110	P	05/05/2025	19:32	LB135674
	Cadmium	2340	2500	94	90 - 110	P	05/05/2025	19:32	LB135674
	Chromium	940	1000	94	90 - 110	P	05/05/2025	19:32	LB135674
	Lead	4680	5000	94	90 - 110	P	05/05/2025	19:32	LB135674
	Selenium	4800	5000	96	90 - 110	P	05/05/2025	19:32	LB135674
	Silver	1190	1250	95	90 - 110	P	05/05/2025	19:32	LB135674
CCV07	Arsenic	4700	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Barium	9380	10000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Cadmium	2350	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
	Chromium	955	1000	96	90 - 110	P	05/05/2025	20:19	LB135674
	Lead	4710	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Selenium	4700	5000	94	90 - 110	P	05/05/2025	20:19	LB135674
	Silver	1180	1250	95	90 - 110	P	05/05/2025	20:19	LB135674
CCV08	Arsenic	4790	5000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Barium	9390	10000	94	90 - 110	P	05/05/2025	21:07	LB135674
	Cadmium	2380	2500	95	90 - 110	P	05/05/2025	21:07	LB135674
	Chromium	961	1000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Lead	4770	5000	95	90 - 110	P	05/05/2025	21:07	LB135674
	Selenium	4800	5000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Silver	1180	1250	94	90 - 110	P	05/05/2025	21:07	LB135674
CCV09	Arsenic	4740	5000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Barium	9390	10000	94	90 - 110	P	05/05/2025	21:59	LB135674
	Cadmium	2410	2500	97	90 - 110	P	05/05/2025	21:59	LB135674
	Chromium	959	1000	96	90 - 110	P	05/05/2025	21:59	LB135674
	Lead	4810	5000	96	90 - 110	P	05/05/2025	21:59	LB135674
	Selenium	4740	5000	95	90 - 110	P	05/05/2025	21:59	LB135674
	Silver	1190	1250	95	90 - 110	P	05/05/2025	21:59	LB135674
CCV10	Arsenic	4740	5000	95	90 - 110	P	05/05/2025	22:33	LB135674
	Barium	9090	10000	91	90 - 110	P	05/05/2025	22:33	LB135674
	Cadmium	2380	2500	95	90 - 110	P	05/05/2025	22:33	LB135674
	Chromium	967	1000	97	90 - 110	P	05/05/2025	22:33	LB135674
	Lead	4750	5000	95	90 - 110	P	05/05/2025	22:33	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	4730	5000	95	90 - 110	P	05/05/2025	22:33	LB135674
	Silver	1180	1250	94	90 - 110	P	05/05/2025	22:33	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	976	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Barium	531	520	102	90 - 110	P	05/06/2025	14:47	LB135689
	Cadmium	498	510	98	90 - 110	P	05/06/2025	14:47	LB135689
	Chromium	498	520	96	90 - 110	P	05/06/2025	14:47	LB135689
	Lead	975	1000	98	90 - 110	P	05/06/2025	14:47	LB135689
	Selenium	1000	1000	100	90 - 110	P	05/06/2025	14:47	LB135689
	Silver	232	250	93	90 - 110	P	05/06/2025	14:47	LB135689



- 2a -

Client:	<u>AECOM</u>	SDG No.:	<u>Q1913</u>				
Contract:	<u>AECO02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1913</u>	SAS No.:	<u>Q1913</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
LLICV01	Arsenic	18.2	20.0	91	80 - 120	P	05/06/2025	15:22	LB135689
	Barium	91.5	100	92	80 - 120	P	05/06/2025	15:22	LB135689
	Cadmium	5.98	6.0	100	80 - 120	P	05/06/2025	15:22	LB135689
	Chromium	9.31	10.0	93	80 - 120	P	05/06/2025	15:22	LB135689
	Lead	11.1	12.0	92	80 - 120	P	05/06/2025	15:22	LB135689
	Selenium	19.1	20.0	95	80 - 120	P	05/06/2025	15:22	LB135689
	Silver	9.80	10.0	98	80 - 120	P	05/06/2025	15:22	LB135689

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM **SDG No.:** Q1913

Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Barium	9340	10000	93	90 - 110	P	05/06/2025	16:29	LB135689
	Cadmium	2390	2500	95	90 - 110	P	05/06/2025	16:29	LB135689
	Chromium	998	1000	100	90 - 110	P	05/06/2025	16:29	LB135689
	Lead	5030	5000	101	90 - 110	P	05/06/2025	16:29	LB135689
	Selenium	4800	5000	96	90 - 110	P	05/06/2025	16:29	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	16:29	LB135689
CCV02	Arsenic	4730	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Barium	9710	10000	97	90 - 110	P	05/06/2025	17:15	LB135689
	Cadmium	2390	2500	96	90 - 110	P	05/06/2025	17:15	LB135689
	Chromium	962	1000	96	90 - 110	P	05/06/2025	17:15	LB135689
	Lead	4890	5000	98	90 - 110	P	05/06/2025	17:15	LB135689
	Selenium	4750	5000	95	90 - 110	P	05/06/2025	17:15	LB135689
	Silver	1200	1250	96	90 - 110	P	05/06/2025	17:15	LB135689
CCV03	Arsenic	4570	5000	92	90 - 110	P	05/06/2025	18:08	LB135689
	Barium	9450	10000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Cadmium	2340	2500	94	90 - 110	P	05/06/2025	18:08	LB135689
	Chromium	954	1000	95	90 - 110	P	05/06/2025	18:08	LB135689
	Lead	4690	5000	94	90 - 110	P	05/06/2025	18:08	LB135689
	Selenium	4550	5000	91	90 - 110	P	05/06/2025	18:08	LB135689
	Silver	1180	1250	94	90 - 110	P	05/06/2025	18:08	LB135689
CCV04	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	19:16	LB135689
	Barium	9930	10000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Cadmium	2470	2500	99	90 - 110	P	05/06/2025	19:16	LB135689
	Chromium	988	1000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Lead	4970	5000	99	90 - 110	P	05/06/2025	19:16	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	19:16	LB135689
	Silver	1240	1250	99	90 - 110	P	05/06/2025	19:16	LB135689
CCV05	Arsenic	4780	5000	96	90 - 110	P	05/06/2025	20:01	LB135689
	Barium	9940	10000	99	90 - 110	P	05/06/2025	20:01	LB135689
	Cadmium	2460	2500	99	90 - 110	P	05/06/2025	20:01	LB135689
	Chromium	995	1000	100	90 - 110	P	05/06/2025	20:01	LB135689
	Lead	4900	5000	98	90 - 110	P	05/06/2025	20:01	LB135689
	Selenium	4730	5000	95	90 - 110	P	05/06/2025	20:01	LB135689



- 2a -

Client:	<u>AECOM</u>	SDG No.:	<u>Q1913</u>				
Contract:	<u>AECO02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1913</u>	SAS No.:	<u>Q1913</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
CCV05	Silver	1230	1250	99	90 - 110	P	05/06/2025	20:01	LB135689
CCV06	Arsenic	4840	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Barium	9850	10000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Cadmium	2440	2500	98	90 - 110	P	05/06/2025	20:31	LB135689
	Chromium	975	1000	98	90 - 110	P	05/06/2025	20:31	LB135689
	Lead	4870	5000	97	90 - 110	P	05/06/2025	20:31	LB135689
	Selenium	4820	5000	96	90 - 110	P	05/06/2025	20:31	LB135689
	Silver	1210	1250	97	90 - 110	P	05/06/2025	20:31	LB135689