

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Matrix: Water **Sample ID:** Q1913-02 **Client ID:** WC-12-A-202504DUP
Percent Solids for Sample: NA **Duplicate ID** Q1913-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	500	U	500	U			P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	122		113		8		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Matrix: Water **Sample ID:** Q1913-02MS **Client ID:** WC-12-A-202504MSD
Percent Solids for Sample: NA **Duplicate ID** Q1913-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4220		4060		4		P
Barium	ug/L	20	1020		965		6		P
Cadmium	ug/L	20	943		908		4		P
Chromium	ug/L	20	2180		2100		4		P
Lead	ug/L	20	4570		4400		4		P
Mercury	ug/L	20	31.2		31.3		0		CV
Selenium	ug/L	20	10100		9630		5		P
Silver	ug/L	20	385		366		5		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Matrix: Water **Sample ID:** Q1913-03 **Client ID:** WC-13-S-202504DUP
Percent Solids for Sample: NA **Duplicate ID** Q1913-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1170		902		26	*	P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Matrix: Water **Sample ID:** Q1913-03MS **Client ID:** WC-13-S-202504MSD
Percent Solids for Sample: NA **Duplicate ID** Q1913-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3430		3610		5		P
Barium	ug/L	20	1920		2110		9		P
Cadmium	ug/L	20	864		907		5		P
Chromium	ug/L	20	1740		1780		2		P
Lead	ug/L	20	4170		4360		4		P
Selenium	ug/L	20	8390		8850		5		P
Silver	ug/L	20	321		332		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>AECOM</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1913</u>
Contract:	<u>AECO02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1913</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1935-08</u>	Client ID:	<u>MH-MMDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1935-08DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: AECOM **Level:** LOW **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913
Matrix: Water **Sample ID:** Q1935-08MS **Client ID:** MH-MMMSD
Percent Solids for Sample: NA **Duplicate ID** Q1935-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	44.8		32.1		33	*	CV