

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167807BS Mercury	ug/L	4.0	4.14		104	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167808BS							
Arsenic	ug/L	4000	3770		94	80 - 120	P
Barium	ug/L	1000	826		83	80 - 120	P
Cadmium	ug/L	1000	943		94	80 - 120	P
Chromium	ug/L	2000	1870		94	80 - 120	P
Lead	ug/L	5000	4630		93	80 - 120	P
Selenium	ug/L	10000	9600		96	80 - 120	P
Silver	ug/L	380	346		91	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167839BS Mercury	ug/L	4.0	4.05		101	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** Q1913
Contract: AECO02 **Lab Code:** CHEM **Case No.:** Q1913 **SAS No.:** Q1913

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167868BS							
Arsenic	ug/L	4000	3800		95	80 - 120	P
Barium	ug/L	1000	825		82	80 - 120	P
Cadmium	ug/L	1000	923		92	80 - 120	P
Chromium	ug/L	2000	1800		90	80 - 120	P
Lead	ug/L	5000	4560		91	80 - 120	P
Selenium	ug/L	10000	9530		95	80 - 120	P
Silver	ug/L	380	337		89	80 - 120	P