

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: AR1660CCC500 Data File : PP071711.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	494.440	500.000	-1.1
Aroclor-1016-2	5.685	5.593	5.793	479.570	500.000	-4.1
Aroclor-1016-3	5.747	5.654	5.854	480.090	500.000	-4.0
Aroclor-1016-4	5.844	5.752	5.952	485.850	500.000	-2.8
Aroclor-1016-5	6.137	6.045	6.245	471.970	500.000	-5.6
Aroclor-1260-1	7.255	7.165	7.365	497.220	500.000	-0.6
Aroclor-1260-2	7.510	7.418	7.618	481.060	500.000	-3.8
Aroclor-1260-3	7.868	7.777	7.977	489.720	500.000	-2.1
Aroclor-1260-4	8.092	8.002	8.202	473.900	500.000	-5.2
Aroclor-1260-5	8.412	8.322	8.522	493.150	500.000	-1.4
Decachlorobiphenyl	10.223	10.137	10.337	50.290	50.000	0.6
Tetrachloro-m-xylene	4.510	4.417	4.617	47.870	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: AR1660CCC500 Data File : PP071711.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.891	4.798	4.998	481.740	500.000	-3.7
Aroclor-1016-2	4.910	4.816	5.016	482.850	500.000	-3.4
Aroclor-1016-3	5.086	4.993	5.193	498.410	500.000	-0.3
Aroclor-1016-4	5.128	5.036	5.236	495.700	500.000	-0.9
Aroclor-1016-5	5.342	5.250	5.450	504.660	500.000	0.9
Aroclor-1260-1	6.376	6.285	6.485	510.780	500.000	2.2
Aroclor-1260-2	6.564	6.473	6.673	498.980	500.000	-0.2
Aroclor-1260-3	6.717	6.627	6.827	483.020	500.000	-3.4
Aroclor-1260-4	7.189	7.098	7.298	491.580	500.000	-1.7
Aroclor-1260-5	7.429	7.339	7.539	495.450	500.000	-0.9
Decachlorobiphenyl	8.836	8.749	8.949	53.350	50.000	6.7
Tetrachloro-m-xylene	3.807	3.712	3.912	50.760	50.000	1.5