

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071725.D Time Analyzed: 02:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	490.410	500.000	-1.9
Aroclor-1016-2	5.686	5.593	5.793	485.270	500.000	-2.9
Aroclor-1016-3	5.748	5.654	5.854	483.700	500.000	-3.3
Aroclor-1016-4	5.846	5.752	5.952	478.380	500.000	-4.3
Aroclor-1016-5	6.139	6.045	6.245	480.660	500.000	-3.9
Aroclor-1260-1	7.257	7.165	7.365	545.630	500.000	9.1
Aroclor-1260-2	7.511	7.418	7.618	501.340	500.000	0.3
Aroclor-1260-3	7.869	7.777	7.977	494.860	500.000	-1.0
Aroclor-1260-4	8.094	8.002	8.202	486.430	500.000	-2.7
Aroclor-1260-5	8.413	8.322	8.522	499.420	500.000	-0.1
Decachlorobiphenyl	10.226	10.137	10.337	51.370	50.000	2.7
Tetrachloro-m-xylene	4.511	4.417	4.617	48.180	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071725.D Time Analyzed: 02:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	479.500	500.000	-4.1
Aroclor-1016-2	4.911	4.816	5.016	483.950	500.000	-3.2
Aroclor-1016-3	5.088	4.993	5.193	493.560	500.000	-1.3
Aroclor-1016-4	5.129	5.036	5.236	492.920	500.000	-1.4
Aroclor-1016-5	5.343	5.250	5.450	525.370	500.000	5.1
Aroclor-1260-1	6.378	6.285	6.485	518.250	500.000	3.7
Aroclor-1260-2	6.566	6.473	6.673	510.590	500.000	2.1
Aroclor-1260-3	6.719	6.627	6.827	505.300	500.000	1.1
Aroclor-1260-4	7.190	7.098	7.298	500.150	500.000	0.0
Aroclor-1260-5	7.431	7.339	7.539	512.030	500.000	2.4
Decachlorobiphenyl	8.839	8.749	8.949	50.470	50.000	0.9
Tetrachloro-m-xylene	3.809	3.712	3.912	52.380	50.000	4.8