

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071731.D Time Analyzed: 08:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	498.980	500.000	-0.2
Aroclor-1016-2	5.685	5.593	5.793	501.490	500.000	0.3
Aroclor-1016-3	5.747	5.654	5.854	497.340	500.000	-0.5
Aroclor-1016-4	5.844	5.752	5.952	502.000	500.000	0.4
Aroclor-1016-5	6.137	6.045	6.245	498.000	500.000	-0.4
Aroclor-1260-1	7.255	7.165	7.365	543.030	500.000	8.6
Aroclor-1260-2	7.509	7.418	7.618	516.520	500.000	3.3
Aroclor-1260-3	7.867	7.777	7.977	520.850	500.000	4.2
Aroclor-1260-4	8.093	8.002	8.202	507.950	500.000	1.6
Aroclor-1260-5	8.412	8.322	8.522	524.890	500.000	5.0
Decachlorobiphenyl	10.225	10.137	10.337	53.410	50.000	6.8
Tetrachloro-m-xylene	4.510	4.417	4.617	50.560	50.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071731.D Time Analyzed: 08:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.798	4.998	515.080	500.000	3.0
Aroclor-1016-2	4.909	4.816	5.016	530.310	500.000	6.1
Aroclor-1016-3	5.086	4.993	5.193	545.890	500.000	9.2
Aroclor-1016-4	5.128	5.036	5.236	540.360	500.000	8.1
Aroclor-1016-5	5.342	5.250	5.450	563.040	500.000	12.6
Aroclor-1260-1	6.376	6.285	6.485	556.270	500.000	11.3
Aroclor-1260-2	6.565	6.473	6.673	542.730	500.000	8.5
Aroclor-1260-3	6.717	6.627	6.827	538.540	500.000	7.7
Aroclor-1260-4	7.188	7.098	7.298	562.110	500.000	12.4
Aroclor-1260-5	7.430	7.339	7.539	563.520	500.000	12.7
Decachlorobiphenyl	8.837	8.749	8.949	56.150	50.000	12.3
Tetrachloro-m-xylene	3.807	3.712	3.912	51.730	50.000	3.5