

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071746.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.668	5.571	5.771	492.260	500.000	-1.5
Aroclor-1016-2	5.690	5.593	5.793	501.960	500.000	0.4
Aroclor-1016-3	5.752	5.654	5.854	502.050	500.000	0.4
Aroclor-1016-4	5.849	5.752	5.952	501.980	500.000	0.4
Aroclor-1016-5	6.142	6.045	6.245	498.240	500.000	-0.4
Aroclor-1260-1	7.261	7.165	7.365	544.810	500.000	9.0
Aroclor-1260-2	7.514	7.418	7.618	508.260	500.000	1.7
Aroclor-1260-3	7.873	7.777	7.977	507.380	500.000	1.5
Aroclor-1260-4	8.098	8.002	8.202	493.200	500.000	-1.4
Aroclor-1260-5	8.417	8.322	8.522	516.130	500.000	3.2
Decachlorobiphenyl	10.231	10.137	10.337	52.670	50.000	5.3
Tetrachloro-m-xylene	4.515	4.417	4.617	49.170	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071746.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.894	4.798	4.998	538.400	500.000	7.7
Aroclor-1016-2	4.912	4.816	5.016	544.140	500.000	8.8
Aroclor-1016-3	5.089	4.993	5.193	550.190	500.000	10.0
Aroclor-1016-4	5.131	5.036	5.236	538.320	500.000	7.7
Aroclor-1016-5	5.345	5.250	5.450	589.120	500.000	17.8
Aroclor-1260-1	6.379	6.285	6.485	556.160	500.000	11.2
Aroclor-1260-2	6.568	6.473	6.673	541.960	500.000	8.4
Aroclor-1260-3	6.721	6.627	6.827	555.050	500.000	11.0
Aroclor-1260-4	7.191	7.098	7.298	558.830	500.000	11.8
Aroclor-1260-5	7.433	7.339	7.539	564.900	500.000	13.0
Decachlorobiphenyl	8.841	8.749	8.949	55.410	50.000	10.8
Tetrachloro-m-xylene	3.810	3.712	3.912	51.880	50.000	3.8