

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071760.D Time Analyzed: 22:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	494.870	500.000	-1.0
Aroclor-1016-2	5.685	5.593	5.793	492.880	500.000	-1.4
Aroclor-1016-3	5.748	5.654	5.854	489.500	500.000	-2.1
Aroclor-1016-4	5.845	5.752	5.952	492.060	500.000	-1.6
Aroclor-1016-5	6.137	6.045	6.245	489.340	500.000	-2.1
Aroclor-1260-1	7.256	7.165	7.365	534.710	500.000	6.9
Aroclor-1260-2	7.510	7.418	7.618	500.240	500.000	0.0
Aroclor-1260-3	7.868	7.777	7.977	497.090	500.000	-0.6
Aroclor-1260-4	8.092	8.002	8.202	484.830	500.000	-3.0
Aroclor-1260-5	8.411	8.322	8.522	503.610	500.000	0.7
Decachlorobiphenyl	10.224	10.137	10.337	51.400	50.000	2.8
Tetrachloro-m-xylene	4.510	4.417	4.617	49.600	50.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071760.D Time Analyzed: 22:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	538.240	500.000	7.6
Aroclor-1016-2	4.911	4.816	5.016	542.610	500.000	8.5
Aroclor-1016-3	5.088	4.993	5.193	552.880	500.000	10.6
Aroclor-1016-4	5.130	5.036	5.236	536.700	500.000	7.3
Aroclor-1016-5	5.344	5.250	5.450	583.930	500.000	16.8
Aroclor-1260-1	6.379	6.285	6.485	550.110	500.000	10.0
Aroclor-1260-2	6.566	6.473	6.673	535.140	500.000	7.0
Aroclor-1260-3	6.719	6.627	6.827	529.890	500.000	6.0
Aroclor-1260-4	7.190	7.098	7.298	540.990	500.000	8.2
Aroclor-1260-5	7.432	7.339	7.539	523.090	500.000	4.6
Decachlorobiphenyl	8.838	8.749	8.949	53.640	50.000	7.3
Tetrachloro-m-xylene	3.809	3.712	3.912	54.140	50.000	8.3