

Cover Page

Order ID : Q1913

Project ID : Meeker Ave Plumes Superfund Site RI FS

Client : AECOM

Lab Sample Number

Q1913-01
Q1913-02
Q1913-03
Q1913-04

Client Sample Number

WC-12-S-202504
WC-12-A-202504
WC-13-S-202504
WC-13-A-202504

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/7/2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1913

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 05/07/2025

LAB CHRONICLE

OrderID:	Q1913	OrderDate:	4/29/2025 2:15:00 PM
Client:	AECOM	Project:	Meeker Ave Plumes Superfund Site RI FS
Contact:	Amit Haryani	Location:	L21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1913-01	WC-12-S-202504	SOIL			04/29/25			04/29/25
			PCB	8082A		05/02/25	05/02/25	
			TCLP Pesticide	8081B		05/02/25	05/05/25	
Q1913-02	WC-12-A-202504	TCLP			04/29/25			04/29/25
			TCLP Pesticide	8081B		05/01/25	05/01/25	
			PCB	8082A		05/01/25	05/01/25	
Q1913-03	WC-13-S-202504	SOIL			04/29/25			04/29/25
			PCB	8082A		05/02/25	05/02/25	
			TCLP Pesticide	8081B		05/02/25	05/05/25	
Q1913-04	WC-13-A-202504	TCLP			04/29/25			04/29/25
			TCLP Pesticide	8081B		05/01/25	05/01/25	
			PCB	8082A		05/01/25	05/02/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1913

Order ID: Q1913

Client: AECOM

Project ID: Meeker Ave Plumes Superfund Site RI

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: Q1913

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071715.D	PIBLK-PP071715.D	Tetrachloro-m-xylene	1	20	15.9	80		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	18.4	92		60	140
PB167818BL	PB167818BL	Tetrachloro-m-xylene	1	20	18.5	93		16	158
		Decachlorobiphenyl	1	20	19.6	98		10	173
		Tetrachloro-m-xylene	2	20	20.3	102		16	158
		Decachlorobiphenyl	2	20	20.6	103		10	173
PB167818BS	PB167818BS	Tetrachloro-m-xylene	1	20	19.8	99		16	158
		Decachlorobiphenyl	1	20	20.6	103		10	173
		Tetrachloro-m-xylene	2	20	20.9	104		16	158
		Decachlorobiphenyl	2	20	21.8	109		10	173
PB167818BSD	PB167818BSD	Tetrachloro-m-xylene	1	20	19.1	96		16	158
		Decachlorobiphenyl	1	20	20.5	103		10	173
		Tetrachloro-m-xylene	2	20	19.4	97		16	158
		Decachlorobiphenyl	2	20	23.1	115		10	173
Q1913-02	WC-12-A-202504	Tetrachloro-m-xylene	1	20	20.5	103		16	158
		Decachlorobiphenyl	1	20	19.9	99		10	173
		Tetrachloro-m-xylene	2	20	22.4	112		16	158
		Decachlorobiphenyl	2	20	20.1	101		10	173
Q1913-04	WC-13-A-202504	Tetrachloro-m-xylene	1	20	14.4	72		16	158
		Decachlorobiphenyl	1	20	16.8	84		10	173
		Tetrachloro-m-xylene	2	20	18.6	93		16	158
		Decachlorobiphenyl	2	20	17.9	90		10	173
I.BLK-PP071729.D	PIBLK-PP071729.D	Tetrachloro-m-xylene	1	20	16.5	83		60	140
		Decachlorobiphenyl	1	20	17.9	90		60	140
		Tetrachloro-m-xylene	2	20	19.0	95		60	140
		Decachlorobiphenyl	2	20	19.2	96		60	140
I.BLK-PP071735.D	PIBLK-PP071735.D	Tetrachloro-m-xylene	1	20	16.7	83		60	140
		Decachlorobiphenyl	1	20	18.2	91		60	140
		Tetrachloro-m-xylene	2	20	17.8	89		60	140
		Decachlorobiphenyl	2	20	20.1	101		60	140
PB167834BL	PB167834BL	Tetrachloro-m-xylene	1	20	20.1	100		32	144
		Decachlorobiphenyl	1	20	21.6	108		32	175
		Tetrachloro-m-xylene	2	20	20.8	104		32	144
		Decachlorobiphenyl	2	20	23.8	119		32	175
PB167834BS	PB167834BS	Tetrachloro-m-xylene	1	20	20.7	103		32	144

Surrogate Summary

SDG No.: Q1913

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
PB167834BS	PB167834BS	Decachlorobiphenyl	1	20	22.1	111		32	175
		Tetrachloro-m-xylene	2	20	20.0	100		32	144
Q1928-01MS	MH-QMS	Decachlorobiphenyl	2	20	23.9	119		32	175
		Tetrachloro-m-xylene	1	20	20.5	102		32	144
		Decachlorobiphenyl	1	20	19.3	97		32	175
		Tetrachloro-m-xylene	2	20	21.9	110		32	144
		Decachlorobiphenyl	2	20	22.0	110		32	175
		Tetrachloro-m-xylene	1	20	20.2	101		32	144
Q1928-01MSD	MH-QMSD	Decachlorobiphenyl	1	20	19.5	97		32	175
		Tetrachloro-m-xylene	2	20	20.1	101		32	144
		Decachlorobiphenyl	2	20	21.0	105		32	175
		Tetrachloro-m-xylene	1	20	16.9	85		60	140
Q1913-01	WC-12-S-202504	Decachlorobiphenyl	1	20	18.6	93		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	20.3	101		60	140
		Tetrachloro-m-xylene	1	20	18.2	91		32	144
		Decachlorobiphenyl	1	20	24.9	124		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
Q1913-03	WC-13-S-202504	Decachlorobiphenyl	2	20	27.2	136		32	175
		Tetrachloro-m-xylene	1	20	26.4	132		32	144
		Decachlorobiphenyl	1	20	16.8	84		32	175
		Tetrachloro-m-xylene	2	20	16.0	80		32	144
		Decachlorobiphenyl	2	20	15.5	77		32	175
		Tetrachloro-m-xylene	1	20	16.5	82		60	140
I.BLK-PP071764.D	PIBLK-PP071764.D	Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	20.5	102		60	140
		Tetrachloro-m-xylene	2	20	20.5	102		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: 8082A

DataFile : PP071740.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1928-01MS	MH-QMS											
	AR1016	191.1	0	198	ug/kg	104				55	146	
	AR1260	191.1	0	195	ug/kg	102				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: 8082A

DataFile : PP071741.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1928-01MSD	MH-QMSD											
	AR1016	190.8	0	186	ug/kg	97		7		55	146	20
	AR1260	190.8	0	182	ug/kg	95		7		31	146	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: **8082A** **Datafile :** PP071717.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167818BS	AR1016	5	4.80	ug/L	96				61	112	
	AR1260	5	4.70	ug/L	94				66	113	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: **8082A** **Datafile :** PP071718.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167818BSD	AR1016	5	4.50	ug/L	90	6			61	112	20
	AR1260	5	4.80	ug/L	96	2			66	113	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: **8082A** **Datafile :** PP071738.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167834BS	AR1016	166.5	157	ug/kg	94				71	120	
	AR1260	166.5	165	ug/kg	99				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167818BL

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM Case No.: Q1913

SAS No.: Q1913 SDG NO.: Q1913

Lab Sample ID: PB167818BL

Lab File ID: PP071716.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/01/2025

Date Analyzed (1): 05/01/2025

Date Analyzed (2): 05/01/2025

Time Analyzed (1): 23:04

Time Analyzed (2): 23:04

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167818BS	PB167818BS	PP071717.D	05/01/2025	05/01/2025
PB167818BSD	PB167818BSD	PP071718.D	05/01/2025	05/01/2025
WC-12-A-202504	Q1913-02	PP071719.D	05/01/2025	05/01/2025
WC-13-A-202504	Q1913-04	PP071720.D	05/02/2025	05/02/2025

COMMENTS: _____

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167834BL

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM Case No.: Q1913

SAS No.: Q1913 SDG NO.: Q1913

Lab Sample ID: PB167834BL

Lab File ID: PP071737.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/02/2025

Date Analyzed (1): 05/02/2025

Date Analyzed (2): 05/02/2025

Time Analyzed (1): 13:19

Time Analyzed (2): 13:19

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB167834BS	PB167834BS	PP071738.D	05/02/2025	05/02/2025
MH-QMS	Q1928-01MS	PP071740.D	05/02/2025	05/02/2025
MH-QMSD	Q1928-01MSD	PP071741.D	05/02/2025	05/02/2025
WC-12-S-202504	Q1913-01	PP071757.D	05/02/2025	05/02/2025
WC-13-S-202504	Q1913-03	PP071758.D	05/02/2025	05/02/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	AECOM	Date Collected:	04/29/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/29/25
Client Sample ID:	WC-12-S-202504	SDG No.:	Q1913
Lab Sample ID:	Q1913-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	80.5
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071757.D	1	05/02/25 09:45	05/02/25 20:09	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.1	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.1	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.1	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.1	ug/kg
12672-29-6	Aroclor-1248	7.30	U	7.30	21.1	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.1	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.1	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.1	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.2		32 - 175	136%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071757.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 20:09
 Operator : YP\AJ
 Sample : Q1913-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-12-S-202504

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 00:24:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.809	36024807	30688533	18.170m	21.786
2) SA Decachlor...	10.231	8.840	36029091	23827480	24.882	27.168

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071757.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 20:09
Operator : YP\AJ
Sample : Q1913-01
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WC-12-S-202504


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:24:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.513 min
Delta R.T.: -0.004 min
Response: 36024807
Conc: 18.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-12-S-202504


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025




#1 Tetrachloro-m-xylene

R.T.: 3.809 min
Delta R.T.: -0.003 min
Response: 30688533
Conc: 21.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: -0.007 min
Response: 36029091
Conc: 24.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 23827480
Conc: 27.17 ng/ml

Report of Analysis

Client:	AECOM	Date Collected:	04/29/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/29/25
Client Sample ID:	WC-12-A-202504	SDG No.:	Q1913
Lab Sample ID:	Q1913-02	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	980	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071719.D	1	05/01/25 08:45	05/01/25 23:53	PB167818

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.099	U	0.099	0.51	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.51	ug/L
11141-16-5	Aroclor-1232	0.098	U	0.098	0.51	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.51	ug/L
12672-29-6	Aroclor-1248	0.072	U	0.072	0.51	ug/L
11097-69-1	Aroclor-1254	0.096	U	0.096	0.51	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.51	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.51	ug/L
11096-82-5	Aroclor-1260	0.083	U	0.083	0.51	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.4		16 - 158	112%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.1		10 - 173	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071719.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 23:53
 Operator : YP\AJ
 Sample : Q1913-02
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-12-A-202504

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:36:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.808	40730900	31629597	20.544	22.454
2) SA Decachlor...	10.225	8.837	28757021	17659906	19.860	20.136

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071719.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 23:53
Operator : YP\AJ
Sample : Q1913-02
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
WC-12-A-202504

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:36:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.511 min
Delta R.T.: -0.006 min
Response: 40730900
Conc: 20.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-12-A-202504




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 31629597
Conc: 22.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: -0.012 min
Response: 28757021
Conc: 19.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.837 min
Delta R.T.: -0.012 min
Response: 17659906
Conc: 20.14 ng/ml

Report of Analysis

Client:	AECOM	Date Collected:	04/29/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/29/25
Client Sample ID:	WC-13-S-202504	SDG No.:	Q1913
Lab Sample ID:	Q1913-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	83.3
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071758.D	1	05/02/25 09:45	05/02/25 20:26	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.4	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.4	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.4	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.4	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.4	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.4	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.4	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.4	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.4		32 - 144	132%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071758.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 20:26
 Operator : YP\AJ
 Sample : Q1913-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-13-S-202504

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 00:25:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.810	52356856	22549117	26.408m	16.008m#
2) SA Decachlor...	10.232	8.843	24342246	13566364	16.811	15.468m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071758.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 20:26
Operator : YP\AJ
Sample : Q1913-03
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WC-13-S-202504


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:25:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.504 min
 Delta R.T.: -0.013 min
 Response: 52356856
 Conc: 26.41 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 WC-13-S-202504


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025




#1 Tetrachloro-m-xylene

R.T.: 3.810 min
 Delta R.T.: -0.002 min
 Response: 22549117
 Conc: 16.01 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.232 min
 Delta R.T.: -0.006 min
 Response: 24342246
 Conc: 16.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.843 min
 Delta R.T.: -0.007 min
 Response: 13566364
 Conc: 15.47 ng/ml m

Report of Analysis

Client:	AECOM	Date Collected:	04/29/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/29/25
Client Sample ID:	WC-13-A-202504	SDG No.:	Q1913
Lab Sample ID:	Q1913-04	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	990	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071720.D	1	05/01/25 08:45	05/02/25 00:09	PB167818

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.098	U	0.098	0.51	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.51	ug/L
11141-16-5	Aroclor-1232	0.097	U	0.097	0.51	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.51	ug/L
12672-29-6	Aroclor-1248	0.072	U	0.072	0.51	ug/L
11097-69-1	Aroclor-1254	0.095	U	0.095	0.51	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.51	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.51	ug/L
11096-82-5	Aroclor-1260	0.082	U	0.082	0.51	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		16 - 158	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		10 - 173	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071720.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 00:09
 Operator : YP\AJ
 Sample : Q1913-04
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-13-A-202504

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/02/2025
 Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:36:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.807	28612636	26192989	14.432m	18.595m#
2) SA Decachlor...	10.225	8.837	24289721	15702065	16.775	17.903m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071720.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 00:09
Operator : YP\AJ
Sample : Q1913-04
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WC-13-A-202504


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:36:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.511 min
Delta R.T.: -0.006 min
Response: 28612636
Conc: 14.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-13-A-202504


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
Supervised By :mohammad ahmed 05/05/2025




#1 Tetrachloro-m-xylene

R.T.: 3.807 min
Delta R.T.: -0.005 min
Response: 26192989
Conc: 18.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: -0.012 min
Response: 24289721
Conc: 16.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.837 min
Delta R.T.: -0.012 min
Response: 15702065
Conc: 17.90 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract: AECO02Lab Code: CHEMCase No.: Q1913SAS No.: Q1913SDG NO.: Q1913Instrument ID: ECD_PCalibration Date(s): 04/22/2025 04/22/2025Calibration Times: 10:29 17:49GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP071389.D</u>	RT 750 = <u>PP071390.D</u>
RT 500 = <u>PP071391.D</u>	RT 250 = <u>PP071392.D</u>	RT 050 = <u>PP071393.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1242-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1242-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1242-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1242-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1242-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.51	4.52	4.51	4.51	4.52	4.51	4.41	4.61
Aroclor-1248-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1248-2	(2)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Aroclor-1248-3	(3)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1248-4	(4)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Aroclor-1248-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.24	10.24	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.51	4.51	4.51	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61
Aroclor-1268-1	(1)	8.73	8.73	8.73	8.74	8.73	8.73	8.63	8.83
Aroclor-1268-2	(2)	8.83	8.83	8.83	8.83	8.83	8.83	8.73	8.93
Aroclor-1268-3	(3)	9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Aroclor-1268-4	(4)	9.48	9.48	9.48	9.48	9.48	9.48	9.38	9.58
Aroclor-1268-5	(5)	9.89	9.89	9.89	9.90	9.89	9.90	9.80	10.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.24	10.23	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene	4.51	4.52	4.52	4.52	4.51	4.52	4.42	4.62

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>AECO02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1913</u>	SAS No.:	<u>Q1913</u>	SDG NO.:	<u>Q1913</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>04/22/2025</u>	<u>04/22/2025</u>		
		Calibration Times:		<u>10:29</u>	<u>17:49</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D
RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: AECO02
Lab Code: CHEM **Case No.:** Q1913 **SAS No.:** Q1913 **SDG NO.:** Q1913
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712
Aroclor-1242-1	(1)	53870259	55684409	58494980	58676784	47130380	54771362
Aroclor-1242-2	(2)	81881116	85641493	90236470	89744200	77399340	84980524
Aroclor-1242-3	(3)	49423643	51294363	53309060	53836868	47752740	51123335
Aroclor-1242-4	(4)	41638031	42855355	45034468	44656444	36858500	42208560
Aroclor-1242-5	(5)	44955701	46392219	48330700	47956460	43817460	46290508
Decachlorobiphenyl		1368547670	1421241013	1488838040	1503303480	1179092600	1392204561
Tetrachloro-m-xylene		1868945350	1926494133	2017272340	2030390160	1613146400	1891249677
Aroclor-1248-1	(1)	42447397	43370463	44490672	45736348	38649560	42938888
Aroclor-1248-2	(2)	57932834	60215229	62023042	63285912	58705480	60432499
Aroclor-1248-3	(3)	65030901	68053485	69372090	70318056	61147600	66784426
Aroclor-1248-4	(4)	80777360	84441085	86910244	89142916	81782860	84610893
Aroclor-1248-5	(5)	76029464	78502729	80776474	83552224	78380620	79448302
Decachlorobiphenyl		1379130770	1440268240	1485619680	1521197800	1358343200	1436911938
Tetrachloro-m-xylene		1879191440	1925518227	1989640420	2037603040	1811949400	1928780505
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235
Aroclor-1268-1	(1)	299594266	309097848	319744804	331826088	282063800	308465361



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	251109850	259047004	267960354	275856908	228123960	256419615	7
Aroclor-1268-3	(3)	211915080	218354113	225284280	231505696	216470880	220706010	4
Aroclor-1268-4	(4)	93231632	96431025	99007040	98059984	82736300	93893196	7
Aroclor-1268-5	(5)	590082071	603740247	621062448	631335032	536707900	596585540	6
Decachlorobiphenyl		2486623660	2556709627	2669414960	2694667240	2263272000	2534137497	7
Tetrachloro-m-xylene		1883977950	1931567613	1982008260	2006991440	1590966000	1879102253	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: AECO02
Lab Code: CHEM **Case No.:** Q1913 **SAS No.:** Q1913 **SDG NO.:** Q1913
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49
GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409
Aroclor-1242-1	(1)	42717527	44660211	47919806	49576672	46343800	46243603
Aroclor-1242-2	(2)	62346958	64901092	68295906	70279172	64290540	66022734
Aroclor-1242-3	(3)	34614031	35703872	38469574	40739776	32714120	36448275
Aroclor-1242-4	(4)	33518120	34659665	37548334	39683016	33143760	35710579
Aroclor-1242-5	(5)	41214500	42688541	44971456	47258656	41407120	43508055
Decachlorobiphenyl		815615080	851726400	897002660	917481280	766770800	849719244
Tetrachloro-m-xylene		1295159620	1325412680	1503876340	1432488040	1389045200	1389196376
Aroclor-1248-1	(1)	30996932	35945815	36538014	39256084	37796080	36106585
Aroclor-1248-2	(2)	43624561	47309389	48616518	53505512	51756020	48962400
Aroclor-1248-3	(3)	45791659	49899927	50774174	56002176	52959580	51085503
Aroclor-1248-4	(4)	53871444	58562581	58680784	64889280	67282380	60657294
Aroclor-1248-5	(5)	52150607	56290108	57608714	65271236	60133720	58290877
Decachlorobiphenyl		826375100	860756013	900400040	940375840	921771600	889935719
Tetrachloro-m-xylene		1324712230	1514156880	1548203640	1434858960	1416930600	1447772462
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971
Aroclor-1268-1	(1)	181860271	188460193	203483780	205414332	201904640	196224643



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	154801411	159518377	172755008	173272408	170561360	166181713	5
Aroclor-1268-3	(3)	128491160	133089205	143083692	144773184	135469120	136981272	5
Aroclor-1268-4	(4)	56964280	59695839	63521398	64177792	55793320	60030526	6
Aroclor-1268-5	(5)	367021219	377695863	385175068	387479012	342861840	372046600	5
Decachlorobiphenyl		1499051200	1546040093	1602790460	1652739760	1464953600	1553115023	5
Tetrachloro-m-xylene		1476470990	1445952813	1464214380	1457737600	1300882000	1429051557	5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.72	4.62	4.82	24116800
		2	4.80	4.70	4.90	18016900
		3	4.88	4.78	4.98	56611200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	44912000
		2	5.40	5.30	5.50	23013000
		3	5.69	5.59	5.79	47217600
		4	5.85	5.75	5.95	23179400
		5	5.94	5.84	6.04	16145700
Aroclor-1262	500	1	8.10	8.00	8.20	144491000
		2	8.42	8.32	8.52	274186000
		3	8.74	8.64	8.84	182878000
		4	8.82	8.72	8.92	134211000
		5	9.48	9.38	9.58	88162000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	21995000
		2	4.11	4.01	4.21	16561700
		3	4.19	4.09	4.29	48477000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	38317800
		2	4.92	4.82	5.02	38187400
		3	5.09	4.99	5.19	21175200
		4	5.18	5.08	5.28	18540400
		5	5.35	5.25	5.45	19901900
Aroclor-1262	500	1	6.94	6.84	7.04	113612000
		2	7.20	7.10	7.30	92339400
		3	7.72	7.62	7.82	75694400
		4	7.79	7.69	7.89	124826000
		5	8.29	8.19	8.39	55493200

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:29
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:18:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.811	194.7E6	138.0E6	97.025	98.760
2) SA Decachlor...	10.230	8.849	141.0E6	83114760	96.154	95.894
Target Compounds						
3) L1 AR-1016-1	5.664	4.897	64709894	48121828	960.134	944.232
4) L1 AR-1016-2	5.686	4.915	98977911	69736709	959.335	949.119
5) L1 AR-1016-3	5.749	5.092	59905708	38517048	964.650	935.461
6) L1 AR-1016-4	5.846	5.134	50563747	31206636	970.673	926.980
7) L1 AR-1016-5	6.139	5.349	47772963	39458687	970.565	925.074
31) L7 AR-1260-1	7.259	6.384	90879847	63219917	958.169	923.272
32) L7 AR-1260-2	7.512	6.572	137.2E6	77338708	963.694	924.770
33) L7 AR-1260-3	7.872	6.726	110.4E6	72062678	967.762	937.334
34) L7 AR-1260-4	8.095	7.197	105.6E6	57333188	973.982	934.793
35) L7 AR-1260-5	8.416	7.439	220.5E6	143.8E6	981.349	959.843



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:29
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:18:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:45
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:20:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.811	150.4E6	103.4E6	74.955	74.302
2) SA Decachlor...	10.233	8.848	110.0E6	65650348	75.027	75.495
Target Compounds						
3) L1 AR-1016-1	5.668	4.897	50739521	37060731	751.897	734.641
4) L1 AR-1016-2	5.690	4.915	76773673	54072327	746.071	740.558
5) L1 AR-1016-3	5.752	5.092	46595357	30341683	750.210	741.220
6) L1 AR-1016-4	5.849	5.134	39314842	24611133	753.145	737.269
7) L1 AR-1016-5	6.142	5.349	37008027	31489328	751.240	742.118
31) L7 AR-1260-1	7.261	6.383	71143144	49592774	750.053	732.641
32) L7 AR-1260-2	7.514	6.572	106.8E6	58161296	750.195	712.735
33) L7 AR-1260-3	7.873	6.725	85834111	53208274	751.834	710.374
34) L7 AR-1260-4	8.097	7.197	83085064	42584936	760.941	711.945
35) L7 AR-1260-5	8.418	7.438	171.8E6	104.4E6	759.599	713.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:45
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:20:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:02
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:15:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.812	103.3E6	70755601	50.000	50.000
2) SA Decachlor...	10.237	8.849	76129158	45115981	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.671	4.898	35041766	26903064	500.000	500.000
4) L1 AR-1016-2	5.693	4.916	53684537	38606872	500.000	500.000
5) L1 AR-1016-3	5.754	5.093	32148154	21915868	500.000	500.000
6) L1 AR-1016-4	5.852	5.136	26809538	18061510	500.000	500.000
7) L1 AR-1016-5	6.145	5.350	25335334	22925304	500.000	500.000
31) L7 AR-1260-1	7.265	6.385	49407493	36863833	500.000	500.000
32) L7 AR-1260-2	7.518	6.573	73749186	44960859	500.000	500.000
33) L7 AR-1260-3	7.877	6.727	58851402	40849117	500.000	500.000
34) L7 AR-1260-4	8.102	7.198	55605534	32665875	500.000	500.000
35) L7 AR-1260-5	8.422	7.439	114.5E6	77921339	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:02
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:15:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:18
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:30:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:29:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	52508817	36208541	25.869	25.761
2) SA Decachlor...	10.233	8.850	38498712	23374964	25.928	26.384
Target Compounds						
3) L1 AR-1016-1	5.668	4.899	17782164	14045149	259.997	270.720
4) L1 AR-1016-2	5.689	4.917	27203473	19630231	260.616	263.876
5) L1 AR-1016-3	5.752	5.094	16462482	11270654	261.124	268.530
6) L1 AR-1016-4	5.849	5.136	13735357	9273773	259.716	270.294
7) L1 AR-1016-5	6.142	5.351	12722141	11586437	256.138	266.906
31) L7 AR-1260-1	7.261	6.386	24914936	19087592	259.387	273.244
32) L7 AR-1260-2	7.515	6.574	37471925	22873239	259.800	272.056
33) L7 AR-1260-3	7.873	6.727	31628881	19836001	269.748	260.958
34) L7 AR-1260-4	8.098	7.199	32815153	17020763	286.081	275.052
35) L7 AR-1260-5	8.417	7.441	63777045	39378070	273.242	264.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:18
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:30:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:29:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:34
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:47:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:46:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8969831	7104871	4.524	5.044
2) SA Decachlor...	10.235	8.850	6503503	4207302	4.491	4.797
Target Compounds						
3) L1 AR-1016-1	5.669	4.898	3083717	3007251	45.991	56.175
4) L1 AR-1016-2	5.691	4.916	4780792	4212537	46.584	55.164
5) L1 AR-1016-3	5.753	5.094	2802822	2158254	45.466	51.131
6) L1 AR-1016-4	5.851	5.135	2247335	1759505	43.809	51.021
7) L1 AR-1016-5	6.144	5.349	2125784	2429782	44.068	54.667
31) L7 AR-1260-1	7.262	6.384	4795547	4103726	49.941	56.760
32) L7 AR-1260-2	7.516	6.573	6922549	5050265	48.383	57.743
33) L7 AR-1260-3	7.875	6.726	4575293	4394430	40.810m	56.060 #
34) L7 AR-1260-4	8.098	7.197	5275083	3401257	46.807m	53.894
35) L7 AR-1260-5	8.417	7.439	9961562	7878032	43.897m	52.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:34
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:47:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:46:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:51
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:26:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	102.0E6	70886063	50.000	50.000
2) SA Decachlor...	10.233	8.849	73129811	43684216	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.715	4.022	12058427	10997465	500.000	500.000
9) L2 AR-1221-2	4.800	4.109	9008471	8280856	500.000	500.000
10) L2 AR-1221-3	4.876	4.185	28305615	24238496	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:51
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:26:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:07
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:29:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	97587735	67524433	50.000	50.000
2) SA Decachlor...	10.233	8.851	71248479	42941843	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.878	4.186	22456038	19158895	500.000	500.000
12) L3 AR-1232-2	5.403	4.917	11506547	19093695	500.000	500.000
13) L3 AR-1232-3	5.690	5.094	23608778	10587643	500.000	500.000
14) L3 AR-1232-4	5.850	5.178	11589674	9270206	500.000	500.000
15) L3 AR-1232-5	5.940	5.351	8072861	9950926	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:07
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:29:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:23
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:20:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	186.9E6	129.5E6	96.183	92.543
2) SA Decachlor...	10.233	8.849	136.9E6	81561508	95.790	95.248
Target Compounds						
16) L4 AR-1242-1	5.668	4.897	53870259	42717527	958.842	942.603
17) L4 AR-1242-2	5.689	4.916	81881116	62346958	951.456	954.464
18) L4 AR-1242-3	5.751	5.093	49423643	34614031	962.179	947.245
19) L4 AR-1242-4	5.848	5.177	41638031	33518120	960.813	943.289
20) L4 AR-1242-5	6.580	5.701	44955701	41214500	963.821	956.409



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:23
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:20:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:39
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:23:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.5E6	99405951	74.571	72.305
2) SA Decachlor...	10.236	8.850	106.6E6	63879480	74.739	74.732
Target Compounds						
16) L4 AR-1242-1	5.669	4.899	41763307	33495158	745.553	742.700
17) L4 AR-1242-2	5.691	4.917	64231120	48675819	747.572	746.776
18) L4 AR-1242-3	5.753	5.094	38470772	26777904	749.299	738.446
19) L4 AR-1242-4	5.850	5.179	32141516	25994749	744.431	737.606
20) L4 AR-1242-5	6.581	5.703	34794164	32016406	747.305	745.293



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:39
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:23:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:56
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:13:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	100.9E6	75193817	50.000	50.000
2) SA Decachlor...	10.235	8.850	74441902	44850133	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	29247490	23959903	500.000	500.000
17) L4 AR-1242-2	5.690	4.916	45118235	34147953	500.000	500.000
18) L4 AR-1242-3	5.752	5.093	26654530	19234787	500.000	500.000
19) L4 AR-1242-4	5.849	5.178	22517234	18774167	500.000	500.000
20) L4 AR-1242-5	6.581	5.701	24165350	22485728	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:56
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:13:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:12
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:34:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:34:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	50759754	35812201	25.888	25.778
2) SA Decachlor...	10.233	8.850	37582587	22937032	26.000	26.351
Target Compounds						
16) L4 AR-1242-1	5.668	4.899	14669196	12394168	258.800	268.164
17) L4 AR-1242-2	5.689	4.917	22436050	17569793	258.254	264.383
18) L4 AR-1242-3	5.751	5.094	13459217	10184944	259.001	272.457
19) L4 AR-1242-4	5.849	5.178	11164111	9920754	256.375	272.906
20) L4 AR-1242-5	6.580	5.702	11989115	11814664	255.584	268.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:12
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:34:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:34:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:28
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:44:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:44:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8065732	6945226	4.265	4.999
2) SA Decachlor...	10.237	8.850	5895463	3833854	4.235	4.512
Target Compounds						
16) L4 AR-1242-1	5.670	4.899	2356519	2317190	43.025	50.108
17) L4 AR-1242-2	5.691	4.917	3869967	3214527	45.539	48.688
18) L4 AR-1242-3	5.754	5.094	2387637	1635706	46.703	44.877
19) L4 AR-1242-4	5.851	5.179	1842925	1657188	43.662	46.406
20) L4 AR-1242-5	6.582	5.702	2190873	2070356	47.329	47.586



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:28
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:44:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:44:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:45
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:30:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	187.9E6	132.5E6	97.145	92.221
2) SA Decachlor...	10.237	8.850	137.9E6	82637510	96.283	95.713
Target Compounds						
21) L5 AR-1248-1	5.669	4.898	42447397	30996932	976.497	917.952
22) L5 AR-1248-2	5.941	5.136	57932834	43624561	965.902	945.881
23) L5 AR-1248-3	6.144	5.178	65030901	45791659	967.700	948.403
24) L5 AR-1248-4	6.543	5.350	80777360	53871444	963.427	957.270
25) L5 AR-1248-5	6.582	5.743	76029464	52150607	969.727	950.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:45
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:30:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:01
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:33:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.813	144.4E6	113.6E6	74.770	77.657
2) SA Decachlor...	10.238	8.850	108.0E6	64556701	75.275	74.847
Target Compounds						
21) L5 AR-1248-1	5.670	4.898	32527847	26959361	748.865	781.576
22) L5 AR-1248-2	5.941	5.135	45161422	35482042	751.976	762.779
23) L5 AR-1248-3	6.144	5.177	51040114	37424945	756.312	766.560
24) L5 AR-1248-4	6.543	5.350	63330814	43921936	753.553	770.043
25) L5 AR-1248-5	6.583	5.743	58877047	42217581	750.636	762.741



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:01
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:33:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:17
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:28:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	99482021	77410182	50.000	50.000
2) SA Decachlor...	10.237	8.850	74280984	45020002	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	22245336	18269007	500.000	500.000
22) L5 AR-1248-2	5.940	5.136	31011521	24308259	500.000	500.000
23) L5 AR-1248-3	6.142	5.178	34686045	25387087	500.000	500.000
24) L5 AR-1248-4	6.542	5.350	43455122	29340392	500.000	500.000
25) L5 AR-1248-5	6.581	5.743	40388237	28804357	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:17
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:28:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:33
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:46:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:46:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	50940076	35871474	26.017	24.646
2) SA Decachlor...	10.236	8.850	38029945	23509396	26.110	26.655
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	11434087	9814021	259.799	275.024
22) L5 AR-1248-2	5.940	5.135	15821478	13376378	259.947	277.150
23) L5 AR-1248-3	6.143	5.178	17579514	14000544	257.788	276.598
24) L5 AR-1248-4	6.542	5.350	22285729	16222320	261.208	274.950
25) L5 AR-1248-5	6.581	5.743	20888056	16317809	262.033	282.168



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:33
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:46:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:46:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:50
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:04:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:04:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	9059747	7084653	4.697	4.893
2) SA Decachlor...	10.236	8.850	6791716	4608858	4.727	5.179
Target Compounds						
21) L5 AR-1248-1	5.667	4.899	1932478	1889804	45.005	52.340
22) L5 AR-1248-2	5.939	5.136	2935274	2587801	48.571	52.853
23) L5 AR-1248-3	6.142	5.178	3057380	2647979	45.780	51.834
24) L5 AR-1248-4	6.541	5.350	4089143	3364119	48.329	55.461
25) L5 AR-1248-5	6.581	5.744	3919031	3006686	49.328	51.581



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:04:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:06
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:51:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	188.8E6	139.1E6	96.818	97.876
2) SA Decachlor...	10.236	8.850	141.0E6	84630750	96.040	96.040
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	78574372	81690839	957.338	952.962
27) L6 AR-1254-2	6.735	5.851	121.6E6	70045476	958.538	950.650
28) L6 AR-1254-3	7.099	6.254	123.4E6	109.7E6	962.759	954.874
29) L6 AR-1254-4	7.381	6.483	112.5E6	70169900	960.617	948.370
30) L6 AR-1254-5	7.798	6.901	106.4E6	92229111	966.455	936.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:06
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:51:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:22
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:54:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	145.3E6	112.6E6	74.658	77.773
2) SA Decachlor...	10.236	8.850	110.2E6	65650190	75.033	74.666
Target Compounds						
26) L6 AR-1254-1	6.519	5.703	60934961	65520948	744.931	759.494
27) L6 AR-1254-2	6.735	5.851	94362074	56602387	745.775	762.037
28) L6 AR-1254-3	7.098	6.255	95849879	85577532	748.410	746.476
29) L6 AR-1254-4	7.381	6.483	87194278	55551540	746.295	750.532
30) L6 AR-1254-5	7.797	6.901	79441918	72082199	721.000m	737.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:22
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:54:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071408.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:38
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:48:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	100.6E6	72545075	50.000	50.000
2) SA Decachlor...	10.237	8.849	76319487	45805245	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	42788667	44877702	500.000	500.000
27) L6 AR-1254-2	6.735	5.851	66073164	38658957	500.000	500.000
28) L6 AR-1254-3	7.098	6.254	66490827	60048409	500.000	500.000
29) L6 AR-1254-4	7.381	6.483	60868718	38905081	500.000	500.000
30) L6 AR-1254-5	7.797	6.901	56892560	52345397	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071408.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:38
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:48:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:55
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:07:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:07:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	51789148	37965204	26.192	25.910
2) SA Decachlor...	10.234	8.850	38742671	23881547	26.022	26.587
Target Compounds						
26) L6 AR-1254-1	6.517	5.703	22149547	24321384	265.267	273.203
27) L6 AR-1254-2	6.733	5.851	34278927	20788594	265.367	271.757
28) L6 AR-1254-3	7.096	6.255	34405813	31095041	263.728	265.596
29) L6 AR-1254-4	7.379	6.483	31098214	20469357	261.934	269.399
30) L6 AR-1254-5	7.796	6.901	29038261	26846100	262.635	268.183



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:55
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:11
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:54:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	8612576	7342515	4.471m	5.009
2) SA Decachlor...	10.236	8.850	6643627	4341119	4.560	4.865
Target Compounds						
26) L6 AR-1254-1	6.520	5.703	3787970	5045448	46.222	55.202
27) L6 AR-1254-2	6.735	5.852	5933913	4378664	46.696	55.629
28) L6 AR-1254-3	7.099	6.255	5851823	6211772	45.798	52.416
29) L6 AR-1254-4	7.382	6.484	5633281	4095094	47.937	53.069
30) L6 AR-1254-5	7.798	6.901	4819584	5346230	44.737	52.689

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:11
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:54:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071411.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:27
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 17:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98220466	71170259	50.000	50.000
2) SA Decachlor...	10.235	8.850	74643123	44870081	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.100	6.939	72245642	56806132	500.000	500.000
37) L8 AR-1262-2	8.420	7.198	137.1E6	46169678	500.000	500.000
38) L8 AR-1262-3	8.737	7.721	91438860	37847231	500.000	500.000
39) L8 AR-1262-4	8.822	7.786	67105487	62412938	500.000	500.000
40) L8 AR-1262-5	9.478	8.286	44080983	27746588	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071411.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:27
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 17:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:44
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:44:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	188.4E6	147.6E6	95.054	100.837
2) SA Decachlor...	10.235	8.850	248.7E6	149.9E6	93.152	93.528
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	299.6E6	181.9E6	936.979	895.922m
42) L9 AR-1268-2	8.826	7.788	251.1E6	154.8E6	937.116	896.075
43) L9 AR-1268-3	9.058	7.993	211.9E6	128.5E6	940.656	898.014
44) L9 AR-1268-4	9.477	8.287	93231632	56964280	941.667	896.773
45) L9 AR-1268-5	9.894	8.588	590.1E6	367.0E6	950.117	952.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC1000

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:44:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:00
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:45:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.9E6	108.4E6	73.091	74.065
2) SA Decachlor...	10.234	8.849	191.8E6	116.0E6	71.833	72.344
Target Compounds						
41) L9 AR-1268-1	8.733	7.721	231.8E6	141.3E6	725.026	696.327m
42) L9 AR-1268-2	8.826	7.787	194.3E6	119.6E6	725.052	692.534
43) L9 AR-1268-3	9.058	7.991	163.8E6	99816904	726.929	697.612
44) L9 AR-1268-4	9.476	8.285	72323269	44771879	730.486	704.831
45) L9 AR-1268-5	9.894	8.587	452.8E6	283.3E6	729.082	735.437

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:00
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:45:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:16
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:46:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.814	99100413	73210719	50.000	50.000
2) SA Decachlor...	10.236	8.851	133.5E6	80139523	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	159.9E6	101.7E6	500.000	501.224m
42) L9 AR-1268-2	8.826	7.788	134.0E6	86377504	500.000	500.000
43) L9 AR-1268-3	9.059	7.993	112.6E6	71541846	500.000	500.000
44) L9 AR-1268-4	9.478	8.287	49503520	31760699	500.000	500.000
45) L9 AR-1268-5	9.894	8.589	310.5E6	192.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:16
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:46:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:33
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.813	50174786	36443440	25.315	24.889
2) SA Decachlor...	10.236	8.850	67366681	41318494	25.236	25.779
Target Compounds						
41) L9 AR-1268-1	8.735	7.721	82956522	51353583	259.446	252.990m
42) L9 AR-1268-2	8.828	7.788	68964227	43318102	257.367	250.749
43) L9 AR-1268-3	9.061	7.991	57876424	36193296	256.904	252.952
44) L9 AR-1268-4	9.480	8.286	24514996	16044448	247.609	252.583
45) L9 AR-1268-5	9.897	8.587	157.8E6	96869753	254.135	251.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:33
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:49
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	7954830	6504410	4.014	4.442
2) SA Decachlor...	10.235	8.850	11316360	7324768	4.239	4.570
Target Compounds						
41) L9 AR-1268-1	8.731	7.721	14103190	10095232	44.108m	49.733m
42) L9 AR-1268-2	8.825	7.787	11406198	8528068	42.567m	49.365
43) L9 AR-1268-3	9.058	7.992	10823544	6773456	48.044	47.339
44) L9 AR-1268-4	9.476	8.287	4136815	2789666	41.783	43.917
45) L9 AR-1268-5	9.894	8.588	26835395	17143092	43.209	44.507

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:49
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:05
 Operator : YP\AJ
 Sample : PP042225ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:27:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98003244	70310326	49.431	49.914
2) SA Decachlor...	10.233	8.849	73424396	44681766	50.707	50.945
Target Compounds						
3) L1 AR-1016-1	5.668	4.898	33546404	27255669	500.321	509.132
4) L1 AR-1016-2	5.690	4.916	50377176	38783246	490.870	507.875
5) L1 AR-1016-3	5.752	5.093	30221252	21921068	490.230	519.328
6) L1 AR-1016-4	5.850	5.135	25444869	18087794	496.019	524.498
7) L1 AR-1016-5	6.142	5.349	24049539	22819712	498.556	513.411
31) L7 AR-1260-1	7.262	6.384	46974832	35744650	489.196	494.399
32) L7 AR-1260-2	7.515	6.573	70731391	43416292	494.358	496.405
33) L7 AR-1260-3	7.875	6.726	58007061	39241127	517.439	500.604
34) L7 AR-1260-4	8.098	7.198	56235072	31343464	498.252	496.643
35) L7 AR-1260-5	8.418	7.439	114.7E6	75416996	506.086	500.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:05
Operator : YP\AJ
Sample : PP042225ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:27:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:38
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 02:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 02:23:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98322267	73653264	48.740	48.976
2) SA Decachlor...	10.234	8.849	73973652	44775920	49.685	49.917
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	28608484	23663233	489.076	493.809
17) L4 AR-1242-2	5.690	4.917	43367430	33449196	480.598	489.769
18) L4 AR-1242-3	5.753	5.094	25764272	18726327	483.300	486.783
19) L4 AR-1242-4	5.850	5.178	21798483	18331184	484.040	488.202
20) L4 AR-1242-5	6.581	5.702	23472650	21942094	485.668	487.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:38
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 02:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 02:23:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:27
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 07:42:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98804591	74071948	49.835	52.584
2) SA Decachlor...	10.234	8.848	74699675	45534064	51.588	51.917
Target Compounds						
21) L5 AR-1248-1	5.667	4.897	22002964	19168563	512.425	530.888
22) L5 AR-1248-2	5.938	5.135	30310191	26184570	501.555	534.789
23) L5 AR-1248-3	6.141	5.177	34304891	27346403	513.666	535.307
24) L5 AR-1248-4	6.541	5.349	43159659	31642146	510.096	521.654
25) L5 AR-1248-5	6.580	5.742	40310718	30029077	507.383	515.159



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:27
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 07:42:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:43
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:28:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 19:27:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	98111190	74345146	50.931	50.716
2) SA Decachlor...	10.232	8.849	73988546	45066273	50.788	50.509
Target Compounds						
26) L6 AR-1254-1	6.517	5.702	41728109	42206013	509.183	461.771
27) L6 AR-1254-2	6.733	5.850	64122746	35938664	504.601	456.583
28) L6 AR-1254-3	7.097	6.254	64574948	56055056	505.381	473.005
29) L6 AR-1254-4	7.379	6.482	58986526	36616929	501.956	474.524
30) L6 AR-1254-5	7.796	6.900	55244441	47442452	512.803	467.562



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:43
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:28:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 19:27:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 20:16
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1268

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.814	98428428	75098106	49.661	51.289
2) SA Decachlor...	10.234	8.850	131.8E6	79464849	49.365	49.579
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	157.9E6	98709054	493.783	486.283m
42) L9 AR-1268-2	8.827	7.788	132.3E6	84218105	493.704	487.500
43) L9 AR-1268-3	9.058	7.992	112.3E6	69254815	498.379	484.016
44) L9 AR-1268-4	9.477	8.286	48904008	30724632	493.945	483.689
45) L9 AR-1268-5	9.894	8.588	309.0E6	189.8E6	497.455	492.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 20:16
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/01/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: AR1660CCC500 Data File : PP071711.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	494.440	500.000	-1.1
Aroclor-1016-2	5.685	5.593	5.793	479.570	500.000	-4.1
Aroclor-1016-3	5.747	5.654	5.854	480.090	500.000	-4.0
Aroclor-1016-4	5.844	5.752	5.952	485.850	500.000	-2.8
Aroclor-1016-5	6.137	6.045	6.245	471.970	500.000	-5.6
Aroclor-1260-1	7.255	7.165	7.365	497.220	500.000	-0.6
Aroclor-1260-2	7.510	7.418	7.618	481.060	500.000	-3.8
Aroclor-1260-3	7.868	7.777	7.977	489.720	500.000	-2.1
Aroclor-1260-4	8.092	8.002	8.202	473.900	500.000	-5.2
Aroclor-1260-5	8.412	8.322	8.522	493.150	500.000	-1.4
Decachlorobiphenyl	10.223	10.137	10.337	50.290	50.000	0.6
Tetrachloro-m-xylene	4.510	4.417	4.617	47.870	50.000	-4.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: AR1660CCC500 Data File : PP071711.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.798	4.998	481.740	500.000	-3.7
Aroclor-1016-2	4.910	4.816	5.016	482.850	500.000	-3.4
Aroclor-1016-3	5.086	4.993	5.193	498.410	500.000	-0.3
Aroclor-1016-4	5.128	5.036	5.236	495.700	500.000	-0.9
Aroclor-1016-5	5.342	5.250	5.450	504.660	500.000	0.9
Aroclor-1260-1	6.376	6.285	6.485	510.780	500.000	2.2
Aroclor-1260-2	6.564	6.473	6.673	498.980	500.000	-0.2
Aroclor-1260-3	6.717	6.627	6.827	483.020	500.000	-3.4
Aroclor-1260-4	7.189	7.098	7.298	491.580	500.000	-1.7
Aroclor-1260-5	7.429	7.339	7.539	495.450	500.000	-0.9
Decachlorobiphenyl	8.836	8.749	8.949	53.350	50.000	6.7
Tetrachloro-m-xylene	3.807	3.712	3.912	50.760	50.000	1.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 21:27
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/02/2025
 Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:32:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.807	94905662	71506944	47.868	50.763
2) SA Decachlor...	10.223	8.836	72822728	46794503	50.292	53.354m
Target Compounds						
3) L1 AR-1016-1	5.663	4.891	33152145	25789092	494.440	481.737
4) L1 AR-1016-2	5.685	4.910	49217831	36872510	479.574	482.854
5) L1 AR-1016-3	5.747	5.086	29596227	21038246	480.091	498.414
6) L1 AR-1016-4	5.844	5.128	24923165	17094849	485.849	495.705
7) L1 AR-1016-5	6.137	5.342	22767179	22430567	471.972	504.655
31) L7 AR-1260-1	7.255	6.376	47745804	36928777	497.225	510.777
32) L7 AR-1260-2	7.510	6.564	68828245	43641487	481.056	498.980
33) L7 AR-1260-3	7.868	6.717	54899590	37863094	489.720	483.024
34) L7 AR-1260-4	8.092	7.189	53486996	31024002	473.903	491.581
35) L7 AR-1260-5	8.412	7.429	111.7E6	74703319	493.147	495.455

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 21:27
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:32:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 02:52 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 02:52 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071725.D Time Analyzed: 02:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	490.410	500.000	-1.9
Aroclor-1016-2	5.686	5.593	5.793	485.270	500.000	-2.9
Aroclor-1016-3	5.748	5.654	5.854	483.700	500.000	-3.3
Aroclor-1016-4	5.846	5.752	5.952	478.380	500.000	-4.3
Aroclor-1016-5	6.139	6.045	6.245	480.660	500.000	-3.9
Aroclor-1260-1	7.257	7.165	7.365	545.630	500.000	9.1
Aroclor-1260-2	7.511	7.418	7.618	501.340	500.000	0.3
Aroclor-1260-3	7.869	7.777	7.977	494.860	500.000	-1.0
Aroclor-1260-4	8.094	8.002	8.202	486.430	500.000	-2.7
Aroclor-1260-5	8.413	8.322	8.522	499.420	500.000	-0.1
Decachlorobiphenyl	10.226	10.137	10.337	51.370	50.000	2.7
Tetrachloro-m-xylene	4.511	4.417	4.617	48.180	50.000	-3.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071725.D Time Analyzed: 02:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	479.500	500.000	-4.1
Aroclor-1016-2	4.911	4.816	5.016	483.950	500.000	-3.2
Aroclor-1016-3	5.088	4.993	5.193	493.560	500.000	-1.3
Aroclor-1016-4	5.129	5.036	5.236	492.920	500.000	-1.4
Aroclor-1016-5	5.343	5.250	5.450	525.370	500.000	5.1
Aroclor-1260-1	6.378	6.285	6.485	518.250	500.000	3.7
Aroclor-1260-2	6.566	6.473	6.673	510.590	500.000	2.1
Aroclor-1260-3	6.719	6.627	6.827	505.300	500.000	1.1
Aroclor-1260-4	7.190	7.098	7.298	500.150	500.000	0.0
Aroclor-1260-5	7.431	7.339	7.539	512.030	500.000	2.4
Decachlorobiphenyl	8.839	8.749	8.949	50.470	50.000	0.9
Tetrachloro-m-xylene	3.809	3.712	3.912	52.380	50.000	4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071725.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 02:52
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:38:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.809	95530112	73790074	48.183	52.384
2) SA Decachlor...	10.226	8.839	74386756	44265815	51.372	50.471
Target Compounds						
3) L1 AR-1016-1	5.665	4.893	32881997	25669251	490.411	479.498
4) L1 AR-1016-2	5.686	4.911	49802664	36956005	485.272	483.947
5) L1 AR-1016-3	5.748	5.088	29818911	20833359	483.703	493.560
6) L1 AR-1016-4	5.846	5.129	24539850	16998638	478.377	492.915
7) L1 AR-1016-5	6.139	5.343	23186218	23351091	480.659	525.366
31) L7 AR-1260-1	7.257	6.378	52393711	37469150	545.628	518.251
32) L7 AR-1260-2	7.511	6.566	71730701	44657044	501.342	510.591
33) L7 AR-1260-3	7.869	6.719	55476099	39609459	494.862	505.303
34) L7 AR-1260-4	8.094	7.190	54900306	31564672	486.425	500.148
35) L7 AR-1260-5	8.413	7.431	113.2E6	77203150	499.419	512.034



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071725.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 02:52
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:38:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 08:14 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 08:14 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071731.D Time Analyzed: 08:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.663	5.571	5.771	498.980	500.000	-0.2
Aroclor-1016-2	5.685	5.593	5.793	501.490	500.000	0.3
Aroclor-1016-3	5.747	5.654	5.854	497.340	500.000	-0.5
Aroclor-1016-4	5.844	5.752	5.952	502.000	500.000	0.4
Aroclor-1016-5	6.137	6.045	6.245	498.000	500.000	-0.4
Aroclor-1260-1	7.255	7.165	7.365	543.030	500.000	8.6
Aroclor-1260-2	7.509	7.418	7.618	516.520	500.000	3.3
Aroclor-1260-3	7.867	7.777	7.977	520.850	500.000	4.2
Aroclor-1260-4	8.093	8.002	8.202	507.950	500.000	1.6
Aroclor-1260-5	8.412	8.322	8.522	524.890	500.000	5.0
Decachlorobiphenyl	10.225	10.137	10.337	53.410	50.000	6.8
Tetrachloro-m-xylene	4.510	4.417	4.617	50.560	50.000	1.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071731.D Time Analyzed: 08:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.798	4.998	515.080	500.000	3.0
Aroclor-1016-2	4.909	4.816	5.016	530.310	500.000	6.1
Aroclor-1016-3	5.086	4.993	5.193	545.890	500.000	9.2
Aroclor-1016-4	5.128	5.036	5.236	540.360	500.000	8.1
Aroclor-1016-5	5.342	5.250	5.450	563.040	500.000	12.6
Aroclor-1260-1	6.376	6.285	6.485	556.270	500.000	11.3
Aroclor-1260-2	6.565	6.473	6.673	542.730	500.000	8.5
Aroclor-1260-3	6.717	6.627	6.827	538.540	500.000	7.7
Aroclor-1260-4	7.188	7.098	7.298	562.110	500.000	12.4
Aroclor-1260-5	7.430	7.339	7.539	563.520	500.000	12.7
Decachlorobiphenyl	8.837	8.749	8.949	56.150	50.000	12.3
Tetrachloro-m-xylene	3.807	3.712	3.912	51.730	50.000	3.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 08:14
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 11:38:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.807	100.2E6	72862470	50.557	51.725
2) SA Decachlor...	10.225	8.837	77344018	49244565	53.414	56.148m
Target Compounds						
3) L1 AR-1016-1	5.663	4.891	33456318	27574298	498.977	515.084
4) L1 AR-1016-2	5.685	4.909	51467369	40496465	501.493	530.310
5) L1 AR-1016-3	5.747	5.086	30659844	23042048	497.344	545.885
6) L1 AR-1016-4	5.844	5.128	25751743	18634813	502.001	540.360
7) L1 AR-1016-5	6.137	5.342	24022596	25025541	497.998	563.039
31) L7 AR-1260-1	7.255	6.376	52144135	40218075	543.029	556.272
32) L7 AR-1260-2	7.509	6.565	73902889	47467800	516.524	542.728
33) L7 AR-1260-3	7.867	6.717	58389844	42214637	520.854	538.537
34) L7 AR-1260-4	8.093	7.188	57329384	35475102	507.947	562.110
35) L7 AR-1260-5	8.412	7.430	118.9E6	84966400	524.895	563.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 08:14
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 11:38:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:37 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-5	(5)	8.42	8.42	8.32	8.52	0.00
Tetrachloro-m-xylene		4.52	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 16:37 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071746.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.668	5.571	5.771	492.260	500.000	-1.5
Aroclor-1016-2	5.690	5.593	5.793	501.960	500.000	0.4
Aroclor-1016-3	5.752	5.654	5.854	502.050	500.000	0.4
Aroclor-1016-4	5.849	5.752	5.952	501.980	500.000	0.4
Aroclor-1016-5	6.142	6.045	6.245	498.240	500.000	-0.4
Aroclor-1260-1	7.261	7.165	7.365	544.810	500.000	9.0
Aroclor-1260-2	7.514	7.418	7.618	508.260	500.000	1.7
Aroclor-1260-3	7.873	7.777	7.977	507.380	500.000	1.5
Aroclor-1260-4	8.098	8.002	8.202	493.200	500.000	-1.4
Aroclor-1260-5	8.417	8.322	8.522	516.130	500.000	3.2
Decachlorobiphenyl	10.231	10.137	10.337	52.670	50.000	5.3
Tetrachloro-m-xylene	4.515	4.417	4.617	49.170	50.000	-1.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071746.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.894	4.798	4.998	538.400	500.000	7.7
Aroclor-1016-2	4.912	4.816	5.016	544.140	500.000	8.8
Aroclor-1016-3	5.089	4.993	5.193	550.190	500.000	10.0
Aroclor-1016-4	5.131	5.036	5.236	538.320	500.000	7.7
Aroclor-1016-5	5.345	5.250	5.450	589.120	500.000	17.8
Aroclor-1260-1	6.379	6.285	6.485	556.160	500.000	11.2
Aroclor-1260-2	6.568	6.473	6.673	541.960	500.000	8.4
Aroclor-1260-3	6.721	6.627	6.827	555.050	500.000	11.0
Aroclor-1260-4	7.191	7.098	7.298	558.830	500.000	11.8
Aroclor-1260-5	7.433	7.339	7.539	564.900	500.000	13.0
Decachlorobiphenyl	8.841	8.749	8.949	55.410	50.000	10.8
Tetrachloro-m-xylene	3.810	3.712	3.912	51.880	50.000	3.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 16:37
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 17:12:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.810	97485949	73085218	49.170	51.884
2) SA Decachlor...	10.231	8.841	76266879	48599036	52.670	55.412m
Target Compounds						
3) L1 AR-1016-1	5.668	4.894	33006066	28822306	492.262	538.397
4) L1 AR-1016-2	5.690	4.912	51515047	41552198	501.957	544.135
5) L1 AR-1016-3	5.752	5.089	30949621	23223912	502.045	550.194
6) L1 AR-1016-4	5.849	5.131	25750719	18564302	501.981	538.315
7) L1 AR-1016-5	6.142	5.345	24034491	26184770	498.244	589.120
31) L7 AR-1260-1	7.261	6.379	52315402	40210154	544.812	556.163
32) L7 AR-1260-2	7.514	6.568	72720473	47400895	508.260	541.964
33) L7 AR-1260-3	7.873	6.721	56879283	43509105	507.379	555.051
34) L7 AR-1260-4	8.098	7.191	55665243	35268306	493.203	558.833
35) L7 AR-1260-5	8.417	7.433	116.9E6	85174758	516.133	564.904

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 16:37
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 17:12:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 22:20 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Continuing Calib Date: 05/02/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 22:20 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071760.D Time Analyzed: 22:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.663	5.571	5.771	494.870	500.000	-1.0
Aroclor-1016-2	5.685	5.593	5.793	492.880	500.000	-1.4
Aroclor-1016-3	5.748	5.654	5.854	489.500	500.000	-2.1
Aroclor-1016-4	5.845	5.752	5.952	492.060	500.000	-1.6
Aroclor-1016-5	6.137	6.045	6.245	489.340	500.000	-2.1
Aroclor-1260-1	7.256	7.165	7.365	534.710	500.000	6.9
Aroclor-1260-2	7.510	7.418	7.618	500.240	500.000	0.0
Aroclor-1260-3	7.868	7.777	7.977	497.090	500.000	-0.6
Aroclor-1260-4	8.092	8.002	8.202	484.830	500.000	-3.0
Aroclor-1260-5	8.411	8.322	8.522	503.610	500.000	0.7
Decachlorobiphenyl	10.224	10.137	10.337	51.400	50.000	2.8
Tetrachloro-m-xylene	4.510	4.417	4.617	49.600	50.000	-0.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP071760.D Time Analyzed: 22:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.893	4.798	4.998	538.240	500.000	7.6
Aroclor-1016-2	4.911	4.816	5.016	542.610	500.000	8.5
Aroclor-1016-3	5.088	4.993	5.193	552.880	500.000	10.6
Aroclor-1016-4	5.130	5.036	5.236	536.700	500.000	7.3
Aroclor-1016-5	5.344	5.250	5.450	583.930	500.000	16.8
Aroclor-1260-1	6.379	6.285	6.485	550.110	500.000	10.0
Aroclor-1260-2	6.566	6.473	6.673	535.140	500.000	7.0
Aroclor-1260-3	6.719	6.627	6.827	529.890	500.000	6.0
Aroclor-1260-4	7.190	7.098	7.298	540.990	500.000	8.2
Aroclor-1260-5	7.432	7.339	7.539	523.090	500.000	4.6
Decachlorobiphenyl	8.838	8.749	8.949	53.640	50.000	7.3
Tetrachloro-m-xylene	3.809	3.712	3.912	54.140	50.000	8.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 22:20
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 00:27:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.809	98330327	76262273	49.596	54.139
2) SA Decachlor...	10.224	8.838	74424695	47046882	51.398	53.642
Target Compounds						
3) L1 AR-1016-1	5.663	4.893	33180638	28813885	494.865	538.239
4) L1 AR-1016-2	5.685	4.911	50583943	41435659	492.885	542.609
5) L1 AR-1016-3	5.748	5.088	30176494	23337183	489.504	552.877
6) L1 AR-1016-4	5.845	5.130	25241953	18508669	492.064	536.702
7) L1 AR-1016-5	6.137	5.344	23605187	25954094	489.345	583.930
31) L7 AR-1260-1	7.256	6.379	51345716	39772465	534.714	550.109
32) L7 AR-1260-2	7.510	6.566	71573693	46803940	500.245	535.138
33) L7 AR-1260-3	7.868	6.719	55726029	41536556	497.092	529.887
34) L7 AR-1260-4	8.092	7.190	54719797	34142222	484.826	540.990
35) L7 AR-1260-5	8.411	7.432	114.1E6	78869404	503.609	523.085



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 22:20
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:27:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

Instrument ID: ECD_P

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025

04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/22/2025	10:13	PP071388.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	10.23	4.51
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	10.23	4.51
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	10.24	4.52
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	10.23	4.51
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	10.24	4.52
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	10.23	4.51
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	10.23	4.51
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	10.24	4.52
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	10.24	4.51
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	10.23	4.51
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	10.24	4.52
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	10.24	4.52
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	10.24	4.52
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	10.24	4.51
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	10.24	4.51
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	10.24	4.51
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	10.24	4.52
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	10.24	4.52
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	10.24	4.52
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	10.23	4.51
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	10.24	4.51
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	10.24	4.52
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	10.24	4.51
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	10.23	4.52
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	10.24	4.52
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	10.24	4.52
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	10.24	4.51
AR1660CCC500	AR1660CCC500	05/01/2025	21:27	PP071711.D	10.22	4.51
IBLK	IBLK	05/01/2025	22:48	PP071715.D	10.23	4.51
PB167818BL	PB167818BL	05/01/2025	23:04	PP071716.D	10.22	4.51
PB167818BS	PB167818BS	05/01/2025	23:21	PP071717.D	10.23	4.51
PB167818BSD	PB167818BSD	05/01/2025	23:37	PP071718.D	10.22	4.51
WC-12-A-202504	Q1913-02	05/01/2025	23:53	PP071719.D	10.23	4.51
WC-13-A-202504	Q1913-04	05/02/2025	00:09	PP071720.D	10.23	4.51
AR1660CCC500	AR1660CCC500	05/02/2025	02:52	PP071725.D	10.23	4.51
IBLK	IBLK	05/02/2025	04:14	PP071729.D	10.22	4.51
AR1660CCC500	AR1660CCC500	05/02/2025	08:14	PP071731.D	10.23	4.51
IBLK	IBLK	05/02/2025	09:22	PP071735.D	10.23	4.51
PB167834BL	PB167834BL	05/02/2025	13:19	PP071737.D	10.22	4.51
PB167834BS	PB167834BS	05/02/2025	13:36	PP071738.D	10.23	4.51

Analytical Sequence

MH-QMS	Q1928-01MS	05/02/2025	14:09	PP071740.D	10.23	4.51
MH-QMSD	Q1928-01MSD	05/02/2025	14:25	PP071741.D	10.23	4.51
AR1660CCC500	AR1660CCC500	05/02/2025	16:37	PP071746.D	10.23	4.52
I.BLK	I.BLK	05/02/2025	18:15	PP071750.D	10.23	4.52
WC-12-S-202504	Q1913-01	05/02/2025	20:09	PP071757.D	10.23	4.51
WC-13-S-202504	Q1913-03	05/02/2025	20:26	PP071758.D	10.23	4.50
AR1660CCC500	AR1660CCC500	05/02/2025	22:20	PP071760.D	10.22	4.51
I.BLK	I.BLK	05/02/2025	23:58	PP071764.D	10.22	4.51

Analytical Sequence

Client: AECOM

SDG No.: Q1913

Project: Meeker Ave Plumes Superfund Site RI FS

Instrument ID: ECD_P

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025

04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/22/2025	10:13	PP071388.D	8.85	3.81
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	8.85	3.81
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	8.85	3.81
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	8.85	3.81
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	8.85	3.81
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	8.85	3.81
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	8.85	3.81
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	8.85	3.81
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	8.85	3.81
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	8.85	3.81
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	8.85	3.81
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	8.85	3.81
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	8.85	3.81
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	8.85	3.81
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	8.85	3.81
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	8.85	3.81
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	8.85	3.81
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	8.85	3.81
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	8.85	3.81
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	8.85	3.81
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	8.85	3.81
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	8.85	3.81
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	8.85	3.81
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	8.85	3.81
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	8.85	3.81
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	8.85	3.81
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	8.85	3.81
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	8.85	3.81
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	8.85	3.81
AR1660CCC500	AR1660CCC500	05/01/2025	21:27	PP071711.D	8.84	3.81
IBLK	IBLK	05/01/2025	22:48	PP071715.D	8.84	3.81
PB167818BL	PB167818BL	05/01/2025	23:04	PP071716.D	8.84	3.81
PB167818BS	PB167818BS	05/01/2025	23:21	PP071717.D	8.84	3.81
PB167818BSD	PB167818BSD	05/01/2025	23:37	PP071718.D	8.84	3.81
WC-12-A-202504	Q1913-02	05/01/2025	23:53	PP071719.D	8.84	3.81
WC-13-A-202504	Q1913-04	05/02/2025	00:09	PP071720.D	8.84	3.81
AR1660CCC500	AR1660CCC500	05/02/2025	02:52	PP071725.D	8.84	3.81
IBLK	IBLK	05/02/2025	04:14	PP071729.D	8.84	3.81
AR1660CCC500	AR1660CCC500	05/02/2025	08:14	PP071731.D	8.84	3.81
IBLK	IBLK	05/02/2025	09:22	PP071735.D	8.84	3.81
PB167834BL	PB167834BL	05/02/2025	13:19	PP071737.D	8.84	3.81
PB167834BS	PB167834BS	05/02/2025	13:36	PP071738.D	8.84	3.81

Analytical Sequence

MH-QMS	Q1928-01MS	05/02/2025	14:09	PP071740.D	8.84	3.81
MH-QMSD	Q1928-01MSD	05/02/2025	14:25	PP071741.D	8.84	3.81
AR1660CCC500	AR1660CCC500	05/02/2025	16:37	PP071746.D	8.84	3.81
I.BLK	I.BLK	05/02/2025	18:15	PP071750.D	8.84	3.81
WC-12-S-202504	Q1913-01	05/02/2025	20:09	PP071757.D	8.84	3.81
WC-13-S-202504	Q1913-03	05/02/2025	20:26	PP071758.D	8.84	3.81
AR1660CCC500	AR1660CCC500	05/02/2025	22:20	PP071760.D	8.84	3.81
I.BLK	I.BLK	05/02/2025	23:58	PP071764.D	8.84	3.81

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167818BS

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

Lab Sample ID: PB167818BS Date(s) Analyzed: 05/01/2025 05/01/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071717.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.664	5.614	5.714	4.48	4.40	8.7
COLUMN 1	2	5.685	5.635	5.735	4.48		
	3	5.747	5.697	5.797	4.24		
	4	5.845	5.795	5.895	4.51		
	5	6.137	6.087	6.187	4.22		
	1	4.891	4.841	4.941	4.73		
COLUMN 2	2	4.91	4.86	4.96	4.76	4.80	
	3	5.087	5.037	5.137	4.86		
	4	5.128	5.078	5.178	4.87		
	5	5.343	5.293	5.393	4.71		
Aroclor-1260	1	7.256	7.206	7.306	4.50	4.30	8.89
COLUMN 1	2	7.51	7.46	7.56	4.59		
	3	7.868	7.818	7.918	3.95		
	4	8.092	8.042	8.142	4.14		
	5	8.413	8.363	8.463	4.15		
	1	6.376	6.326	6.426	4.93		
COLUMN 2	2	6.565	6.515	6.615	4.94	4.70	
	3	6.717	6.667	6.767	4.95		
	4	7.189	7.139	7.239	4.41		
	5	7.43	7.38	7.48	4.29		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167818BSD

Contract: AEEO02

Lab Code: CHEM

Case No.: Q1913

SAS No.: Q1913

SDG NO.: Q1913

Lab Sample ID: PB167818BSD

Date(s) Analyzed: 05/01/2025

05/01/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071718.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.662	5.612	5.712	4.46	4.40	2.25
COLUMN 1	2	5.683	5.633	5.733	4.59		
	3	5.745	5.695	5.795	4.35		
	4	5.843	5.793	5.893	4.49		
	5	6.135	6.085	6.185	4.34		
	1	4.891	4.841	4.941	4.42		
COLUMN 2	2	4.909	4.859	4.959	4.47	4.50	
	3	5.086	5.036	5.136	4.59		
	4	5.128	5.078	5.178	4.59		
	5	5.343	5.293	5.393	4.52		
Aroclor-1260	1	7.254	7.204	7.304	4.65	4.40	8.7
COLUMN 1	2	7.508	7.458	7.558	4.70		
	3	7.867	7.817	7.917	4.08		
	4	8.091	8.041	8.141	4.25		
	5	8.41	8.36	8.46	4.25		
	1	6.376	6.326	6.426	4.86		
COLUMN 2	2	6.565	6.515	6.615	4.93	4.80	
	3	6.717	6.667	6.767	5.06		
	4	7.189	7.139	7.239	4.65		
	5	7.43	7.38	7.48	4.66		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167834BS

Contract: AECO02

Lab Code: CHEM

Case No.: Q1913

SAS No.: Q1913

SDG NO.: Q1913

Lab Sample ID: PB167834BS

Date(s) Analyzed: 05/02/2025

05/02/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071738.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.666	5.616	5.716	148	151	3.9
COLUMN 1	2	5.688	5.638	5.738	153		
	3	5.749	5.699	5.799	151		
	4	5.847	5.797	5.897	154		
	5	6.14	6.09	6.19	148		
	1	4.892	4.842	4.942	160		
COLUMN 2	2	4.91	4.86	4.96	153	157	
	3	5.088	5.038	5.138	160		
	4	5.129	5.079	5.179	159		
	5	5.343	5.293	5.393	155		
Aroclor-1260	1	7.259	7.209	7.309	155	147	11.54
COLUMN 1	2	7.513	7.463	7.563	156		
	3	7.871	7.821	7.921	137		
	4	8.094	8.044	8.144	143		
	5	8.415	8.365	8.465	143		
	1	6.376	6.326	6.426	170	165	
COLUMN 2	2	6.566	6.516	6.616	171		
	3	6.718	6.668	6.768	174		
	4	7.19	7.14	7.24	156		
	5	7.431	7.381	7.481	154		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

MH-QMS

Contract: AECO02

Lab Code: CHEM

Case No.: Q1913

SAS No.: Q1913

SDG NO.: Q1913

Lab Sample ID: Q1928-01MS

Date(s) Analyzed: 05/02/2025

05/02/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071740.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.663	5.613	5.713	188	191	3.6
COLUMN 1	2	5.685	5.635	5.735	189		
	3	5.747	5.697	5.797	188		
	4	5.845	5.795	5.895	201		
	5	6.136	6.086	6.186	187		
	1	4.891	4.841	4.941	196		
COLUMN 2	2	4.91	4.86	4.96	203	198	
	3	5.087	5.037	5.137	203		
	4	5.129	5.079	5.179	198		
	5	5.343	5.293	5.393	189		
Aroclor-1260	1	7.257	7.207	7.307	184	175	10.81
COLUMN 1	2	7.51	7.46	7.56	187		
	3	7.869	7.819	7.919	162		
	4	8.093	8.043	8.143	173		
	5	8.413	8.363	8.463	169		
	1	6.377	6.327	6.427	206		
COLUMN 2	2	6.565	6.515	6.615	196	195	
	3	6.718	6.668	6.768	202		
	4	7.189	7.139	7.239	188		
	5	7.432	7.382	7.482	182		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

MH-QMSD

Contract: AECO02

Lab Code: CHEM

Case No.: Q1913

SAS No.: Q1913

SDG NO.: Q1913

Lab Sample ID: Q1928-01MSD

Date(s) Analyzed: 05/02/2025

05/02/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071741.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.666	5.616	5.716	186	185	0.54
COLUMN 1	2	5.687	5.637	5.737	183		
	3	5.75	5.7	5.8	183		
	4	5.847	5.797	5.897	199		
	5	6.138	6.088	6.188	176		
	1	4.893	4.843	4.943	192		
COLUMN 2	2	4.911	4.861	4.961	187	186	
	3	5.089	5.039	5.139	190		
	4	5.13	5.08	5.18	184		
	5	5.344	5.294	5.394	176		
Aroclor-1260	1	7.258	7.208	7.308	182	173	5.07
COLUMN 1	2	7.512	7.462	7.562	184		
	3	7.871	7.821	7.921	162		
	4	8.095	8.045	8.145	171		
	5	8.414	8.364	8.464	168		
	1	6.378	6.328	6.428	190	182	
COLUMN 2	2	6.567	6.517	6.617	188		
	3	6.719	6.669	6.769	190		
	4	7.19	7.14	7.24	173		
	5	7.432	7.382	7.482	167		



QC SAMPLE DATA

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	
Client Sample ID:	PB167818BL	SDG No.:	Q1913
Lab Sample ID:	PB167818BL	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071716.D	1	05/01/25 08:45	05/01/25 23:04	PB167818

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.3		16 - 158	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.6		10 - 173	103%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071716.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 23:04
 Operator : YP\AJ
 Sample : PB167818BL
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167818BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
 Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:34:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

 System Monitoring Compounds

1) SA Tetrachlo...	4.510	3.807	36703742	28605416	18.513	20.307
2) SA Decachlor...	10.224	8.837	28361863	18035705	19.587	20.564m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071716.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 23:04
Operator : YP\AJ
Sample : PB167818BL
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB167818BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:34:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 36703742
Conc: 18.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167818BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2025
Supervised By :mohammad ahmed 05/05/2025




#1 Tetrachloro-m-xylene

R.T.: 3.807 min
Delta R.T.: -0.005 min
Response: 28605416
Conc: 20.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: -0.013 min
Response: 28361863
Conc: 19.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.837 min
Delta R.T.: -0.012 min
Response: 18035705
Conc: 20.56 ng/ml m

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	
Client Sample ID:	PB167834BL	SDG No.:	Q1913
Lab Sample ID:	PB167834BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071737.D	1	05/02/25 09:45	05/02/25 13:19	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.8		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 13:19
 Operator : YP\AJ
 Sample : PB167834BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167834BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 13:53:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.807	39796682	29279728	20.073	20.786
2) SA Decachlor...	10.224	8.837	31354254	20852206	21.653	23.775m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 13:19
Operator : YP\AJ
Sample : PB167834BL
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB167834BL

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 13:53:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP071737.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 4.509 min
Delta R.T.: -0.008 min
Response: 39796682
Conc: 20.07 ng/ml

Instrument :

ECD_P

ClientSampleId :


PB167834BL

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025


Response_ Signal: PP071737.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.807 min
Delta R.T.: -0.005 min
Response: 29279728
Conc: 20.79 ng/ml


Response_ Signal: PP071737.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: -0.013 min
Response: 31354254
Conc: 21.65 ng/ml

Response_ Signal: PP071737.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.837 min
Delta R.T.: -0.012 min
Response: 20852206
Conc: 23.78 ng/ml m

Report of Analysis

Client:	AECOM	Date Collected:	04/22/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/22/25
Client Sample ID:	PIBLK-PP071388.D	SDG No.:	Q1913
Lab Sample ID:	I.BLK-PP071388.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071388.D	1		04/22/25	PP042225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 140	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071388.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:13
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:33:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.811	33865777	24238749	17.081	17.207
2) SA Decachlor...	10.233	8.847	25533987	15462668	17.634	17.630

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 33865777
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: -0.001 min
Response: 24238749
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.005 min
Response: 25533987
Conc: 17.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.847 min
Delta R.T.: -0.002 min
Response: 15462668
Conc: 17.63 ng/ml

Report of Analysis

Client:	AECOM		Date Collected:	05/01/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	05/01/25	
Client Sample ID:	PIBLK-PP071715.D		SDG No.:	Q1913	
Lab Sample ID:	I.BLK-PP071715.D		Matrix:	WATER	
Analytical Method:	SW8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	5030				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071715.D	1		05/01/25	PP050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.9		60 - 140	80%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 140	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071715.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 22:48
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:34:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.808	31622600	23438554	15.950	16.639
2) SA Decachlor...	10.226	8.838	25504964	16106916	17.614	18.365

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071715.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 22:48
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:34:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 31622600
Conc: 15.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.005 min
Response: 23438554
Conc: 16.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: -0.011 min
Response: 25504964
Conc: 17.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.838 min
Delta R.T.: -0.011 min
Response: 16106916
Conc: 18.36 ng/ml

Report of Analysis

Client:	AECOM	Date Collected:	05/02/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	05/02/25
Client Sample ID:	PIBLK-PP071729.D	SDG No.:	Q1913
Lab Sample ID:	I.BLK-PP071729.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071729.D	1		05/02/25	PP050125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		60 - 140	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		60 - 140	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071729.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 04:14
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:41:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.808	32707911	26816686	16.497	19.037
2) SA Decachlor...	10.224	8.838	25955383	16828234	17.925	19.187

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071729.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 04:14
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:41:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 32707911
Conc: 16.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 26816686
Conc: 19.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: -0.013 min
Response: 25955383
Conc: 17.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.838 min
Delta R.T.: -0.011 min
Response: 16828234
Conc: 19.19 ng/ml

Report of Analysis

Client:	AECOM	Date Collected:	05/02/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	05/02/25
Client Sample ID:	PIBLK-PP071735.D	SDG No.:	Q1913
Lab Sample ID:	I.BLK-PP071735.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071735.D	1		05/02/25	PP050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.7		60 - 140	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.2		60 - 140	91%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071735.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 09:22
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 11:40:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.808	33078850	25012083	16.684	17.756
2) SA Decachlor...	10.228	8.837	26385582	17660920	18.222	20.137m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 09:22
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 11:40:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: -0.003 min
Response: 33078850
Conc: 16.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 25012083
Conc: 17.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.009 min
Response: 26385582
Conc: 18.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.837 min
Delta R.T.: -0.012 min
Response: 17660920
Conc: 20.14 ng/ml m

Report of Analysis

Client:	AECOM	Date Collected:	05/02/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	05/02/25
Client Sample ID:	PIBLK-PP071750.D	SDG No.:	Q1913
Lab Sample ID:	I.BLK-PP071750.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071750.D	1		05/02/25	PP050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071750.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 18:15
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 00:17:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.809	33548431	26508098	16.921	18.818
2) SA Decachlor...	10.231	8.840	26939358	17785975	18.604	20.279

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071750.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 18:15
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:17:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: -0.002 min
Response: 33548431
Conc: 16.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.809 min
Delta R.T.: -0.003 min
Response: 26508098
Conc: 18.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: -0.006 min
Response: 26939358
Conc: 18.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 17785975
Conc: 20.28 ng/ml

Report of Analysis

Client:	AECOM		Date Collected:	05/02/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	05/02/25	
Client Sample ID:	PIBLK-PP071764.D		SDG No.:	Q1913	
Lab Sample ID:	I.BLK-PP071764.D		Matrix:	WATER	
Analytical Method:	SW8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	5030				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071764.D	1		05/02/25	PP050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		60 - 140	82%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.8		60 - 140	89%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 23:58
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:34:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.510	3.808	32628579	26442664	16.457	18.772
2)	SA Decachlor...	10.223	8.836	25726435	17965808	17.767	20.484

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 23:58
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 00:34:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 32628579
Conc: 16.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 26442664
Conc: 18.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.014 min
Response: 25726435
Conc: 17.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.836 min
Delta R.T.: -0.013 min
Response: 17965808
Conc: 20.48 ng/ml

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	
Client Sample ID:	PB167818BS	SDG No.:	Q1913
Lab Sample ID:	PB167818BS	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071717.D	1	05/01/25 08:45	05/01/25 23:21	PB167818

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.80		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.70		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.9		16 - 158	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.8		10 - 173	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071717.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 23:21
 Operator : YP\AJ
 Sample : PB167818BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167818BS

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:35:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.808	39276061	29434131	19.810	20.895
2) SA Decachlor...	10.225	8.836	29790758	19106193	20.574	21.785m
Target Compounds						
3) L1 AR-1016-1	5.664	4.891	30032755	25319249	447.917	472.960
4) L1 AR-1016-2	5.685	4.910	45929914	36319086	447.536	475.607
5) L1 AR-1016-3	5.747	5.087	26129668	20523319	423.859	486.215
6) L1 AR-1016-4	5.845	5.128	23131898	16786569	450.930	486.766
7) L1 AR-1016-5	6.137	5.343	20374634	20925822	422.374	470.801
31) L7 AR-1260-1	7.256	6.376	43212809	35641174	450.018	492.968
32) L7 AR-1260-2	7.510	6.565	65698504	43189803	459.182	493.816
33) L7 AR-1260-3	7.868	6.717	44285105	38819748	395.036	495.228 #
34) L7 AR-1260-4	8.092	7.189	46764527	27826891	414.341	440.923
35) L7 AR-1260-5	8.413	7.430	93953927	64706695	414.674	429.154

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071717.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 23:21
Operator : YP\AJ
Sample : PB167818BS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167818BS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:35:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	AECOM	Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	
Client Sample ID:	PB167834BS	SDG No.:	Q1913
Lab Sample ID:	PB167834BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071738.D	1	05/02/25 09:45	05/02/25 13:36	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	157		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	165		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.7		32 - 144	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.9		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071738.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 13:36
 Operator : YP\AJ
 Sample : PB167834BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167834BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 13:54:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.808	40997660	28128831	20.678	19.969
2) SA Decachlor...	10.228	8.838	32001601	20915047	22.100	23.847m
Target Compounds						
3) L1 AR-1016-1	5.666	4.892	29769725	25727314	443.994	480.583m
4) L1 AR-1016-2	5.688	4.910	47267712	35009191	460.572	458.453m
5) L1 AR-1016-3	5.749	5.088	27960921	20250532	453.564	479.752
6) L1 AR-1016-4	5.847	5.129	23723669	16495600	462.466	478.328
7) L1 AR-1016-5	6.140	5.343	21446555	20655344	444.595	464.715
31) L7 AR-1260-1	7.259	6.376	44594898	36910497	464.411	510.524
32) L7 AR-1260-2	7.513	6.566	66812401	44912381	466.967	513.511
33) L7 AR-1260-3	7.871	6.718	46193290	40991050	412.057	522.928 #
34) L7 AR-1260-4	8.094	7.190	48365231	29546051	428.523	468.163
35) L7 AR-1260-5	8.415	7.431	97551691	69892815	430.553	463.550

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071738.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 13:36
Operator : YP\AJ
Sample : PB167834BS
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB167834BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 13:54:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	AECOM	Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	
Client Sample ID:	PB167818BSD	SDG No.:	Q1913
Lab Sample ID:	PB167818BSD	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071718.D	1	05/01/25 08:45	05/01/25 23:37	PB167818

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.50		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.80		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.4		16 - 158	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.1		10 - 173	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
 Data File : PP071718.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2025 23:37
 Operator : YP\AJ
 Sample : PB167818BSD
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167818BSD

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:35:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.807	37901227	27357640	19.117	19.421
2) SA Decachlor...	10.224	8.836	29681132	20251625	20.498	23.091m
Target Compounds						
3) L1 AR-1016-1	5.662	4.891	29919741	23642455	446.231	441.638
4) L1 AR-1016-2	5.683	4.909	47118580	34165426	459.119	447.404
5) L1 AR-1016-3	5.745	5.086	26791602	19383117	434.596	459.202
6) L1 AR-1016-4	5.843	5.128	23012184	15819989	448.597	458.737
7) L1 AR-1016-5	6.135	5.343	20958977	20074200	434.488	451.641
31) L7 AR-1260-1	7.254	6.376	44629541	35147048	464.772	486.133
32) L7 AR-1260-2	7.508	6.565	67314399	43159917	470.475	493.474
33) L7 AR-1260-3	7.867	6.717	45775779	39656909	408.333	505.908
34) L7 AR-1260-4	8.091	7.189	48000924	29328163	425.296	464.711
35) L7 AR-1260-5	8.410	7.430	96319981	70305261	425.116	466.285

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050125\
Data File : PP071718.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2025 23:37
Operator : YP\AJ
Sample : PB167818BSD
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB167818BSD

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 05/02/2025

Supervised By :mohammad ahmed 05/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:35:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	AECOM	Date Collected:	05/01/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	05/01/25
Client Sample ID:	MH-QMS	SDG No.:	Q1913
Lab Sample ID:	Q1928-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	87.1
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071740.D	1	05/02/25 09:45	05/02/25 14:09	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	198		4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	195		3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.9		32 - 144	110%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.0		32 - 175	110%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 14:09
 Operator : YP\AJ
 Sample : Q1928-01MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MH-QMS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 14:26:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.808	40590559	30910500	20.473	21.943
2) SA Decachlor...	10.226	8.838	27950831	19298489	19.303	22.004
Target Compounds						
3) L1 AR-1016-1	5.663	4.891	33032711	27489939	492.659	513.508m
4) L1 AR-1016-2	5.685	4.910	50844986	40564945	495.428	531.207m
5) L1 AR-1016-3	5.747	5.087	30279953	22405655	491.182	530.809
6) L1 AR-1016-4	5.845	5.129	26971126	17865955	525.772	518.065
7) L1 AR-1016-5	6.136	5.343	23661712	21966071	490.516m	494.205
31) L7 AR-1260-1	7.257	6.377	46297786	38941475	482.145	538.615
32) L7 AR-1260-2	7.510	6.565	69940065	44960973	488.827	514.066
33) L7 AR-1260-3	7.869	6.718	47490882	41341969	423.632	527.405
34) L7 AR-1260-4	8.093	7.189	51036383	31057515	452.190	492.112
35) L7 AR-1260-5	8.413	7.432	100.1E6	71936933	441.730	477.107

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 14:09
Operator : YP\AJ
Sample : Q1928-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MH-QMS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 14:26:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	AECOM	Date Collected:	05/01/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	05/01/25
Client Sample ID:	MH-QMSD	SDG No.:	Q1913
Lab Sample ID:	Q1928-01MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	87.1
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071741.D	1	05/02/25 09:45	05/02/25 14:25	PB167834

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	186		4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	182		3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.0		32 - 175	105%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
 Data File : PP071741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 14:25
 Operator : YP\AJ
 Sample : Q1928-01MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MH-QMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2025
 Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 15:03:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.809	40134609	28356810	20.243	20.131
2) SA Decachlor...	10.228	8.838	28171916	18380573	19.456	20.957m
Target Compounds						
3) L1 AR-1016-1	5.666	4.893	32590661	26990971	486.066	504.188m
4) L1 AR-1016-2	5.687	4.911	49321385	37358555	480.583	489.219m
5) L1 AR-1016-3	5.750	5.089	29543870	20985080	479.242	497.154
6) L1 AR-1016-4	5.847	5.130	26711141	16650236	520.704	482.812
7) L1 AR-1016-5	6.138	5.344	22269597	20455876	461.657m	460.228
31) L7 AR-1260-1	7.258	6.378	45862983	35907043	477.617	496.645
32) L7 AR-1260-2	7.512	6.567	69117869	42965499	483.080	491.251
33) L7 AR-1260-3	7.871	6.719	47511476	39096588	423.816	498.760
34) L7 AR-1260-4	8.095	7.190	50598117	28634145	448.307	453.714
35) L7 AR-1260-5	8.414	7.432	99582703	66073357	439.517	438.218

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050225\
Data File : PP071741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 14:25
Operator : YP\AJ
Sample : Q1928-01MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MH-QMSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2025
Supervised By :mohammad ahmed 05/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 15:03:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP042225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP071393.D	AR-1260-3	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-4	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-5	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC750	PP071407.D	AR-1254-5	yogesh	4/23/2025 7:32:40 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC050	PP071410.D	Tetrachloro-m-xylene	yogesh	4/23/2025 7:32:41 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC1000	PP071412.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:43 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC750	PP071413.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:44 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC500	PP071414.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:46 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC250	PP071415.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:48 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICV500	PP071421.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:52 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP042225

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	PP050125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071681.D	AR-1016-1 #2	yogesh	5/1/2025 4:05:39 PM	mohammad	5/2/2025 3:22:57	Peak Integrated by Software
AR1660CCC500	PP071681.D	AR-1016-2 #2	yogesh	5/1/2025 4:05:39 PM	mohammad	5/2/2025 3:22:57	Peak Integrated by Software
AR1660CCC500	PP071681.D	Decachlorobiphenyl #2	yogesh	5/1/2025 4:05:39 PM	mohammad	5/2/2025 3:22:57	Peak Integrated by Software
AR1242CCC500	PP071682.D	Decachlorobiphenyl #2	yogesh	5/1/2025 4:05:41 PM	mohammad	5/2/2025 3:23:00	Peak Integrated by Software
AR1248CCC500	PP071683.D	AR-1248-1 #2	yogesh	5/1/2025 4:05:42 PM	mohammad	5/2/2025 3:23:01	Peak Integrated by Software
AR1248CCC500	PP071683.D	Decachlorobiphenyl #2	yogesh	5/1/2025 4:05:42 PM	mohammad	5/2/2025 3:23:01	Peak Integrated by Software
I.BLK	PP071685.D	Decachlorobiphenyl #2	yogesh	5/1/2025 4:05:44 PM	mohammad	5/2/2025 3:23:03	Peak Integrated by Software
AR1254CCC500	PP071699.D	AR-1254-5	yogesh	5/2/2025 7:17:19 AM	mohammad	5/5/2025 4:32:43	Peak Integrated by Software
AR1660CCC500	PP071711.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:08 AM	mohammad	5/5/2025 4:34:11	Peak Integrated by Software
AR1242CCC500	PP071712.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:12 AM	mohammad	5/5/2025 4:34:14	Peak Integrated by Software
AR1248CCC500	PP071713.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:13 AM	mohammad	5/5/2025 4:34:16	Peak Integrated by Software
AR1254CCC500	PP071714.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:15 AM	mohammad	5/5/2025 4:34:20	Peak Integrated by Software
PB167818BL	PP071716.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:16 AM	mohammad	5/5/2025 4:34:24	Peak Integrated by Software

Manual Integration Report

Sequence:

PP050125

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB167818BS	PP071717.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:18 AM	mohammad	5/5/2025 4:34:47	Peak Integrated by Software
PB167818BSD	PP071718.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:20 AM	mohammad	5/5/2025 4:34:30	Peak Integrated by Software
Q1913-04	PP071720.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:22 AM	mohammad	5/5/2025 4:34:53	Peak Integrated by Software
Q1913-04	PP071720.D	Tetrachloro-m-xylene	yogesh	5/2/2025 9:26:22 AM	mohammad	5/5/2025 4:34:53	Peak Integrated by Software
Q1913-04	PP071720.D	Tetrachloro-m-xylene #2	yogesh	5/2/2025 9:26:22 AM	mohammad	5/5/2025 4:34:53	Peak Integrated by Software
AR1248CCC500	PP071727.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:23 AM	mohammad	5/5/2025 4:35:01	Peak Integrated by Software
AR1254CCC500	PP071728.D	Decachlorobiphenyl #2	yogesh	5/2/2025 9:26:25 AM	mohammad	5/5/2025 4:35:04	Peak Integrated by Software

Manual Integration Report

Sequence:	PP050225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071731.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:03 AM	mohammad	5/6/2025 4:43:46	Peak Integrated by Software
AR1242CCC500	PP071732.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:06 AM	mohammad	5/6/2025 4:43:49	Peak Integrated by Software
AR1248CCC500	PP071733.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:08 AM	mohammad	5/6/2025 4:43:52	Peak Integrated by Software
AR1254CCC500	PP071734.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:09 AM	mohammad	5/6/2025 4:43:54	Peak Integrated by Software
I.BLK	PP071735.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:11 AM	mohammad	5/6/2025 4:43:58	Peak Integrated by Software
PB167834BL	PP071737.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:16 AM	mohammad	5/6/2025 4:44:04	Peak Integrated by Software
PB167834BS	PP071738.D	AR-1016-1 #2	yogesh	5/5/2025 8:39:18 AM	mohammad	5/6/2025 4:44:07	Peak Integrated by Software
PB167834BS	PP071738.D	AR-1016-2 #2	yogesh	5/5/2025 8:39:18 AM	mohammad	5/6/2025 4:44:07	Peak Integrated by Software
PB167834BS	PP071738.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:18 AM	mohammad	5/6/2025 4:44:07	Peak Integrated by Software
Q1928-01MS	PP071740.D	AR-1016-1 #2	yogesh	5/5/2025 8:39:22 AM	mohammad	5/6/2025 4:44:09	Peak Integrated by Software
Q1928-01MS	PP071740.D	AR-1016-2 #2	yogesh	5/5/2025 8:39:22 AM	mohammad	5/6/2025 4:44:09	Peak Integrated by Software
Q1928-01MS	PP071740.D	AR-1016-5	yogesh	5/5/2025 8:39:22 AM	mohammad	5/6/2025 4:44:09	Peak Integrated by Software
Q1928-01MSD	PP071741.D	AR-1016-1 #2	yogesh	5/5/2025 8:39:24 AM	mohammad	5/6/2025 4:44:12	Peak Integrated by Software

Manual Integration Report

Sequence:	PP050225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1928-01MSD	PP071741.D	AR-1016-2 #2	yogesh	5/5/2025 8:39:24 AM	mohammad	5/6/2025 4:44:12	Peak Integrated by Software
Q1928-01MSD	PP071741.D	AR-1016-5	yogesh	5/5/2025 8:39:24 AM	mohammad	5/6/2025 4:44:12	Peak Integrated by Software
Q1928-01MSD	PP071741.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:24 AM	mohammad	5/6/2025 4:44:12	Peak Integrated by Software
AR1660CCC500	PP071746.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:32 AM	mohammad	5/6/2025 4:44:23	Peak Integrated by Software
AR1254CCC500	PP071749.D	AR-1254-5	yogesh	5/5/2025 8:39:33 AM	mohammad	5/6/2025 4:44:32	Peak Integrated by Software
Q1913-01	PP071757.D	Tetrachloro-m-xylene	yogesh	5/5/2025 8:39:41 AM	mohammad	5/6/2025 4:44:46	Peak Integrated by Software
Q1913-03	PP071758.D	Decachlorobiphenyl #2	yogesh	5/5/2025 8:39:43 AM	mohammad	5/6/2025 4:44:50	Peak Integrated by Software
Q1913-03	PP071758.D	Tetrachloro-m-xylene	yogesh	5/5/2025 8:39:43 AM	mohammad	5/6/2025 4:44:50	Peak Integrated by Software
Q1913-03	PP071758.D	Tetrachloro-m-xylene #2	yogesh	5/5/2025 8:39:43 AM	mohammad	5/6/2025 4:44:50	Peak Integrated by Software
AR1254CCC500	PP071763.D	AR-1254-5	yogesh	5/5/2025 8:39:47 AM	mohammad	5/6/2025 4:44:58	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071387.D	22 Apr 2025 09:56	YP\AJ	Ok
2	I.BLK	PP071388.D	22 Apr 2025 10:13	YP\AJ	Ok
3	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29	YP\AJ	Ok
4	AR1660ICC750	PP071390.D	22 Apr 2025 10:45	YP\AJ	Ok
5	AR1660ICC500	PP071391.D	22 Apr 2025 11:02	YP\AJ	Ok
6	AR1660ICC250	PP071392.D	22 Apr 2025 11:18	YP\AJ	Ok
7	AR1660ICC050	PP071393.D	22 Apr 2025 11:34	YP\AJ	Ok,M
8	AR1221ICC500	PP071394.D	22 Apr 2025 11:51	YP\AJ	Ok
9	AR1232ICC500	PP071395.D	22 Apr 2025 12:07	YP\AJ	Ok
10	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23	YP\AJ	Ok
11	AR1242ICC750	PP071397.D	22 Apr 2025 12:39	YP\AJ	Ok
12	AR1242ICC500	PP071398.D	22 Apr 2025 12:56	YP\AJ	Ok
13	AR1242ICC250	PP071399.D	22 Apr 2025 13:12	YP\AJ	Ok
14	AR1242ICC050	PP071400.D	22 Apr 2025 13:28	YP\AJ	Ok
15	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45	YP\AJ	Ok
16	AR1248ICC750	PP071402.D	22 Apr 2025 14:01	YP\AJ	Ok
17	AR1248ICC500	PP071403.D	22 Apr 2025 14:17	YP\AJ	Ok
18	AR1248ICC250	PP071404.D	22 Apr 2025 14:33	YP\AJ	Ok
19	AR1248ICC050	PP071405.D	22 Apr 2025 14:50	YP\AJ	Ok
20	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06	YP\AJ	Ok
21	AR1254ICC750	PP071407.D	22 Apr 2025 15:22	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP071408.D	22 Apr 2025 15:38	YP\AJ	Ok
23	AR1254ICC250	PP071409.D	22 Apr 2025 15:55	YP\AJ	Ok
24	AR1254ICC050	PP071410.D	22 Apr 2025 16:11	YP\AJ	Ok,M
25	AR1262ICC500	PP071411.D	22 Apr 2025 16:27	YP\AJ	Ok
26	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44	YP\AJ	Ok,M
27	AR1268ICC750	PP071413.D	22 Apr 2025 17:00	YP\AJ	Ok,M
28	AR1268ICC500	PP071414.D	22 Apr 2025 17:16	YP\AJ	Ok,M
29	AR1268ICC250	PP071415.D	22 Apr 2025 17:33	YP\AJ	Ok,M
30	AR1268ICC050	PP071416.D	22 Apr 2025 17:49	YP\AJ	Ok,M
31	PP042225ICV500	PP071417.D	22 Apr 2025 18:05	YP\AJ	Ok
32	AR1242ICV500	PP071418.D	22 Apr 2025 18:38	YP\AJ	Ok
33	AR1248ICV500	PP071419.D	22 Apr 2025 19:27	YP\AJ	Ok
34	AR1254ICV500	PP071420.D	22 Apr 2025 19:43	YP\AJ	Ok
35	AR1268ICV500	PP071421.D	22 Apr 2025 20:16	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071680.D	01 May 2025 10:35	YP\AJ	Ok
2	AR1660CCC500	PP071681.D	01 May 2025 10:51	YP\AJ	Ok,M
3	AR1242CCC500	PP071682.D	01 May 2025 11:08	YP\AJ	Ok,M
4	AR1248CCC500	PP071683.D	01 May 2025 11:24	YP\AJ	Ok,M
5	AR1254CCC500	PP071684.D	01 May 2025 11:40	YP\AJ	Ok
6	I.BLK	PP071685.D	01 May 2025 11:56	YP\AJ	Ok,M
7	PB167813BL	PP071686.D	01 May 2025 12:13	YP\AJ	Ok,M
8	PB167813BS	PP071687.D	01 May 2025 12:29	YP\AJ	Ok,M
9	Q1919-01	PP071688.D	01 May 2025 12:45	YP\AJ	Ok,M
10	Q1919-02	PP071689.D	01 May 2025 13:01	YP\AJ	Ok
11	Q1920-01	PP071690.D	01 May 2025 13:18	YP\AJ	Ok,M
12	Q1922-01	PP071691.D	01 May 2025 13:34	YP\AJ	Ok
13	Q1922-05	PP071692.D	01 May 2025 13:50	YP\AJ	Ok,M
14	Q1923-01	PP071693.D	01 May 2025 14:07	YP\AJ	Ok,M
15	Q1925-01	PP071694.D	01 May 2025 14:23	YP\AJ	Dilution
16	Q1925-02	PP071695.D	01 May 2025 14:39	YP\AJ	Ok,M
17	AR1660CCC500	PP071696.D	01 May 2025 16:01	YP\AJ	Ok
18	AR1242CCC500	PP071697.D	01 May 2025 16:34	YP\AJ	Ok
19	AR1248CCC500	PP071698.D	01 May 2025 16:50	YP\AJ	Ok
20	AR1254CCC500	PP071699.D	01 May 2025 17:06	YP\AJ	Ok,M
21	I.BLK	PP071700.D	01 May 2025 17:22	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB167821BL	PP071701.D	01 May 2025 17:39	YP\AJ	Ok
23	PB167821BS	PP071702.D	01 May 2025 17:55	YP\AJ	Ok,M
24	Q1916-01	PP071703.D	01 May 2025 18:11	YP\AJ	Ok
25	Q1916-01MS	PP071704.D	01 May 2025 18:28	YP\AJ	Ok,M
26	Q1916-01MSD	PP071705.D	01 May 2025 18:44	YP\AJ	Ok,M
27	Q1917-01	PP071706.D	01 May 2025 19:00	YP\AJ	Ok
28	Q1925-03	PP071707.D	01 May 2025 19:16	YP\AJ	Dilution
29	Q1925-04	PP071708.D	01 May 2025 19:33	YP\AJ	Ok,M
30	Q1925-05	PP071709.D	01 May 2025 19:49	YP\AJ	Ok,M
31	Q1919-08	PP071710.D	01 May 2025 20:05	YP\AJ	Ok,M
32	AR1660CCC500	PP071711.D	01 May 2025 21:27	YP\AJ	Ok,M
33	AR1242CCC500	PP071712.D	01 May 2025 21:59	YP\AJ	Ok,M
34	AR1248CCC500	PP071713.D	01 May 2025 22:16	YP\AJ	Ok,M
35	AR1254CCC500	PP071714.D	01 May 2025 22:32	YP\AJ	Ok,M
36	I.BLK	PP071715.D	01 May 2025 22:48	YP\AJ	Ok
37	PB167818BL	PP071716.D	01 May 2025 23:04	YP\AJ	Ok,M
38	PB167818BS	PP071717.D	01 May 2025 23:21	YP\AJ	Ok,M
39	PB167818BSD	PP071718.D	01 May 2025 23:37	YP\AJ	Ok,M
40	Q1913-02	PP071719.D	01 May 2025 23:53	YP\AJ	Ok
41	Q1913-04	PP071720.D	02 May 2025 00:09	YP\AJ	Ok,M
42	Q1919-05	PP071721.D	02 May 2025 00:26	YP\AJ	Ok
43	Q1926-01	PP071722.D	02 May 2025 00:42	YP\AJ	Ok
44	Q1925-01DL	PP071723.D	02 May 2025 00:58	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q1925-02DL	PP071724.D	02 May 2025 01:15	YP\AJ	Not Ok
46	AR1660CCC500	PP071725.D	02 May 2025 02:52	YP\AJ	Ok
47	AR1242CCC500	PP071726.D	02 May 2025 03:25	YP\AJ	Ok
48	AR1248CCC500	PP071727.D	02 May 2025 03:41	YP\AJ	Ok,M
49	AR1254CCC500	PP071728.D	02 May 2025 03:58	YP\AJ	Ok,M
50	I.BLK	PP071729.D	02 May 2025 04:14	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050225

Review By	yogesh	Review On	5/2/2025 9:59:25 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:45:55 AM
SubDirectory	PP050225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071730.D	02 May 2025 07:58	YP\AJ	Ok
2	AR1660CCC500	PP071731.D	02 May 2025 08:14	YP\AJ	Ok,M
3	AR1242CCC500	PP071732.D	02 May 2025 08:33	YP\AJ	Ok,M
4	AR1248CCC500	PP071733.D	02 May 2025 08:49	YP\AJ	Ok,M
5	AR1254CCC500	PP071734.D	02 May 2025 09:05	YP\AJ	Ok,M
6	I.BLK	PP071735.D	02 May 2025 09:22	YP\AJ	Ok,M
7	Q1925-03DL	PP071736.D	02 May 2025 09:38	YP\AJ	Ok,M
8	PB167834BL	PP071737.D	02 May 2025 13:19	YP\AJ	Ok,M
9	PB167834BS	PP071738.D	02 May 2025 13:36	YP\AJ	Ok,M
10	Q1928-01	PP071739.D	02 May 2025 13:52	YP\AJ	Ok
11	Q1928-01MS	PP071740.D	02 May 2025 14:09	YP\AJ	Ok,M
12	Q1928-01MSD	PP071741.D	02 May 2025 14:25	YP\AJ	Ok,M
13	Q1929-02	PP071742.D	02 May 2025 14:42	YP\AJ	Ok,M
14	Q1929-06	PP071743.D	02 May 2025 14:59	YP\AJ	Ok,M
15	Q1929-10	PP071744.D	02 May 2025 15:15	YP\AJ	Ok,M
16	Q1932-01	PP071745.D	02 May 2025 15:31	YP\AJ	Ok
17	AR1660CCC500	PP071746.D	02 May 2025 16:37	YP\AJ	Ok,M
18	AR1242CCC500	PP071747.D	02 May 2025 17:26	YP\AJ	Ok
19	AR1248CCC500	PP071748.D	02 May 2025 17:42	YP\AJ	Ok
20	AR1254CCC500	PP071749.D	02 May 2025 17:59	YP\AJ	Ok,M
21	I.BLK	PP071750.D	02 May 2025 18:15	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050225

Review By	yogesh	Review On	5/2/2025 9:59:25 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:45:55 AM
SubDirectory	PP050225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1932-03	PP071751.D	02 May 2025 18:31	YP\AJ	Ok,M
23	Q1932-05	PP071752.D	02 May 2025 18:48	YP\AJ	Ok
24	Q1932-07	PP071753.D	02 May 2025 19:04	YP\AJ	Ok
25	Q1933-01	PP071754.D	02 May 2025 19:20	YP\AJ	Ok,M
26	Q1935-01	PP071755.D	02 May 2025 19:37	YP\AJ	Ok
27	Q1935-05	PP071756.D	02 May 2025 19:53	YP\AJ	Ok,M
28	Q1913-01	PP071757.D	02 May 2025 20:09	YP\AJ	Ok,M
29	Q1913-03	PP071758.D	02 May 2025 20:26	YP\AJ	Ok,M
30	Q1913-03DL	PP071759.D	02 May 2025 20:42	YP\AJ	Not Ok
31	AR1660CCC500	PP071760.D	02 May 2025 22:20	YP\AJ	Ok
32	AR1242CCC500	PP071761.D	02 May 2025 23:09	YP\AJ	Ok
33	AR1248CCC500	PP071762.D	02 May 2025 23:25	YP\AJ	Ok
34	AR1254CCC500	PP071763.D	02 May 2025 23:42	YP\AJ	Ok,M
35	I.BLK	PP071764.D	02 May 2025 23:58	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YP\AJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP042225ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:27		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071680.D	01 May 2025 10:35		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071681.D	01 May 2025 10:51		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP071682.D	01 May 2025 11:08		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP071683.D	01 May 2025 11:24		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP071684.D	01 May 2025 11:40		YP\AJ	Ok
6	I.BLK	I.BLK	PP071685.D	01 May 2025 11:56		YP\AJ	Ok,M
7	PB167813BL	PB167813BL	PP071686.D	01 May 2025 12:13		YP\AJ	Ok,M
8	PB167813BS	PB167813BS	PP071687.D	01 May 2025 12:29		YP\AJ	Ok,M
9	Q1919-01	50680	PP071688.D	01 May 2025 12:45		YP\AJ	Ok,M
10	Q1919-02	50681	PP071689.D	01 May 2025 13:01	AR1242 Hit	YP\AJ	Ok
11	Q1920-01	291	PP071690.D	01 May 2025 13:18		YP\AJ	Ok,M
12	Q1922-01	MH-R	PP071691.D	01 May 2025 13:34		YP\AJ	Ok
13	Q1922-05	MH-S	PP071692.D	01 May 2025 13:50		YP\AJ	Ok,M
14	Q1923-01	TR-05-04302025	PP071693.D	01 May 2025 14:07		YP\AJ	Ok,M
15	Q1925-01	AUD-25-0068	PP071694.D	01 May 2025 14:23	AR1242 + AR1260 Hit, Need 5X Dilution	YP\AJ	Dilution
16	Q1925-02	AUD-25-0069	PP071695.D	01 May 2025 14:39	AR1242 + AR1260 Hit,	YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP071696.D	01 May 2025 16:01		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP071697.D	01 May 2025 16:34		YPIAJ	Ok
19	AR1248CCC500	AR1248CCC500	PP071698.D	01 May 2025 16:50		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071699.D	01 May 2025 17:06		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP071700.D	01 May 2025 17:22		YPIAJ	Ok
22	PB167821BL	PB167821BL	PP071701.D	01 May 2025 17:39		YPIAJ	Ok
23	PB167821BS	PB167821BS	PP071702.D	01 May 2025 17:55		YPIAJ	Ok,M
24	Q1916-01	WC-12	PP071703.D	01 May 2025 18:11		YPIAJ	Ok
25	Q1916-01MS	WC-12MS	PP071704.D	01 May 2025 18:28		YPIAJ	Ok,M
26	Q1916-01MSD	WC-12MSD	PP071705.D	01 May 2025 18:44		YPIAJ	Ok,M
27	Q1917-01	MH-JJ	PP071706.D	01 May 2025 19:00		YPIAJ	Ok
28	Q1925-03	AUD-25-0070	PP071707.D	01 May 2025 19:16	AR1242 + AR1260 Hit, Need 5x Dilution	YPIAJ	Dilution
29	Q1925-04	AUD-25-0072	PP071708.D	01 May 2025 19:33		YPIAJ	Ok,M
30	Q1925-05	AUD-25-0073	PP071709.D	01 May 2025 19:49		YPIAJ	Ok,M
31	Q1919-08	0428	PP071710.D	01 May 2025 20:05		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP071711.D	01 May 2025 21:27		YPIAJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PP071712.D	01 May 2025 21:59		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP071713.D	01 May 2025 22:16		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP071714.D	01 May 2025 22:32		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP071715.D	01 May 2025 22:48		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050125

Review By	yogesh	Review On	5/1/2025 11:42:54 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:23:29 AM
SubDirectory	PP050125	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167818BL	PB167818BL	PP071716.D	01 May 2025 23:04		YPIAJ	Ok,M
38	PB167818BS	PB167818BS	PP071717.D	01 May 2025 23:21		YPIAJ	Ok,M
39	PB167818BSD	PB167818BSD	PP071718.D	01 May 2025 23:37		YPIAJ	Ok,M
40	Q1913-02	WC-12-A-202504	PP071719.D	01 May 2025 23:53		YPIAJ	Ok
41	Q1913-04	WC-13-A-202504	PP071720.D	02 May 2025 00:09		YPIAJ	Ok,M
42	Q1919-05	0423	PP071721.D	02 May 2025 00:26		YPIAJ	Ok
43	Q1926-01	TRE-1104	PP071722.D	02 May 2025 00:42		YPIAJ	Ok
44	Q1925-01DL	AUD-25-0068DL	PP071723.D	02 May 2025 00:58	AR1242 + AR1260 Hit,	YPIAJ	Ok,M
45	Q1925-02DL	AUD-25-0069DL	PP071724.D	02 May 2025 01:15	AR1242 + AR1260 Hit; Not Required	YPIAJ	Not Ok
46	AR1660CCC500	AR1660CCC500	PP071725.D	02 May 2025 02:52		YPIAJ	Ok
47	AR1242CCC500	AR1242CCC500	PP071726.D	02 May 2025 03:25		YPIAJ	Ok
48	AR1248CCC500	AR1248CCC500	PP071727.D	02 May 2025 03:41		YPIAJ	Ok,M
49	AR1254CCC500	AR1254CCC500	PP071728.D	02 May 2025 03:58		YPIAJ	Ok,M
50	I.BLK	I.BLK	PP071729.D	02 May 2025 04:14		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050225

Review By	yogesh	Review On	5/2/2025 9:59:25 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:45:55 AM
SubDirectory	PP050225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071730.D	02 May 2025 07:58		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071731.D	02 May 2025 08:14		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP071732.D	02 May 2025 08:33		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP071733.D	02 May 2025 08:49		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP071734.D	02 May 2025 09:05		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP071735.D	02 May 2025 09:22		YP\AJ	Ok,M
7	Q1925-03DL	AUD-25-0070DL	PP071736.D	02 May 2025 09:38	AR1242 + AR1260 Hit	YP\AJ	Ok,M
8	PB167834BL	PB167834BL	PP071737.D	02 May 2025 13:19		YP\AJ	Ok,M
9	PB167834BS	PB167834BS	PP071738.D	02 May 2025 13:36		YP\AJ	Ok,M
10	Q1928-01	MH-Q	PP071739.D	02 May 2025 13:52		YP\AJ	Ok
11	Q1928-01MS	MH-QMS	PP071740.D	02 May 2025 14:09		YP\AJ	Ok,M
12	Q1928-01MSD	MH-QMSD	PP071741.D	02 May 2025 14:25		YP\AJ	Ok,M
13	Q1929-02	WC-A4-02-C	PP071742.D	02 May 2025 14:42		YP\AJ	Ok,M
14	Q1929-06	WC-A1-03-C	PP071743.D	02 May 2025 14:59		YP\AJ	Ok,M
15	Q1929-10	WC-A1-04-C	PP071744.D	02 May 2025 15:15		YP\AJ	Ok,M
16	Q1932-01	COMP-4	PP071745.D	02 May 2025 15:31		YP\AJ	Ok
17	AR1660CCC500	AR1660CCC500	PP071746.D	02 May 2025 16:37		YP\AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP071747.D	02 May 2025 17:26		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP050225

Review By	yogesh	Review On	5/2/2025 9:59:25 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:45:55 AM
SubDirectory	PP050225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP071748.D	02 May 2025 17:42		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071749.D	02 May 2025 17:59		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP071750.D	02 May 2025 18:15		YPIAJ	Ok
22	Q1932-03	COMP-5	PP071751.D	02 May 2025 18:31	AR1254 Hit	YPIAJ	Ok,M
23	Q1932-05	COMP-6	PP071752.D	02 May 2025 18:48	AR1254 Hit	YPIAJ	Ok
24	Q1932-07	COMP-7	PP071753.D	02 May 2025 19:04		YPIAJ	Ok
25	Q1933-01	OK-01-050125	PP071754.D	02 May 2025 19:20		YPIAJ	Ok,M
26	Q1935-01	MH-NN	PP071755.D	02 May 2025 19:37		YPIAJ	Ok
27	Q1935-05	MH-MM	PP071756.D	02 May 2025 19:53		YPIAJ	Ok,M
28	Q1913-01	WC-12-S-202504	PP071757.D	02 May 2025 20:09		YPIAJ	Ok,M
29	Q1913-03	WC-13-S-202504	PP071758.D	02 May 2025 20:26		YPIAJ	Ok,M
30	Q1913-03DL	WC-13-S-202504DL	PP071759.D	02 May 2025 20:42	Not required	YPIAJ	Not Ok
31	AR1660CCC500	AR1660CCC500	PP071760.D	02 May 2025 22:20		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP071761.D	02 May 2025 23:09		YPIAJ	Ok
33	AR1248CCC500	AR1248CCC500	PP071762.D	02 May 2025 23:25		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP071763.D	02 May 2025 23:42		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP071764.D	02 May 2025 23:58		YPIAJ	Ok

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/5/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 05/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135646

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1913-01	WC-12-S-202504	1	1.18	10.32	11.5	9.49	80.5	
Q1913-03	WC-13-S-202504	2	1.19	10.37	11.56	9.83	83.3	
Q1936-01	SMALL-PILE-A	33	1.15	9.94	11.09	10.18	90.8	
Q1936-02	SMALL-PILE-A	34	1.17	9.94	11.11	10.22	91.0	
Q1936-03	SMALL-PILE-B	35	1.18	10.39	11.57	10.93	93.8	
Q1936-04	SMALL-PILE-B	36	1.18	9.88	11.06	10.44	93.7	
Q1936-05	SMALL-PILE-C	37	1.15	10.03	11.18	10.9	97.2	
Q1936-06	SMALL-PILE-C	38	1.15	10.25	11.4	11.01	96.2	
Q1936-07	SMALL-PILE-D	39	1.16	10.41	11.57	10.97	94.2	
Q1936-08	SMALL-PILE-D	40	1.18	9.96	11.14	10.55	94.1	
Q1937-01	LARGE-PILE-A	41	1.15	10.13	11.28	10.7	94.3	
Q1937-02	LARGE-PILE-A	42	1.18	9.58	10.76	10.2	94.2	
Q1937-03	LARGE-PILE-B	43	1.15	10.03	11.18	10.82	96.4	
Q1937-04	LARGE-PILE-B	44	1.17	10.04	11.21	10.72	95.1	
Q1937-05	LARGE-PILE-C	45	1.16	10.19	11.35	10.83	94.9	
Q1937-06	LARGE-PILE-C	46	1.19	10.06	11.25	10.63	93.8	
Q1937-07	LARGE-PILE-D	47	1.18	10.33	11.51	10.95	94.6	
Q1937-08	LARGE-PILE-D	48	1.17	9.75	10.92	10.33	93.9	
Q1937-09	LARGE-PILE-E	49	1.16	10.14	11.3	10.52	92.3	
Q1937-10	LARGE-PILE-E	50	1.18	10.30	11.48	10.76	93.0	
Q1937-11	LARGE-PILE-F	51	1.18	10.48	11.66	10.58	89.7	
Q1937-12	LARGE-PILE-F	52	1.18	10.25	11.43	10.44	90.3	
Q1938-01	LOWER-WALL-PILE-A	53	1.15	10.33	11.48	10.77	93.1	
Q1938-02	LOWER-WALL-PILE-A	54	1.17	10.50	11.67	10.71	90.9	
Q1938-03	LOWER-WALL-PILE-B	55	1.14	10.15	11.29	10.94	96.6	
Q1938-04	LOWER-WALL-PILE-B	56	1.18	9.90	11.08	10.68	96.0	
Q1938-05	LOWER-WALL-PILE-C	57	1.17	10.23	11.4	11.07	96.8	
Q1938-06	LOWER-WALL-PILE-C	58	1.12	9.82	10.94	10.7	97.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/5/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 05/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135646

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1938-07	LOWER-WALL-PILE-D	59	1.19	10.01	11.2	10.7	95.0	
Q1938-08	LOWER-WALL-PILE-D	60	1.17	10.23	11.4	10.86	94.7	
Q1942-05	SVOC-GPC-BLANK	3	1.00	1.00	2.00	2.00	100.0	
Q1942-06	PEST-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
Q1942-07	PEST-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
Q1942-08	PCB-GPC-BLANK	6	1.00	1.00	2.00	2.00	100.0	
Q1942-09	PCB-GPC-BLANK-SPIKE	7	1.00	1.00	2.00	2.00	100.0	
Q1942-10	SVOC-GPC2-BLANK	8	1.00	1.00	2.00	2.00	100.0	
Q1942-11	PEST-GPC2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
Q1942-12	PEST-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
Q1942-13	PCB-GPC2-BLANK	11	1.00	1.00	2.00	2.00	100.0	
Q1942-14	PCB-GCP2-BLANK-SPIKE	12	1.00	1.00	2.00	2.00	100.0	
Q1944-01	MELTED-RUBBER-SAMPLE	13	1.00	1.00	2.00	2.00	100.0	MELTED-RUBBER-SAMPLE
Q1946-01	BB-CONCRETE	14	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q1946-02	BB-CONCRETE	15	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q1947-01	MH-P	16	1.18	10.19	11.37	9.28	79.5	
Q1947-02	MH-P-EPH	17	1.15	10.84	11.99	10.64	87.5	
Q1947-03	MH-P-VOC	18	1.16	10.71	11.87	10.5	87.2	
Q1947-05	MH-O	19	1.14	10.59	11.73	10.52	88.6	
Q1947-06	MH-O-EPH	20	1.12	10.00	11.12	9.91	87.9	
Q1947-07	MH-O-VOC	21	1.17	10.00	11.17	9.95	87.8	
Q1951-01	MH-KK	22	1.18	10.40	11.58	10.01	84.9	
Q1951-02	MH-KK-EPH	23	1.16	10.29	11.45	9.91	85.0	
Q1951-03	MH-KK-VOC	24	1.17	10.25	11.42	9.86	84.8	
Q1952-01	CLEAN-FILL	25	1.18	10.49	11.67	8.98	74.4	
Q1952-02	CLEAN-FILL-E2	26	1.13	10.10	11.23	8.47	72.7	
Q1953-01	AUD-25-0075	27	1.00	1.00	2.00	2.00	100.0	pile
Q1953-02	AUD-25-0076	28	1.00	1.00	2.00	2.00	100.0	pile



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/5/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:25
Out Date: 05/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135646

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1953-03	AUD-25-0077	29	1.00	1.00	2.00	2.00	100.0	oilc
Q1954-01	BC260874-1-1	30	1.00	1.00	2.00	2.00	100.0	oilc
Q1954-02	BC260874-1-2	31	1.00	1.00	2.00	2.00	100.0	oilc
Q1955-01	LAW-25-0066	32	1.00	1.00	2.00	2.00	100.0	stone sample, 100 % solids

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

173546

WorkList Name : %1-050225

WorkList ID : 189289

Department : Wet-Chemistry

Date : 05-02-2025 14:36:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-01	WC-12-S-202504	Solid	Percent Solids	Cool 4 deg C	AECO02	L21	04/29/2025	Chemtech -SO
Q1913-03	WC-13-S-202504	Solid	Percent Solids	Cool 4 deg C	AECO02	L21	04/29/2025	Chemtech -SO
Q1936-01	SMALL-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-02	SMALL-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-03	SMALL-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-04	SMALL-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-05	SMALL-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-06	SMALL-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-07	SMALL-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1936-08	SMALL-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L51	05/01/2025	Chemtech -SO
Q1937-01	LARGE-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-02	LARGE-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-03	LARGE-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-04	LARGE-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-05	LARGE-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-06	LARGE-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-07	LARGE-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-08	LARGE-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-09	LARGE-PILE-E	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-10	LARGE-PILE-E	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1937-11	LARGE-PILE-F	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO

Date/Time 05/02/25 15:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05/02/25 17:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

172546

WorkList Name : %1-050225

WorkList ID : 189289

Department : Wet-Chemistry

Date : 05-02-2025 14:36:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1937-12	LARGE-PILE-F	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-01	LOWER-WALL-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-02	LOWER-WALL-PILE-A	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-03	LOWER-WALL-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-04	LOWER-WALL-PILE-B	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-05	LOWER-WALL-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-06	LOWER-WALL-PILE-C	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-07	LOWER-WALL-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1938-08	LOWER-WALL-PILE-D	Solid	Percent Solids	Cool 4 deg C	SAXT01	L41	05/01/2025	Chemtech -SO
Q1942-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	05/01/2025	Chemtech -SO
Q1942-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1942-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L41	04/25/2025	Chemtech -SO
Q1944-01	MELTED-RUBBER-SAMPLE	Solid	Percent Solids	Cool 4 deg C	METE01	L41	05/02/2025	Chemtech -SO
Q1946-01	BB-CONCRETE	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO

Date/Time 05/02/25 15:10

Raw Sample Received by: *Se Lee*

Raw Sample Relinquished by: *Se Lee*

Date/Time 05/02/25

Raw Sample Received by: *Se Lee*

Raw Sample Relinquished by: *Se Lee*

WORKLIST(Hardcopy Internal Chain)

8/23/25

WorkList Name : %1-050225

WorkList ID : 189289

Department : Wet-Chemistry

Date : 05-02-2025 14:36:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1946-02	BB-CONCRETE	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-01	MH-P	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-02	MH-P-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-03	MH-P-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-05	MH-O	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-06	MH-O-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1947-07	MH-O-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1951-01	MH-KK	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/02/2025	Chemtech -SO
Q1951-02	MH-KK-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/02/2025	Chemtech -SO
Q1951-03	MH-KK-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/02/2025	Chemtech -SO
Q1952-01	CLEAN-FILL	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/02/2025	Chemtech -SO
Q1952-02	CLEAN-FILL-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/02/2025	Chemtech -SO
Q1953-01	AUD-25-0075	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1953-02	AUD-25-0076	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1953-03	AUD-25-0077	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1954-01	BC260874-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1954-02	BC260874-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO
Q1955-01	LAW-25-0066	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/02/2025	Chemtech -SO

Date/Time 05/02/25 15:10

Raw Sample Received by: 88 WOC,

Raw Sample Relinquished by: 88 WOC,

Date/Time 05/02/25 17:10

Raw Sample Received by: 88 WOC,

Raw Sample Relinquished by: 88 WOC,

SOP ID: M3510C,3580A-Extraction PCB-14

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 05/01/2025

Matrix : Water

Extraction Start Time : 08:45

Weigh By: N/A

Extraction By: RS

Extraction End Date : 05/01/2025

Balance check: N/A

Filter By: RS

Extraction End Time : 13:25

Balance ID: N/A

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: E3880

Hood ID: 4,6,7

Supervisor By : RUPESH

Extraction Method: ☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3926
Baked Na2SO4	N/A	EP2607
Hexane	N/A	E3928
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40BTS723, Q1926-01 Used Limited volume as sample is oily. Q1913-02,04 pH Adjusted with 1:1 H2SO4.

KD Bath ID: WATER BATH-1,2

Envap ID: NEVAP-02

KD Bath Temperature: 60 °C


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/1/25	RS (B4-Gcb)	Y-Pipest IP40.
13:30	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction PCB-14

Concentration Date: 05/01/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167818BL	ABLK818	PCB	1000	6	RUPESH	Evelyn	10			SEP-1
PB167818BS	ALCS818	PCB	1000	6	RUPESH	Evelyn	10			2
PB167818BS D	ALCSD818	PCB	1000	6	RUPESH	Evelyn	10			3
Q1913-02	WC-12-A-202504	PCB	980	13	RUPESH	Evelyn	10	H		4
Q1913-04	WC-13-A-202504	PCB	990	13	RUPESH	Evelyn	10	H		5
Q1919-05	0423	PCB	990	6	RUPESH	Evelyn	10	B		6
Q1919-08	0428	PCB	990	6	RUPESH	Evelyn	10	B		7
Q1925-04	AUD-25-0072	PCB	1000	6	RUPESH	Evelyn	10	B		8
Q1925-05	AUD-25-0073	PCB	1000	6	RUPESH	Evelyn	10	B		9
Q1926-01	TRE-1104	PCB	100	6	RUPESH	Evelyn	10	B		10



RS
5/1

8/18/25
5m:8

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1926 WorkList ID : 189249 Department : Extraction Date : 05-01-2025 08:40:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-02	WC-12-A-202504	Water	PCB	Cool 4 deg C	AECO02	L21	04/29/2025	8082A
Q1913-04	WC-13-A-202504	Water	PCB	Cool 4 deg C	AECO02	L21	04/29/2025	8082A
Q1919-05	0423	Water	PCB	Cool 4 deg C	PSEG03	L41	04/30/2025	8082A
Q1919-08	0428	Water	PCB	Cool 4 deg C	PSEG03	L41	04/30/2025	8082A
Q1925-04	AUD-25-0072	Water	PCB	Cool 4 deg C	PSEG03	L31	04/30/2025	8082A
Q1925-05	AUD-25-0073	Water	PCB	Cool 4 deg C	PSEG03	L31	04/30/2025	8082A
Q1926-01	TRE-1104	Water	PCB	Cool 4 deg C	PSEG03	L31	04/30/2025	8082A

Date/Time 5/1/25 8:40
Raw Sample Received by: RS (Ext 64)
Raw Sample Relinquished by: CP

Date/Time 5/1/25 9:15
Raw Sample Received by: CP
Raw Sample Relinquished by: RS (Ext 64)

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Welgh By: EH

Balance check: RJ

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RS

Filter By: EH

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 05/02/2025

Extraction Start Time : 09:45

Extraction End Date : 05/02/2025

Extraction End Time : 12:45

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2601
Baked Na2SO4	N/A	EP2607
Sand	N/A	E2865
Hexane	N/A	E3928
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS723.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/2/25	RS (Ext Lab)	R. Pest-PCB Lab
12:50	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 05/02/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167834BL	ABLK834	PCB	30.01	N/A	RUPESEH	Evelyn	10			U7-1
PB167834BS	ALCS834	PCB	30.03	N/A	RUPESEH	Evelyn	10			2
Q1913-01	WC-12-S-202504	PCB	30.07	N/A	RUPESEH	Evelyn	10	E		3
Q1913-03	WC-13-S-202504	PCB	30.08	N/A	RUPESEH	Evelyn	10	E		4
Q1928-01	MH-Q	PCB	30.05	N/A	RUPESEH	Evelyn	10	E		5
Q1928-01MS	MH-QMS	PCB	30.04	N/A	RUPESEH	Evelyn	10	E		6
Q1928-01MS D	MH-QMSD	PCB	30.08	N/A	RUPESEH	Evelyn	10	E		U3-1
Q1929-02	WC-AA-02-C	PCB	30.05	N/A	RUPESEH	Evelyn	10	D		2
Q1929-06	WC-AA-03-C	PCB	30.01	N/A	RUPESEH	Evelyn	10	D		3
Q1929-10	WC-AA-04-C	PCB	30.03	N/A	RUPESEH	Evelyn	10	D		4
Q1932-01	COMP-4	PCB	30.05	N/A	RUPESEH	Evelyn	10	F		5
Q1932-03	COMP-5	PCB	30.08	N/A	RUPESEH	Evelyn	10	E		6
Q1932-05	COMP-6	PCB	30.02	N/A	RUPESEH	Evelyn	10	F		U6-1
Q1932-07	COMP-7	PCB	30.09	N/A	RUPESEH	Evelyn	10	F		2
Q1933-01	OK-01-050125	PCB	30.07	N/A	RUPESEH	Evelyn	10	E		3
Q1935-01	MH-NN	PCB	30.01	N/A	RUPESEH	Evelyn	10	D		4
Q1935-05	MH-MM	PCB	30.03	N/A	RUPESEH	Evelyn	10	D		5

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1929 WorkList ID : 189282 Department : Extraction Date : 05-02-2025 09:00:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-01	WC-12-S-202504	Solid	PCB	Cool 4 deg C	AECO02	L21	04/29/2025	8082A
Q1913-03	WC-13-S-202504	Solid	PCB	Cool 4 deg C	AECO02	L21	04/29/2025	8082A
Q1928-01	MH-Q	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1929-02	WC-A4-02-C	Solid	PCB	Cool 4 deg C	ENTA05	L41	05/29/2025	8082A
Q1929-06	WC-A4-03-C	Solid	PCB	Cool 4 deg C	ENTA05	L41	05/30/2025	8082A
Q1929-10	WC-A4-04-C	Solid	PCB	Cool 4 deg C	ENTA05	L41	05/30/2025	8082A
Q1932-01	COMP-4	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1932-03	COMP-5	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1932-05	COMP-6	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1932-07	COMP-7	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1933-01	OK-01-050125	Solid	PCB	Cool 4 deg C	PSEG05	L31	05/01/2025	8082A
Q1935-01	MH-NN	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A
Q1935-05	MH-MM	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/01/2025	8082A

Date/Time 5/2/25 9:40
Raw Sample Received by: RS (Ext-Verb)
Raw Sample Relinquished by: CP SR

Date/Time 5/2/25 10:10
Raw Sample Received by: CP SR
Raw Sample Relinquished by: RS (Ext-Verb)

Prep Standard - Chemical Standard Summary

Order ID : Q1913

Test : PCB

Prepbatch ID : PB167818,PB167834,

Sequence ID/Qc Batch ID: PP050125,PP050225,

Standard ID :

EP2565,EP2601,EP2607,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24460,PP24461,

Chemical ID :

E2865,E3551,E3804,E3877,E3916,E3917,E3926,E3928,M5173,P 11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13355,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W 3112,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2601	04/07/2025	10/03/2025	Rajesh Parikh	None	None	Riteshkumar Patel 04/07/2025

FROM 8000.00000ml of E3916 + 8000.00000ml of E3917 = Final Quantity: 8000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2607	04/25/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 04/25/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24460	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 1.00000ml of P13355 + 999.00000ml of E3917 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
04/16/2025								
<u>FROM</u> 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3916

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	10/08/2025	04/08/2025 / Rajesh	02/07/2025 / Rajesh	E3926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	10/22/2025	04/18/2025 / RUPESH	04/16/2025 / RUPESH	E3928

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	10/11/2025	04/11/2025 / Abdul	04/22/2024 / Abdul	P13355

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16\%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 **E 3551**

RC-02-01, Ed. 3

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A

Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Harout Sahagian - Quality Control Manager - Fair Lawn

Recd by RP on 3/31/25

E 3946

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.
 *Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor



Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date: 2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	99.9%
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.8 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3926

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors[™]**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3928



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials U.S.

Hydrochloric Acid, 36.5–38.0%
 BAKER INSTRA-ANALYZED® Reagent
 For Trace Metal Analysis



Material No.: 9530-33
 Batch No.: 0000281827
 Manufactured Date: 2021/03/30
 Retest Date: 2026/03/29
 Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
 Avantor Performance Materials, LLC
 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration (µg/mL): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5µL, Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1210967
P1210977
P1210973
P1210973

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial #

1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL.):
200.0

Balance Uncertainty
0.010

Flask Uncertainty

SDS Information
(Solvent Safety Info. On Attached pg.)

Expanded Uncertainty
(+/-) (µg/mL)

Actual Conc (µg/mL)

Actual Weight(g)

Target Weight(g)

Purity (%)

Nominal Conc (µg/mL)


Lot Number

RM#

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL):
200.0

Balance Uncertainty
0.010

Flask Uncertainty
0.010

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL):
200.0


Balance Uncertainty
0.010

Flask Uncertainty
0.010

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE

P12956 / Y.P.
12/18/23
P12957

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 5E-05 Balance Uncertainty
20.0 0.003 Flask Uncertainty


Note: Aroclor 1254 is a mix of isomers.

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty (mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
									(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA) LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = Edaq Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025

Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dylan Murphy
Dylan Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } (2)
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902 } Y.P.
P13703 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13882

↓

AJ
01/28/25

P13883

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 Balance Serial # C322230531


Britiany Federinko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: AECOM
ADDRESS: 605 3rd Ave
CITY NEW YORK STATE: NY ZIP: 10158
ATTENTION: Amit Haryani
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Meeker Ave Superfund
PROJECT NO.: 60705866 LOCATION: Brooklyn
PROJECT MANAGER: Amit Haryani
e-mail: amit.haryani@aecom.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY SAME STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) 3 days DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. TCLP VOA
2. TCLP BBA
3. TCLP
4. Paint Filter
5. Metals Mercury
6. Reactive organics
7. Reactive sulfide
8. pH
9. PCBs Pesticides
PCB+Fluoride

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	WC-12-A-202504	WATER	X		4/29/25	7:00	11	✓	✓	✓	✓	✓	✓	✓	✓	✓	
2.	WC-12-S-202504	SOIL	X		4/29/25	7:00	7	✓	✓								
3.	WC-13-A-202504	WATER	X		4/29/25	7:50	11	✓	✓	✓	✓	✓	✓	✓	✓	✓	
4.	WC-13-S-202504	SOIL	X		4/29/25	7:50	7	✓	✓								
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. Ewa Pencyliviz	DATE/TIME: 1215 4/29/25	RECEIVED BY: 1. [Signature] 4-29-25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.2°C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: Temp 2.2°C (Adjustment factor #1) IR Gun #1
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME: 1650 4-29-25	RECEIVED BY: 3.	Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

From: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Sent: Thursday, May 01, 2025 8:04 PM
To: Yazmeen Gomez
Cc: Fairbanks, Peter; Haryani, Amit; Tan, Paul
Subject: RE: Chemtech/Alliance - Samples - 11/14

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Dear Yazmeen,

I believe that I missed parameter (corrosivity) for samples picked up on Tuesday. Below is a table with parameters for each sample. I believe that this is standard analysis but just want to make sure that everything is correct.

Sample Identifier/location	CLP Sample ID	Matrix	Analyses	Sample Type	Sample Date	Date Shipped
Alliance Technical Group Company						
WC-12-S-202504	NA	IDW - SOLID	Paint Filter 9095B Flashpoint 1010 PCB 8082A TCLP BNA 8270E TCLP Pesticide 8081B Corrosivity 9045C Reactive Cyanide 9014 Reactive Sulfide 9034 TCLP Extraction 1311 TCLP ICP Metals 6010D TCLP VOAS 8260 TCLP MERCURY – 7470A	IDW Sample	4/29/25	4/29/25

WC-12-A-202504	NA	IDW - AQUEOUS	TCLP BNA - 8270E pH-9040C FLASH POINT- 1010A PCB- 8082A TCLP PESTICIDES- 8081B REACTIVE CYANIDE - 9012B REACTIVE SULFIDE- 9034 PAINT FILTER-9095B TCLP VOAS- 8260D TCLP ICP METAL - 6010B TCLP MERCURY – 7470A	IDW Sample	4/29/25	4/29/25
WC-13-S-202504	NA	IDW - SOLID	Paint Filter 9095B Flashpoint 1010 PCB 8082A TCLP BNA 8270E TCLP Pesticide 8081B Corrosivity 9045C Reactive Cyanide 9014 Reactive Sulfide 9034 TCLP Extraction 1311 TCLP ICP Metals 6010D TCLP VOAS 8260 TCLP MERCURY – 7470A	IDW Sample	4/29/25	4/29/25
WC-13-A-202504	NA	IDW - AQUEOUS	TCLP BNA - 8270E pH-9040C FLASH POINT- 1010A PCB- 8082A TCLP PESTICIDES- 8081B REACTIVE CYANIDE - 9012B REACTIVE SULFIDE- 9034 PAINT FILTER-9095B TCLP VOAS- 8260D TCLP ICP METAL - 6010B TCLP MERCURY – 7470A	IDW Sample	4/29/25	4/29/25

I apologize for any inconvenience.

Ewa

From: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>

Sent: Tuesday, April 29, 2025 9:59 AM

To: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>

Cc: Haryani, Amit <Amit.Haryani@aecom.com>; Tan, Paul <Paul.Tan@aecom.com>; Goldstein, Charlotte <charlotte.goldstein@aecom.com>

Subject: RE: Chemtech/Alliance - Samples - 11/14

Good morning,

5 days if you need results in a week or 10 days if you need standard TAT.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Sent: Tuesday, April 29, 2025 5:53 AM
To: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Cc: Haryani, Amit <Amit.Haryani@aecom.com>; Tan, Paul <Paul.Tan@aecom.com>; Goldstein, Charlotte <charlotte.goldstein@aecom.com>
Subject: Re: Chemtech/Alliance - Samples - 11/14

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning,

Can you please kindly help me and let me know what type of way turnaround I have to place on COC in order to receive the result in one week?

I appreciate your help.

Ewa

Get [Outlook for iOS](#)

From: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Sent: Monday, April 28, 2025 3:21:51 PM
To: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Cc: Haryani, Amit <Amit.Haryani@aecom.com>; Tan, Paul <Paul.Tan@aecom.com>; Goldstein, Charlotte <charlotte.goldstein@aecom.com>
Subject: RE: Chemtech/Alliance - Samples - 11/14

Good afternoon,

No problem, Pick up confirmed for tomorrow afternoon.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Sent: Monday, April 28, 2025 2:24 PM
To: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Cc: Haryani, Amit <Amit.Haryani@aecom.com>; Tan, Paul <Paul.Tan@aecom.com>; Goldstein, Charlotte <charlotte.goldstein@aecom.com>
Subject: Re: Chemtech/Alliance - Samples - 11/14

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Is there any chance we can have samples pick up tmrr afternoon? Adres is 3660 Meeker Ave, however I may not be in the storage so please let the courier know to call me prior his arrival. I will let him know where to meet me.

If tomorrow doesn't work let me know if Thursday works for you.

I appreciate your help.

Ewa

Get [Outlook for iOS](#)

From: Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>
Sent: Tuesday, April 15, 2025 9:37:20 AM
To: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Cc: Haryani, Amit <Amit.Haryani@aecom.com>; Tan, Paul <Paul.Tan@aecom.com>; Goldstein, Charlotte <charlotte.goldstein@aecom.com>
Subject: RE: Chemtech/Alliance - Samples - 11/14

Good morning,

Bottle order delivery confirmed – the courier will give you a call with an ETA.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com   

From: "Poncyliusz, Ewa" <Ewa.Poncyliusz@aecom.com>
Date: April 15, 2025 at 7:08:41 AM EDT
To: Kiran Saleem <kiran.saleem@alliancetg.com>, "Haryani, Amit" <Amit.Haryani@aecom.com>
Cc: "Tan, Paul" <Paul.Tan@aecom.com>, "Goldstein, Charlotte" <charlotte.goldstein@aecom.com>
Subject: RE: Chemtech/Alliance - Samples - 11/14

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning,

Can we please have deliver 2 sets on aqueous waste characterization bottlewares for our Meeker Ave Superfund project? Delivery address is 3066 Review Ave. Please make sure that your courier will call me prior arrival because most likely I will be at the field nearby.

Thank you

Ewa

From: Kiran Saleem <Kiran.Saleem@alliancetg.com>
Sent: Tuesday, March 4, 2025 1:06 PM
To: Haryani, Amit <Amit.Haryani@aecom.com>
Cc: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Subject: Re: Chemtech/Alliance - Samples - 11/14

Hello Amit and Ewa,

Hoping you are doing well. I am reaching out to check with you if you have any current or future projects with which we can help you with. Please let me know.

Thank you!

NOTE: Chemtech is now an Alliance Technical Group company. Please add AllianceTG.com to your safe senders list to ensure receipt of important emails.

Regards,



Kiran Saleem
Project Manager
Alliance Technical Group
Main: 908-789-8900
Direct: 908-728-3148
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Kiran Saleem <Kiran.Saleem@alliancetg.com>
Sent: Friday, November 15, 2024 4:17 PM
To: Haryani, Amit <Amit.Haryani@aecom.com>
Cc: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Subject: Re: Chemtech/Alliance - Samples - 11/14

Amit,

Thankyou.

NOTE: Chemtech is now an Alliance Technical Group company. Please add AllianceTG.com to your safe senders list to ensure receipt of important emails.

Regards,

<image001.png>

Kiran Saleem
Project Manager
Alliance Technical Group
Main: 908-789-8900
Direct: 908-728-3148
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Haryani, Amit <Amit.Haryani@aecom.com>
Sent: Friday, November 15, 2024 1:44 PM
To: Kiran Saleem <Kiran.Saleem@alliancetg.com>
Cc: Poncyliusz, Ewa <Ewa.Poncyliusz@aecom.com>
Subject: RE: Chemtech/Alliance - Samples - 11/14

Dear Kiran

Please proceed as detailed below in your email.

Regards
Amit

From: Kiran Saleem <Kiran.Saleem@alliancetg.com>
Sent: Friday, November 15, 2024 10:04 AM
To: Haryani, Amit <Amit.Haryani@aecom.com>
Subject: Chemtech/Alliance - Samples - 11/14

Good Morning Amit,

I hope this message finds you well. I wanted to inform you that the samples WC-10-A-202411 and WC-11-A-202411, which we received on 11/14, have been scheduled for cyanide analysis. As part of our standard protocol, we will be preserving a small volume of each sample for cyanide testing prior to conducting the analysis.

We are notifying you in advance as part of our procedure to keep you informed before proceeding with this step.

Thanks.

NOTE: Chemtech is now an Alliance Technical Group company. Please add AllianceTG.com to your safe senders list to ensure receipt of important emails.

Regards,

<image001.png>

Kiran Saleem
Project Manager
Alliance Technical Group
Main: 908-789-8900
Direct: 908-728-3148
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488