

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088369.D Time Analyzed: 12:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.073	8.975	9.175	46.430	50.000	-7.1
Endrin	6.575	6.476	6.676	52.080	50.000	4.2
gamma-BHC (Lindane)	4.331	4.232	4.432	54.230	50.000	8.5
Heptachlor	4.930	4.831	5.031	53.020	50.000	6.0
Heptachlor epoxide	5.691	5.592	5.792	53.870	50.000	7.7
Methoxychlor	7.494	7.395	7.595	45.160	50.000	-9.7
Tetrachloro-m-xylene	3.551	3.452	3.652	52.750	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: PSTDCCC050 Data File : PD088369.D Time Analyzed: 12:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.075	7.977	8.177	41.500	50.000	-17.0
Endrin	5.791	5.693	5.893	46.350	50.000	-7.3
gamma-BHC (Lindane)	3.731	3.633	3.833	48.020	50.000	-4.0
Heptachlor	4.084	3.986	4.186	46.300	50.000	-7.4
Heptachlor epoxide	4.875	4.777	4.977	48.240	50.000	-3.5
Methoxychlor	6.756	6.659	6.859	41.540	50.000	-16.9
Tetrachloro-m-xylene	2.882	2.783	2.983	48.750	50.000	-2.5