

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088415.D Time Analyzed: 10:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.072	8.975	9.175	50.520	50.000	1.0
Endrin	6.574	6.476	6.676	53.970	50.000	7.9
gamma-BHC (Lindane)	4.330	4.232	4.432	56.130	50.000	12.3
Heptachlor	4.929	4.831	5.031	54.910	50.000	9.8
Heptachlor epoxide	5.691	5.592	5.792	55.230	50.000	10.5
Methoxychlor	7.492	7.395	7.595	49.610	50.000	-0.8
Tetrachloro-m-xylene	3.550	3.452	3.652	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088415.D Time Analyzed: 10:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.074	7.977	8.177	44.790	50.000	-10.4
Endrin	5.790	5.693	5.893	47.530	50.000	-4.9
gamma-BHC (Lindane)	3.730	3.633	3.833	48.340	50.000	-3.3
Heptachlor	4.084	3.986	4.186	46.860	50.000	-6.3
Heptachlor epoxide	4.874	4.777	4.977	48.350	50.000	-3.3
Methoxychlor	6.756	6.659	6.859	43.090	50.000	-13.8
Tetrachloro-m-xylene	2.881	2.783	2.983	48.350	50.000	-3.3