

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL05 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088422.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.080	8.975	9.175	51.710	50.000	3.4
Endrin	6.580	6.476	6.676	55.530	50.000	11.1
gamma-BHC (Lindane)	4.336	4.232	4.432	57.260	50.000	14.5
Heptachlor	4.935	4.831	5.031	55.950	50.000	11.9
Heptachlor epoxide	5.697	5.592	5.792	55.910	50.000	11.8
Methoxychlor	7.499	7.395	7.595	51.610	50.000	3.2
Tetrachloro-m-xylene	3.555	3.452	3.652	54.400	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL05 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088422.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	46.850	50.000	-6.3
Endrin	5.791	5.693	5.893	49.370	50.000	-1.3
gamma-BHC (Lindane)	3.731	3.633	3.833	49.300	50.000	-1.4
Heptachlor	4.084	3.986	4.186	48.290	50.000	-3.4
Heptachlor epoxide	4.875	4.777	4.977	49.290	50.000	-1.4
Methoxychlor	6.757	6.659	6.859	45.930	50.000	-8.1
Tetrachloro-m-xylene	2.881	2.783	2.983	48.930	50.000	-2.1