

Surrogate Summary

SDG No.: Q1913

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	22.7	113		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
I.BLK-PD088368.D	PIBLK-PD088368.D	Decachlorobiphenyl	1	20	20.7	103		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	18.9	94		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
PB167820BL	PB167820BL	Decachlorobiphenyl	1	20	18.6	93		43	140
		Tetrachloro-m-xylene	1	20	18.0	90		77	126
		Decachlorobiphenyl	2	20	16.8	84		43	140
		Tetrachloro-m-xylene	2	20	18.5	92		77	126
PB167820BS	PB167820BS	Decachlorobiphenyl	1	20	20.5	102		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	18.7	93		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
PB167774TB	PB167774TB	Decachlorobiphenyl	1	20	19.1	96		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	17.2	86		43	140
		Tetrachloro-m-xylene	2	20	18.9	94		77	126
PB167805TB	PB167805TB	Decachlorobiphenyl	1	20	18.9	95		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	17.1	85		43	140
		Tetrachloro-m-xylene	2	20	18.9	94		77	126
Q1913-02	WC-12-A-202504	Decachlorobiphenyl	1	20	19.2	96		43	140
		Tetrachloro-m-xylene	1	20	20.1	100		77	126
		Decachlorobiphenyl	2	20	17.3	86		43	140
		Tetrachloro-m-xylene	2	20	19.5	97		77	126
Q1913-04	WC-13-A-202504	Decachlorobiphenyl	1	20	19.3	97		43	140
		Tetrachloro-m-xylene	1	20	17.7	89		77	126
		Decachlorobiphenyl	2	20	16.6	83		43	140
		Tetrachloro-m-xylene	2	20	18.4	92		77	126
Q1901-08MS	B-167-SB01MS	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	19.0	95		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
Q1901-08MSD	B-167-SB01MSD	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	19.0	95		43	140
		Tetrachloro-m-xylene	2	20	18.9	94		77	126
I.BLK-PD088383.D	PIBLK-PD088383.D	Decachlorobiphenyl	1	20	21.0	105		43	140

Surrogate Summary

SDG No.: Q1913

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088383.D	PIBLK-PD088383.D	Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	19.9	99		43	140
I.BLK-PD088406.D	PIBLK-PD088406.D	Tetrachloro-m-xylene	2	20	19.9	100		77	126
		Decachlorobiphenyl	1	20	16.6	83		43	140
		Tetrachloro-m-xylene	1	20	16.2	81		77	126
		Decachlorobiphenyl	2	20	14.9	75		43	140
		Tetrachloro-m-xylene	2	20	15.7	78		77	126
PB167848BL	PB167848BL	Decachlorobiphenyl	1	20	20.5	103		43	140
		Tetrachloro-m-xylene	1	20	19.3	96		77	126
		Decachlorobiphenyl	2	20	18.5	92		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
PB167848BS	PB167848BS	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	20.7	104		77	126
Q1913-01	WC-12-S-202504	Decachlorobiphenyl	1	20	17.6	88		43	140
		Tetrachloro-m-xylene	1	20	23.6	118		77	126
		Decachlorobiphenyl	2	20	15.8	79		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
		Decachlorobiphenyl	1	20	14.1	70		43	140
Q1913-03	WC-13-S-202504	Tetrachloro-m-xylene	1	20	22.6	113		77	126
		Decachlorobiphenyl	2	20	12.6	63		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
		Decachlorobiphenyl	1	20	17.3	86		43	140
		Tetrachloro-m-xylene	1	20	16.2	81		77	126
I.BLK-PD088414.D	PIBLK-PD088414.D	Decachlorobiphenyl	2	20	15.6	78		43	140
		Tetrachloro-m-xylene	2	20	15.5	78		77	126
		Decachlorobiphenyl	1	20	22.3	112		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
Q1929-03MS	WC-A4-02-CMS	Decachlorobiphenyl	2	20	19.9	100		43	140
		Tetrachloro-m-xylene	2	20	15.5	77		77	126
		Decachlorobiphenyl	1	20	22.3	112		43	140
Q1929-03MSD	WC-A4-02-CMSD	Tetrachloro-m-xylene	1	20	18.9	95		77	126
		Decachlorobiphenyl	2	20	20.1	101		43	140
		Tetrachloro-m-xylene	2	20	15.8	79		77	126
		Decachlorobiphenyl	1	20	19.4	97		43	140
I.BLK-PD088421.D	PIBLK-PD088421.D	Tetrachloro-m-xylene	1	20	18.2	91		77	126
		Decachlorobiphenyl	2	20	18.0	90		43	140
		Tetrachloro-m-xylene	2	20	17.3	87		77	126
		Decachlorobiphenyl	1	20	19.4	97		43	140