

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: **8081B** **Datafile :** PD088410.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167848BS	gamma-BHC (Lindane)	0.5	0.54	ug/L	108				82	129	
	Heptachlor	0.5	0.54	ug/L	108				79	127	
	Heptachlor epoxide	0.5	0.55	ug/L	110				81	124	
	Endrin	0.5	0.53	ug/L	106				81	128	
	Methoxychlor	0.5	0.49	ug/L	98				78	108	