

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO02
Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913
Instrument ID: BNA_P Calibration Date(s): 04/14/2025 04/14/2025
Calibration Time(s): 11:06 17:13

LAB FILE ID:		RRF2.5 = BP024275.D		RRF005 = BP024276.D		RRF010 = BP024277.D		
		RRF020 = BP024278.D		RRF040 = BP024279.D		RRF050 = BP024280.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.265	1.280	1.422	1.396	1.333	1.351	4.4
2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.210	4.9
Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.656	6.7
1,4-Dichlorobenzene		1.496	1.489	1.550	1.494	1.436	1.475	3.3
2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.075	8.0
3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.491	9.6
Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.351	4.8
Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	2.9
Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.347	4.0
Hexachlorobutadiene		0.181	0.181	0.186	0.189	0.181	0.183	1.6
2,4,6-Trichlorophenol		0.301	0.335	0.369	0.389	0.386	0.366	9.5
2-Fluorobiphenyl		1.332	1.334	1.359	1.366	1.263	1.306	4.1
2,4,5-Trichlorophenol		0.332	0.371	0.409	0.427	0.429	0.405	9.6
2,4-Dinitrotoluene		0.334	0.368	0.439	0.448	0.442	0.421	11.8
2,4,6-Tribromophenol		0.235	0.256	0.278	0.289	0.289	0.277	8.3
Hexachlorobenzene		0.247	0.247	0.260	0.271	0.261	0.260	3.7
Pentachlorophenol		0.143	0.155	0.175	0.196	0.197	0.181	13.3
Terphenyl-d14		0.967	0.998	1.055	1.032	1.005	0.992	4.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1