

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913
 Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17
 Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.