

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.