

LAB CHRONICLE

OrderID:	Q1913	OrderDate:	4/29/2025 2:15:00 PM
Client:	AECOM	Project:	Meeker Ave Plumes Superfund Site RI FS
Contact:	Amit Haryani	Location:	L21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1913-01	WC-12-S-202504	TCLP	TCLP BNA	8270E	04/29/25	04/30/25	05/01/25	04/29/25
Q1913-02	WC-12-A-202504	TCLP	TCLP BNA	8270E	04/29/25	04/30/25	05/01/25	04/29/25
Q1913-03	WC-13-S-202504	TCLP	TCLP BNA	8270E	04/29/25	04/30/25	05/01/25	04/29/25
Q1913-04	WC-13-A-202504	TCLP	TCLP BNA	8270E	04/29/25	04/30/25	05/01/25	04/29/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1913
Client: AECOM

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	AECOM		Date Collected:	04/30/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/30/25	
Client Sample ID:	PB167774TB		SDG No.:	Q1913	
Lab Sample ID:	PB167774TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	116		10 - 139	78%	SPK: 150
13127-88-3	Phenol-d6	116		10 - 134	77%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.3		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.1		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	131		44 - 137	87%	SPK: 150
1718-51-0	Terphenyl-d14	67.9		48 - 125	68%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	218000		6.904		
1146-65-2	Naphthalene-d8	840000		8.186		
15067-26-2	Acenaphthene-d10	443000		9.945		
1517-22-2	Phenanthrene-d10	773000		11.427		
1719-03-5	Chrysene-d12	539000		14.068		
1520-96-3	Perylene-d12	394000		15.562		

Report of Analysis

Client:	AECOM		Date Collected:	04/30/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/30/25	
Client Sample ID:	PB167774TB		SDG No.:	Q1913	
Lab Sample ID:	PB167774TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/30/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/30/25	
Client Sample ID:	PB167805TB		SDG No.:	Q1913	
Lab Sample ID:	PB167805TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024487.D	1	04/30/25 13:15	05/01/25 11:18	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	128		10 - 139	85%	SPK: 150
13127-88-3	Phenol-d6	113		10 - 134	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.2		49 - 133	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	89.8		52 - 132	90%	SPK: 100
118-79-6	2,4,6-Tribromophenol	155		44 - 137	103%	SPK: 150
1718-51-0	Terphenyl-d14	86.0		48 - 125	86%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	147000	7.71			
1146-65-2	Naphthalene-d8	562000	10.481			
15067-26-2	Acenaphthene-d10	333000	14.345			
1517-22-2	Phenanthrene-d10	681000	17.151			
1719-03-5	Chrysene-d12	822000	21.616			
1520-96-3	Perylene-d12	1000000	24.962			

Report of Analysis

Client:	AECOM		Date Collected:	04/30/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/30/25	
Client Sample ID:	PB167805TB		SDG No.:	Q1913	
Lab Sample ID:	PB167805TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024487.D	1	04/30/25 13:15	05/01/25 11:18	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-12-S-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-01		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024489.D	1	04/30/25 13:15	05/01/25 12:44	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	134		10 - 139	89%	SPK: 150
13127-88-3	Phenol-d6	116		10 - 134	77%	SPK: 150
4165-60-0	Nitrobenzene-d5	91.9		49 - 133	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	97.1		52 - 132	97%	SPK: 100
118-79-6	2,4,6-Tribromophenol	169		44 - 137	113%	SPK: 150
1718-51-0	Terphenyl-d14	91.4		48 - 125	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	160000	7.711			
1146-65-2	Naphthalene-d8	611000	10.481			
15067-26-2	Acenaphthene-d10	350000	14.345			
1517-22-2	Phenanthrene-d10	678000	17.145			
1719-03-5	Chrysene-d12	789000	21.598			
1520-96-3	Perylene-d12	991000	24.951			

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-12-S-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-01		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024489.D	1	04/30/25 13:15	05/01/25 12:44	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-12-A-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-02		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024490.D	1	04/30/25 13:15	05/01/25 13:25	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	102		10 - 139	68%	SPK: 150
13127-88-3	Phenol-d6	78.0		10 - 134	52%	SPK: 150
4165-60-0	Nitrobenzene-d5	80.9		49 - 133	81%	SPK: 100
321-60-8	2-Fluorobiphenyl	83.4		52 - 132	83%	SPK: 100
118-79-6	2,4,6-Tribromophenol	154		44 - 137	103%	SPK: 150
1718-51-0	Terphenyl-d14	85.3		48 - 125	85%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	148000	7.71			
1146-65-2	Naphthalene-d8	578000	10.481			
15067-26-2	Acenaphthene-d10	347000	14.339			
1517-22-2	Phenanthrene-d10	702000	17.157			
1719-03-5	Chrysene-d12	848000	21.598			
1520-96-3	Perylene-d12	1030000	24.951			

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-12-A-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-02		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024490.D	1	04/30/25 13:15	05/01/25 13:25	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-13-S-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-03		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024499.D	1	04/30/25 13:15	05/01/25 19:36	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	115		10 - 139	77%	SPK: 150
13127-88-3	Phenol-d6	96.3		10 - 134	64%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.3		49 - 133	85%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.6		52 - 132	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	166		44 - 137	111%	SPK: 150
1718-51-0	Terphenyl-d14	111		48 - 125	111%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	208000	7.71			
1146-65-2	Naphthalene-d8	830000	10.487			
15067-26-2	Acenaphthene-d10	529000	14.345			
1517-22-2	Phenanthrene-d10	1080000	17.145			
1719-03-5	Chrysene-d12	861000	21.598			
1520-96-3	Perylene-d12	995000	24.951			

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-13-S-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-03		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024499.D	1	04/30/25 13:15	05/01/25 19:36	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-13-A-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-04		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142260.D	1	04/30/25 13:15	05/01/25 15:06	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	103		10 - 139	69%	SPK: 150
13127-88-3	Phenol-d6	96.1		10 - 134	64%	SPK: 150
4165-60-0	Nitrobenzene-d5	88.4		49 - 133	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	74.9		52 - 132	75%	SPK: 100
118-79-6	2,4,6-Tribromophenol	115		44 - 137	77%	SPK: 150
1718-51-0	Terphenyl-d14	55.4		48 - 125	55%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	205000	6.904			
1146-65-2	Naphthalene-d8	767000	8.187			
15067-26-2	Acenaphthene-d10	376000	9.939			
1517-22-2	Phenanthrene-d10	513000	11.428			
1719-03-5	Chrysene-d12	437000	14.063			
1520-96-3	Perylene-d12	450000	15.557			

Report of Analysis

Client:	AECOM		Date Collected:	04/29/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/29/25	
Client Sample ID:	WC-13-A-202504		SDG No.:	Q1913	
Lab Sample ID:	Q1913-04		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142260.D	1	04/30/25 13:15	05/01/25 15:06	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167774TB	PB167774TB	2-Fluorophenol	150	116	78		10	139
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	87.3	87		49	133
		2-Fluorobiphenyl	100	72.1	72		52	132
		2,4,6-Tribromophenol	150	131	87		44	137
		Terphenyl-d14	100	67.9	68		48	125
PB167805TB	PB167805TB	2-Fluorophenol	150	128	85		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	86.2	86		49	133
		2-Fluorobiphenyl	100	89.8	90		52	132
		2,4,6-Tribromophenol	150	155	103		44	137
		Terphenyl-d14	100	86.0	86		48	125
PB167810BL	PB167810BL	2-Fluorophenol	150	110	73		10	139
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	83.3	83		49	133
		2-Fluorobiphenyl	100	69.1	69		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	63.3	63		48	125
PB167810BS	PB167810BS	2-Fluorophenol	150	111	74		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	82.4	82		49	133
		2-Fluorobiphenyl	100	68.5	69		52	132
		2,4,6-Tribromophenol	150	132	88		44	137
		Terphenyl-d14	100	74.1	74		48	125
Q1901-08MS	B-167-SB01MS	2-Fluorophenol	150	96.6	64		10	139
		Phenol-d6	150	89.8	60		10	134
		Nitrobenzene-d5	100	84.0	84		49	133
		2-Fluorobiphenyl	100	70.7	71		52	132
		2,4,6-Tribromophenol	150	126	84		44	137
		Terphenyl-d14	100	70.4	70		48	125
Q1901-08MSD	B-167-SB01MSD	2-Fluorophenol	150	99.4	66		10	139
		Phenol-d6	150	92.8	62		10	134
		Nitrobenzene-d5	100	87.8	88		49	133
		2-Fluorobiphenyl	100	72.3	72		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	72.8	73		48	125
Q1913-01	WC-12-S-202504	2-Fluorophenol	150	134	89		10	139
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	91.9	92		49	133
		2-Fluorobiphenyl	100	97.1	97		52	132
		2,4,6-Tribromophenol	150	169	113		44	137
		Terphenyl-d14	100	91.4	91		48	125
Q1913-02	WC-12-A-202504	2-Fluorophenol	150	102	68		10	139
		Phenol-d6	150	78.0	52		10	134
		Nitrobenzene-d5	100	80.9	81		49	133
		2-Fluorobiphenyl	100	83.4	83		52	132
		2,4,6-Tribromophenol	150	154	103		44	137
		Terphenyl-d14	100	85.3	85		48	125
Q1913-03	WC-13-S-202504	2-Fluorophenol	150	115	77		10	139
		Phenol-d6	150	96.3	64		10	134
		Nitrobenzene-d5	100	85.3	85		49	133

Surrogate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q1913-03	WC-13-S-202504	2-Fluorobiphenyl	100	85.6	86		52	132
		2,4,6-Tribromophenol	150	166	111		44	137
		Terphenyl-d14	100	111	111		48	125
Q1913-04	WC-13-A-202504	2-Fluorophenol	150	103	69		10	139
		Phenol-d6	150	96.1	64		10	134
		Nitrobenzene-d5	100	88.4	88		49	133
		2-Fluorobiphenyl	100	74.9	75		52	132
		2,4,6-Tribromophenol	150	115	77		44	137
		Terphenyl-d14	100	55.4	55		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MS	Client Sample ID:	B-167-SB01MS					DataFile:	BF142277.D		
Pyridine	500	24.1	270	ug/L	49				10	109	
1,4-Dichlorobenzene	500	0	340	ug/L	68				55	125	
2-Methylphenol	500	0	360	ug/L	72				60	131	
3+4-Methylphenols	500	0	360	ug/L	72				54	136	
Hexachloroethane	500	0	350	ug/L	70				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	380	ug/L	76				52	125	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	500	ug/L	100				74	137	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	940	ug/L	94				52	162	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MSD	Client Sample ID:	B-167-SB01MSD					DataFile:	BF142278.D		
Pyridine	500	24.1	280	ug/L	51		4		10	109	20
1,4-Dichlorobenzene	500	0	340	ug/L	68		0		55	125	20
2-Methylphenol	500	0	370	ug/L	74		3		60	131	20
3+4-Methylphenols	500	0	370	ug/L	74		3		54	136	20
Hexachloroethane	500	0	370	ug/L	74		6		19	146	20
Nitrobenzene	500	0	460	ug/L	92		4		62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78		3		52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96		2		78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94		2		71	111	20
2,4-Dinitrotoluene	500	0	530	ug/L	106		6		74	137	20
Hexachlorobenzene	500	0	450	ug/L	90		2		72	115	20
Pentachlorophenol	1000	0	970	ug/L	97		3		52	162	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: 8270E DataFile: BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB167810BS	Pyridine	50	37.2	ug/L	74				29	97	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				76	103	
	2-Methylphenol	50	41.0	ug/L	82				69	109	
	3+4-Methylphenols	50	40.8	ug/L	82				67	106	
	Hexachloroethane	50	41.8	ug/L	84				76	118	
	Nitrobenzene	50	45.8	ug/L	92				58	106	
	Hexachlorobutadiene	50	41.5	ug/L	83				69	101	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				61	110	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				60	115	
	Hexachlorobenzene	50	41.6	ug/L	83				73	106	
	Pentachlorophenol	100	93.4	ug/L	93				47	114	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167810BL

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM Case No.: Q1913

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BF142254.D

Lab Sample ID: PB167810BL

Instrument ID: BNA_F

Date Extracted: 04/30/2025

Matrix: (soil/water) water

Date Analyzed: 05/01/2025

Level: (low/med) LOW

Time Analyzed: 12:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167810BS	PB167810BS	BF142255.D	05/01/2025
WC-13-A-202504	Q1913-04	BF142260.D	05/01/2025
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025
PB167805TB	PB167805TB	BP024487.D	05/01/2025
WC-12-S-202504	Q1913-01	BP024489.D	05/01/2025
WC-12-A-202504	Q1913-02	BP024490.D	05/01/2025
WC-13-S-202504	Q1913-03	BP024499.D	05/01/2025
PB167774TB	PB167774TB	BF142256.D	05/01/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BF142238.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07
WC-13-A-202504	Q1913-04	BF142260.D	05/01/2025	15:06

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BF142273.D

DFTPP Injection Date: 05/02/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.7
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	27.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	37.6
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142274.D	05/02/2025	10:39
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025	12:27
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025	12:55

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913

SDG NO.: Q1913

Lab File ID: BP024274.D

DFTPP Injection Date: 04/14/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.4
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.5
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	29.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	99.1
443	15.0 - 24.0% of mass 442	19.2 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024275.D	04/14/2025	11:06
SSTDICC005	SSTDICC005	BP024276.D	04/14/2025	11:47
SSTDICC010	SSTDICC010	BP024277.D	04/14/2025	12:27
SSTDICC020	SSTDICC020	BP024278.D	04/14/2025	13:08
SSTDICCC040	SSTDICCC040	BP024279.D	04/14/2025	13:49
SSTDICC050	SSTDICC050	BP024280.D	04/14/2025	15:10
SSTDICC060	SSTDICC060	BP024281.D	04/14/2025	16:32
SSTDICC080	SSTDICC080	BP024282.D	04/14/2025	17:13

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BP024485.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:57

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.7
68	Less than 2.0% of mass 69	0.5 (2) 1
69	Mass 69 relative abundance	25.9
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.5
275	10.0 - 60.0% of mass 198	26.6
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024486.D	05/01/2025	10:37
PB167805TB	PB167805TB	BP024487.D	05/01/2025	11:18
WC-12-S-202504	Q1913-01	BP024489.D	05/01/2025	12:44
WC-12-A-202504	Q1913-02	BP024490.D	05/01/2025	13:25
WC-13-S-202504	Q1913-03	BP024499.D	05/01/2025	19:36

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 PB167774TB	217516	6.90	839699	8.19	443070	9.95
02 PB167810BL	223997	6.90	857825	8.19	449926	9.95
03 PB167810BS	228833	6.91	892782	8.19	472284	9.95
04 WC-13-A-202504	204789	6.90	767468	8.19	376081	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
EPA SAMPLE NO.						
01 PB167774TB	773002	11.43	539235	14.07	394051	15.56
02 PB167810BL	794711	11.43	568207	14.07	404816	15.56
03 PB167810BS	804874	11.43	464382	14.07	446826	15.57
04 WC-13-A-202504	513009	11.43	437449	14.06	449964	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	189457	6.91	747235	8.19	389025	9.95
UPPER LIMIT	378914	7.41	1494470	8.692	778050	10.445
LOWER LIMIT	94728.5	6.41	373618	7.692	194513	9.445
EPA SAMPLE NO.						
01 B-167-SB01MS	212656	6.91	808361	8.19	402059	9.95
02 B-167-SB01MSD	217028	6.91	832936	8.19	419407	9.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	652685	11.427	354752	14.068	327562	15.557
UPPER LIMIT	1305370	11.927	709504	14.568	655124	16.057
LOWER LIMIT	326343	10.927	177376	13.568	163781	15.057
EPA SAMPLE NO.						
01 B-167-SB01MS	629997	11.43	370460	14.07	379220	15.56
02 B-167-SB01MSD	684254	11.43	400079	14.07	401381	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BP024486.D Time Analyzed: 10:37

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	147752	7.711	596585	10.48	373130	14.35
UPPER LIMIT	295504	8.211	1193170	10.981	746260	14.846
LOWER LIMIT	73876	7.211	298293	9.981	186565	13.846
EPA SAMPLE NO.						
01 WC-13-S-202504	207548	7.71	829684	10.49	528997	14.35
02 PB167805TB	147117	7.71	562398	10.48	333420	14.35
03 WC-12-S-202504	159742	7.71	610644	10.48	349744	14.35
04 WC-12-A-202504	147765	7.71	578416	10.48	346607	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BP024486.D Time Analyzed: 10:37

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	737949	17.145	820131	21.592	979134	24.939
UPPER LIMIT	1475900	17.645	1640260	22.092	1958270	25.439
LOWER LIMIT	368975	16.645	410066	21.092	489567	24.439
EPA SAMPLE NO.						
01 WC-13-S-202504	1075370	17.15	861399	21.60	995064	24.95
02 PB167805TB	681369	17.15	822212	21.62	999851	24.96
03 WC-12-S-202504	678455	17.15	788921	21.60	990879	24.95
04 WC-12-A-202504	702177	17.16	848293	21.60	1030920	24.95

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:		
Client Sample ID:	PB167810BL		SDG No.:	Q1913	
Lab Sample ID:	PB167810BL		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.53	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	110		10 - 139	73%	SPK: 150
13127-88-3	Phenol-d6	109		10 - 134	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.3		49 - 133	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.1		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	124		44 - 137	83%	SPK: 150
1718-51-0	Terphenyl-d14	63.3		48 - 125	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	224000		6.904		
1146-65-2	Naphthalene-d8	858000		8.187		
15067-26-2	Acenaphthene-d10	450000		9.945		
1517-22-2	Phenanthrene-d10	795000		11.428		
1719-03-5	Chrysene-d12	568000		14.069		
1520-96-3	Perylene-d12	405000		15.563		

Report of Analysis

Client:	AECOM		Date Collected:	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	
Client Sample ID:	PB167810BL		SDG No.:	Q1913
Lab Sample ID:	PB167810BL		Matrix:	TCLP
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:		
Client Sample ID:	PB167810BS		SDG No.:	Q1913	
Lab Sample ID:	PB167810BS		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	37.2		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	39.6		0.53	5.00	ug/L
95-48-7	2-Methylphenol	41.0		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	40.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	41.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.8		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	41.5		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	43.4		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.8		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	41.6		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	93.4	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	111		10 - 139	74%	SPK: 150
13127-88-3	Phenol-d6	113		10 - 134	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.4		49 - 133	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.5		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	74.1		48 - 125	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	229000		6.91		
1146-65-2	Naphthalene-d8	893000		8.192		
15067-26-2	Acenaphthene-d10	472000		9.945		
1517-22-2	Phenanthrene-d10	805000		11.433		
1719-03-5	Chrysene-d12	464000		14.074		
1520-96-3	Perylene-d12	447000		15.568		

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:		
Client Sample ID:	PB167810BS		SDG No.:	Q1913	
Lab Sample ID:	PB167810BS		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/26/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/28/25	
Client Sample ID:	B-167-SB01MS		SDG No.:	Q1913	
Lab Sample ID:	Q1901-08MS		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	270		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	360		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	360		11.0	100	ug/L
67-72-1	Hexachloroethane	350		6.50	50.0	ug/L
98-95-3	Nitrobenzene	440		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	380		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	460		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	500		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	940	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	96.6		10 - 139	64%	SPK: 150
13127-88-3	Phenol-d6	89.8		10 - 134	60%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.0		49 - 133	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		52 - 132	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.4		48 - 125	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	213000	6.91			
1146-65-2	Naphthalene-d8	808000	8.193			
15067-26-2	Acenaphthene-d10	402000	9.945			
1517-22-2	Phenanthrene-d10	630000	11.433			
1719-03-5	Chrysene-d12	370000	14.069			
1520-96-3	Perylene-d12	379000	15.557			

Report of Analysis

Client:	AECOM		Date Collected:	04/26/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/28/25	
Client Sample ID:	B-167-SB01MS		SDG No.:	Q1913	
Lab Sample ID:	Q1901-08MS		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	04/26/25	
Project:	Meeker Ave Plumes Superfund Site RI FS		Date Received:	04/28/25	
Client Sample ID:	B-167-SB01MSD		SDG No.:	Q1913	
Lab Sample ID:	Q1901-08MSD		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	280		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	370		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	370		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	460		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	530		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	450		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	970	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.4		10 - 139	66%	SPK: 150
13127-88-3	Phenol-d6	92.8		10 - 134	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.8		49 - 133	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.3		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	72.8		48 - 125	73%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	217000		6.91		
1146-65-2	Naphthalene-d8	833000		8.193		
15067-26-2	Acenaphthene-d10	419000		9.945		
1517-22-2	Phenanthrene-d10	684000		11.433		
1719-03-5	Chrysene-d12	400000		14.069		
1520-96-3	Perylene-d12	401000		15.563		

Report of Analysis

Client:	AECOM	Date Collected:	04/26/25
Project:	Meeker Ave Plumes Superfund Site RI FS	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1913
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 30 16:00:01 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF142246.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559	4.03	
3) Pyridine	1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454	3.62	
4) n-Nitrosodimet...	0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752	4.95	
5) S 2-Fluorophenol	1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252	5.35	
6) Aniline	2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140	4.29	
7) S Phenol-d6	1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512	4.85	
8) 2-Chlorophenol	1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294	4.64	
9) Benzaldehyde	1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057	8.91	
10) C Phenol	1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628	5.15	
11) bis(2-Chloroet...	1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262	5.46	
12) 1,3-Dichlorobe...	1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464	6.18	
13) C 1,4-Dichlorobe...	1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475	6.86	
14) 1,2-Dichlorobe...	1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408	5.67	
15) Benzyl Alcohol	1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073	4.69	
16) 2,2'-oxybis(1-...	2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278	5.19	
17) 2-Methylphenol	1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060	4.19	
18) Hexachloroethane	0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483	4.84	
19) P n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20) 3+4-Methylphenols	1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326	5.90	

21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463	7.70	
23) S Nitrobenzene-d5	0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304	16.88	
24) Nitrobenzene	0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299	13.31	
25) Isophorone	0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639	3.36	
26) C 2-Nitrophenol	0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102	38.38#	
27) 2,4-Dimethylph...	0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322	3.96	
28) bis(2-Chloroet...	0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403	4.98	
29) C 2,4-Dichloroph...	0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269	6.12	
30) 1,2,4-Trichlor...	0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301	5.28	
31) Naphthalene	1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997	8.21	
32) Benzoic acid	0.048	0.087	0.130	0.146	0.159	0.179	0.125	0.125	39.24	
33) 4-Chloroaniline	0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415	5.98	
34) C Hexachlorobuta...	0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175	4.35	
35) Caprolactam	0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081	9.45	
36) C 4-Chloro-3-met...	0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277	4.84	
37) 2-Methylnaphth...	0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613	7.69	
38) 1-Methylnaphth...	0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635	8.29	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88
41) P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23
42) S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73
43) C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50
44)	2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17
45) S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11
46)	1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38
47)	2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51
48)	2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25
49)	Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93
50)	Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42
51)	2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45
52) C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63
53)	3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09
54) P	2,4-Dinitrophenol		0.025	0.036	0.057	0.069	0.076	0.085	0.058	40.36
55)	Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67
56) P	4-Nitrophenol		0.131	0.177	0.216	0.233	0.230	0.220	0.201	19.82
57)	2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96
58)	Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74
59)	2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65
60)	Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61
61)	4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06
62)	4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57
63)	Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...		0.022	0.034	0.055	0.065	0.071	0.075	0.054	39.23
66) c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11
67)	4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39
68)	Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28
69)	Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29
70) C	Pentachlorophenol		0.093	0.121	0.144	0.145	0.144	0.144	0.132	16.25
71)	Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45
72)	Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77
73)	Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77
74)	Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88
75) C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67
78)	Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54
79) S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41
80)	Butylbenzylphth...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32
81)	Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06
82)	3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37
83)	Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12
84)	Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69
85) c	Di-n-octyl pht...		0.668	0.875	1.262	1.320	1.347	1.329	1.133	25.55

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.323	1.384	1.466	1.613	1.556	1.490	1.422	1.465		6.78
88)	Benzo(b)fluora...	1.317	1.220	1.208	1.302	1.271	1.177	1.202	1.242		4.36
89)	Benzo(k)fluora...	1.211	1.159	1.222	1.145	1.146	1.110	0.962	1.136		7.59
90) C	Benzo(a)pyrene	1.121	1.098	1.146	1.203	1.180	1.137	1.076	1.137		3.91
91)	Dibenzo(a,h)an...	1.075	1.131	1.197	1.312	1.271	1.192	1.144	1.189		6.90
92)	Benzo(g,h,i)pe...	1.076	1.138	1.204	1.308	1.267	1.213	1.162	1.195		6.55

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP041425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Apr 15 04:48:42 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BP024275.D 5 =BP024276.D 10 =BP024277.D 20 =BP024278.D 40 =BP024279.D 50 =BP024280.D 60 =BP024281.D 80 =BP024282.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.551	0.517	0.552	0.512	0.481	0.490	0.477	0.512	6.10	
3)	Pyridine	1.265	1.280	1.422	1.396	1.333	1.383	1.378	1.351	4.43	
4)	n-Nitrosodimet...	0.424	0.443	0.469	0.457	0.442	0.460	0.444	0.448	3.29	
5) S	2-Fluorophenol	1.123	1.145	1.262	1.254	1.210	1.277	1.196	1.210	4.93	
6)	Aniline	1.439	1.487	1.636	1.587	1.403	1.428	1.380	1.480	6.52	
7) S	Phenol-d6	1.463	1.552	1.718	1.748	1.697	1.769	1.647	1.656	6.75	
8)	2-Chlorophenol	1.260	1.297	1.412	1.399	1.359	1.400	1.328	1.350	4.30	
9)	Benzaldehyde	0.943	0.934	0.942	0.836	0.732	0.726	0.621	0.819	15.70	
10) C	Phenol	1.475	1.565	1.720	1.750	1.698	1.779	1.643	1.661	6.54	
11)	bis(2-Chloroet...	1.278	1.317	1.399	1.398	1.321	1.358	1.300	1.339	3.56	
12)	1,3-Dichlorobe...	1.464	1.486	1.534	1.486	1.402	1.433	1.373	1.454	3.79	
13) C	1,4-Dichlorobe...	1.496	1.489	1.550	1.494	1.436	1.463	1.399	1.475	3.28	
14)	1,2-Dichlorobe...	1.472	1.441	1.502	1.445	1.362	1.411	1.339	1.424	4.10	
15)	Benzyl Alcohol	0.857	0.935	1.106	1.158	1.142	1.178	1.122	1.071	11.58	
16)	2,2'-oxybis(1-...	1.452	1.496	1.517	1.514	1.405	1.397	1.348	1.447	4.54	
17)	2-Methylphenol	0.915	1.002	1.123	1.140	1.105	1.150	1.087	1.075	7.97	
18)	Hexachloroethane	0.532	0.537	0.558	0.542	0.521	0.533	0.510	0.533	2.88	
19) P	n-Nitroso-di-n...	0.941	1.008	1.008	1.088	1.084	1.041	1.038	0.989	1.025	4.76
20)	3+4-Methylphenols	1.224	1.365	1.560	1.592	1.567	1.607	1.525	1.491	9.57	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.470	0.487	0.522	0.519	0.490	0.495	0.483	0.495	3.85	
23) S	Nitrobenzene-d5	0.322	0.337	0.367	0.368	0.357	0.356	0.347	0.351	4.76	
24)	Nitrobenzene	0.328	0.334	0.362	0.364	0.349	0.353	0.340	0.347	3.96	
25)	Isophorone	0.575	0.601	0.653	0.673	0.650	0.654	0.639	0.635	5.41	
26) C	2-Nitrophenol	0.130	0.143	0.172	0.183	0.184	0.191	0.186	0.170	13.97	
27)	2,4-Dimethylph...	0.180	0.194	0.219	0.226	0.225	0.228	0.221	0.213	8.81	
28)	bis(2-Chloroet...	0.412	0.426	0.444	0.452	0.430	0.425	0.412	0.429	3.51	
29) C	2,4-Dichloroph...	0.242	0.268	0.297	0.308	0.307	0.311	0.302	0.291	8.94	
30)	1,2,4-Trichlor...	0.308	0.305	0.318	0.321	0.309	0.315	0.306	0.311	2.00	
31)	Naphthalene	1.044	1.054	1.091	1.079	1.037	1.048	1.023	1.054	2.27	
32)	Benzoic acid	0.185	0.226	0.242	0.270	0.285	0.282	0.248		15.56	
33)	4-Chloroaniline	0.328	0.349	0.389	0.397	0.378	0.379	0.378	0.371	6.46	
34) C	Hexachlorobuta...	0.181	0.181	0.186	0.189	0.181	0.182	0.182	0.183	1.62	
35)	Caprolactam	0.085	0.092	0.111	0.112	0.115	0.121	0.115	0.107	12.53	
36) C	4-Chloro-3-met...	0.284	0.310	0.349	0.357	0.356	0.363	0.351	0.339	8.82	
37)	2-Methylnaphth...	0.700	0.720	0.755	0.755	0.733	0.740	0.712	0.731	2.89	
38)	1-Methylnaphth...	0.700	0.714	0.737	0.731	0.712	0.718	0.691	0.715	2.25	

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP041425.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.522	0.538	0.568	0.576	0.544	0.549	0.552	0.550	3.33	
41) P	Hexachlorocycl...	0.162	0.185	0.197	0.218	0.199	0.197	0.204	0.195	8.92	
42) S	2,4,6-Tribromo...	0.235	0.256	0.278	0.289	0.289	0.296	0.294	0.277	8.29	
43) C	2,4,6-Trichlor...	0.301	0.335	0.369	0.389	0.386	0.390	0.390	0.366	9.51	
44)	2,4,5-Trichlor...	0.332	0.371	0.409	0.427	0.429	0.436	0.430	0.405	9.62	
45) S	2-Fluorobiphenyl	1.332	1.334	1.359	1.366	1.263	1.251	1.240	1.306	4.08	
46)	1,1'-Biphenyl	1.454	1.489	1.528	1.529	1.434	1.428	1.429	1.470	3.07	
47)	2-Chloronaphth...	1.063	1.100	1.135	1.139	1.085	1.090	1.078	1.099	2.60	
48)	2-Nitroaniline	0.250	0.270	0.317	0.330	0.324	0.335	0.331	0.308	10.98	
49)	Acenaphthylene	1.604	1.680	1.793	1.814	1.726	1.763	1.730	1.730	4.11	
50)	Dimethylphthalate	1.429	1.416	1.495	1.495	1.460	1.460	1.426	1.454	2.22	
51)	2,6-Dinitrotol...	0.264	0.290	0.319	0.325	0.319	0.328	0.317	0.309	7.58	
52) C	Acenaphthene	1.087	1.087	1.141	1.126	1.081	1.079	1.076	1.097	2.35	
53)	3-Nitroaniline	0.250	0.289	0.338	0.354	0.342	0.340	0.358	0.325	12.25	
54) P	2,4-Dinitrophenol		0.125	0.163	0.186	0.196	0.210	0.211	0.182	18.22	
55)	Dibenzofuran	1.801	1.807	1.871	1.833	1.746	1.768	1.723	1.793	2.84	
56) P	4-Nitrophenol	0.190	0.231	0.289	0.305	0.306	0.323	0.317	0.280	17.90	
57)	2,4-Dinitrotol...	0.334	0.368	0.439	0.448	0.442	0.463	0.456	0.421	11.82	
58)	Fluorene	1.376	1.397	1.454	1.398	1.368	1.366	1.356	1.388	2.40	
59)	2,3,4,6-Tetrac...	0.327	0.351	0.380	0.389	0.387	0.390	0.389	0.373	6.57	
60)	Diethylphthalate	1.451	1.484	1.528	1.550	1.499	1.499	1.481	1.499	2.16	
61)	4-Chlorophenyl...	0.667	0.675	0.699	0.680	0.667	0.670	0.659	0.674	1.95	
62)	4-Nitroaniline	0.263	0.300	0.353	0.369	0.364	0.380	0.376	0.344	12.99	
63)	Azobenzene	1.269	1.383	1.446	1.431	1.385	1.385	1.344	1.378	4.26	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.095	0.118	0.131	0.134	0.140	0.138	0.126	13.50	
66) c	n-Nitrosodiphe...	0.567	0.585	0.616	0.630	0.591	0.608	0.580	0.597	3.71	
67)	4-Bromophenyl-...	0.203	0.212	0.220	0.230	0.220	0.225	0.222	0.219	4.11	
68)	Hexachlorobenzene	0.247	0.247	0.260	0.271	0.261	0.269	0.263	0.260	3.67	
69)	Atrazine	0.176	0.167	0.134	0.121	0.162			0.152	15.57	
70) C	Pentachlorophenol	0.143	0.155	0.175	0.196	0.197	0.203	0.201	0.181	13.35	
71)	Phenanthrene	1.098	1.086	1.117	1.127	1.070	1.082	1.058	1.091	2.27	
72)	Anthracene	0.995	1.017	1.084	1.111	1.053	1.071	1.030	1.052	3.85	
73)	Carbazole	0.968	0.998	1.060	1.084	1.014	1.037	1.031	1.027	3.78	
74)	Di-n-butylphth...	1.147	1.230	1.253	1.398	1.310	1.338	1.284	1.280	6.30	
75) C	Fluoranthene	1.289	1.274	1.316	1.347	1.280	1.304	1.274	1.298	2.07	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.144	0.141	0.109	0.436	0.333	0.302	0.307	0.254	48.39	
78)	Pyrene	1.192	1.236	1.345	1.305	1.294	1.305	1.219	1.271	4.38	
79) S	Terphenyl-d14	0.967	0.998	1.055	1.032	1.005	0.964	0.926	0.992	4.42	
80)	Butylbenzylpht...	0.434	0.488	0.546	0.596	0.580	0.593	0.574	0.544	11.25	
81)	Benzo(a)anthra...	1.195	1.214	1.292	1.293	1.234	1.253	1.221	1.243	3.08	
82)	3,3'-Dichlorob...	0.345	0.383	0.431	0.486	0.463	0.485	0.460	0.436	12.30	
83)	Chrysene	1.180	1.181	1.229	1.213	1.173	1.193	1.167	1.191	1.88	
84)	Bis(2-ethylhex...	0.639	0.735	0.779	0.893	0.853	0.867	0.819	0.798	11.09	
85) c	Di-n-octyl pht...	0.945	1.093	1.222	1.461	1.430	1.486	1.446	1.297	16.46	

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.252	1.304	1.407	1.444	1.401	1.445	1.466	1.388	5.77
88)	Benzo(b)fluora...	1.130	1.174	1.217	1.258	1.196	1.205	1.212	1.199	3.30
89)	Benzo(k)fluora...	1.088	1.131	1.217	1.212	1.154	1.166	1.138	1.158	3.94
90) C	Benzo(a)pyrene	0.939	0.971	1.057	1.091	1.053	1.070	1.058	1.034	5.44
91)	Dibenzo(a,h)an...	1.046	1.089	1.172	1.203	1.169	1.186	1.202	1.152	5.28
92)	Benzo(g,h,i)pe...	1.073	1.110	1.198	1.218	1.180	1.209	1.224	1.173	4.99

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913

Instrument ID: BNA_P Calibration Date/Time: 05/01/2025 10:37

Lab File ID: BP024486.D Init. Calib. Date(s): 04/14/2025 04/14/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:06 17:13

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.351	1.120		-17.1	
2-Fluorophenol	1.210	1.140		-5.8	
Phenol-d6	1.656	1.494		-9.8	
1,4-Dichlorobenzene	1.475	1.388		-5.9	20.0
2-Methylphenol	1.075	1.000		-7.0	
3+4-Methylphenols	1.491	1.403		-5.9	
Nitrobenzene-d5	0.351	0.340		-3.1	
Hexachloroethane	0.533	0.513		-3.8	
Nitrobenzene	0.347	0.327		-5.8	
Hexachlorobutadiene	0.183	0.199		8.7	20.0
2,4,6-Trichlorophenol	0.366	0.391		6.8	20.0
2-Fluorobiphenyl	1.306	1.322		1.2	
2,4,5-Trichlorophenol	0.405	0.427		5.4	
2,4-Dinitrotoluene	0.421	0.448		6.4	
2,4,6-Tribromophenol	0.277	0.329		18.8	
Hexachlorobenzene	0.260	0.285		9.6	
Pentachlorophenol	0.181	0.199		9.9	20.0
Terphenyl-d14	0.992	0.998		0.6	

All other compounds must meet a minimum RRF of 0.010.