

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MS	Client Sample ID:	B-167-SB01MS					DataFile:	BF142277.D		
Pyridine	500	24.1	270	ug/L	49				10	109	
1,4-Dichlorobenzene	500	0	340	ug/L	68				55	125	
2-Methylphenol	500	0	360	ug/L	72				60	131	
3+4-Methylphenols	500	0	360	ug/L	72				54	136	
Hexachloroethane	500	0	350	ug/L	70				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	380	ug/L	76				52	125	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	500	ug/L	100				74	137	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	940	ug/L	94				52	162	