

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MSD	Client Sample ID:	B-167-SB01MSD					DataFile:	BF142278.D		
Pyridine	500	24.1	280	ug/L	51		4		10	109	20
1,4-Dichlorobenzene	500	0	340	ug/L	68		0		55	125	20
2-Methylphenol	500	0	370	ug/L	74		3		60	131	20
3+4-Methylphenols	500	0	370	ug/L	74		3		54	136	20
Hexachloroethane	500	0	370	ug/L	74		6		19	146	20
Nitrobenzene	500	0	460	ug/L	92		4		62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78		3		52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96		2		78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94		2		71	111	20
2,4-Dinitrotoluene	500	0	530	ug/L	106		6		74	137	20
Hexachlorobenzene	500	0	450	ug/L	90		2		72	115	20
Pentachlorophenol	1000	0	970	ug/L	97		3		52	162	20