

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 PB167774TB	217516	6.90	839699	8.19	443070	9.95
02 PB167810BL	223997	6.90	857825	8.19	449926	9.95
03 PB167810BS	228833	6.91	892782	8.19	472284	9.95
04 WC-13-A-202504	204789	6.90	767468	8.19	376081	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.